

AMENDMENT NO.1 Dated 17.05.2020 Ref No: TENDER NO: HLL/SD/RBD/2019-20/TENDER/15 DT:05.03.2020 WHICH HAS NOW BEEN CLOSED, FRESH REQUEST FOR PROPOSAL FOR THE SUPPLY OF PERSONAL PROTECTIVE EQUIPMENTS -GOGGLES, N95 FACE MASK WITH / WITHOUT EXPIRATORY VALVE AND NITRILE GLOVES floated on 11 May 2020

The following amendment has been incorporated to the bid document for the above tender:

1. Product requirement with Quantity

SI No	Item	Total tentative Quantity required	UoM
1	Personal Protective Goggles	90 lakh	Nos
2	N95 face mask with / without expiratory valve	150 lakh	nos
3	Nitrile Gloves 6.5 Size	125 lakh	Pairs
4	Nitrile Gloves 7.0 Size	125 lakh	Pairs

MAY BE READ AS

1. Product requirement with Quantity

SI No	Item	Total tentative Quantity required	UoM
1	Personal Protective Goggles	45 lakh	Nos
2	N95 face mask with / without expiratory valve	150 lakh	nos
3	Nitrile Gloves 6.5 Size / Latex Examination Gloves Size Medium	125 lakh	Pairs
4	Nitrile Gloves 7.0 Size / Latex Examination Gloves Size Large	125 lakh	Pairs

Specification

1. Goggles

1. With transparent glasses, Zero power, well fitting, covered from all sides with elastic band/or adjustable holder.
2. Good seal with the skin of the face.
3. Flexible frame to easily fit all face contours without too much pressure.
4. Covers the eyes and surrounding areas and accommodates for prescription glasses.
5. Fog and scratch resistant.
6. Adjustable band to secure firmly so as not to become loose during clinical activity.
7. Indirect venting to reduce fogging.
8. May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable.
9. Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010

MAY BE READ AS

1. Goggles

1. With transparent glasses, Zero power, well fitting, covered from all sides with elastic band/or adjustable holder.
2. Good seal with the skin of the face.
3. Flexible frame to easily fit all face contours without too much pressure.
4. Covers the eyes and surrounding areas and accommodates for prescription glasses.
5. Fog and scratch resistant.
6. Adjustable band to secure firmly so as not to become loose during clinical activity.
7. Indirect venting to reduce fogging.
8. May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable.
9. Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010
 - c. IS 5983-1980, reaffirmed in 2018

6. Gloves in pairs of Size 6.5 & 7.0 each
- Nitrile
 - Non-sterile
 - Powder free
 - Outer gloves preferably reach mid-forearm (minimum 280mm total length)
 - Different sizes in pairs of 6.5 & 7.0 each
 - Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89/686/EEC category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10
 - Packing – The Product should be supplied in boxes of 100 pieces (The gloves of size 6.5 & 7.0 shall be individually packed).

MAY BE READ AS

6. Gloves in pairs of Size 6.5 & 7.0 each
- Nitrile
 - Non-sterile
 - Powder free
 - Outer gloves preferably reach mid-forearm (minimum 280mm total length)
 - Different sizes in pairs of 6.5 & 7.0 each
 - Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89/686/EEC category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10
 - Packing – The Product should be supplied in boxes of 100 pieces (The gloves of size 6.5 & 7.0 shall be individually packed).

OR

Latex Medical Examination Gloves

- Ambidextrous shape – fits either hand
- Size: Medium and Large in pairs
- Single use
- Non Sterile
- Should conform to IS 15354 Part I 2018 or EN455 or ISO 11193 -1 of 2008.

- Expiry: minimum 3 years from date of manufacture
- Label on primary packaging – name and/ or trademark of manufacturer, manufacturer's product reference, expiry date by year and month

All other relevant specifications and clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President Sourcing Division HLL Lifecare Ltd. HLL Bhavan, Poojappura, Thiruvananthapuram.

Offers may be submitted to:

sdcovidep.hll@lifecarehll.com / sdcovidep.hll@gmail.com

Contact: 0471 2354949 Ext: 323 (9.15 AM to 5.45 PM)