

GAS OULETS FOR TRAUMA, AIIMS RISHIKESH										
LEVEL	DESCRIPTION OF ROOM	ROOM NO	NO OF OULETS						SINGLE ARM MOVABLE	BEDHEAD PANEL
			OXYGEN	VACUUM	AIR 4 bar	AIR 7 bar	N2O	AGSS		
GROUND	CT SCAN ROOM	TB 1020	1	1	0	0	0	0	0	0
GROUND	MRI ROOM	TB 1027	1	1	0	0	0	0	0	0
GROUND	POST OPERATIVE WARD(4 BED)	TB 1048	8	8	4	0	0	0	0	4
GROUND	DISASTER WARD(10 BED)	TB 1053	6	6	0	0	0	0	0	0
GROUND	TRIAGE(8 BED)	TB 1055	6	6	0	0	0	0	0	0
GROUND	TRIAGE(8 BED)	TB 1056	6	6	0	0	0	0	0	0
GROUND	EMERGENCY MINOR OT	TB 1064	2	2	2	0	1	1	1	0
GROUND	RESUSCITATION ROOM	TB 1065	2	2	1	0	0	0	0	1
FIRST	VIP ROOM	TB 2001	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2003	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2005	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2007	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2009	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2011	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2017	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2019	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2020	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2021	1	1	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2026	4	4	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2029	4	4	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2042	4	4	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2045	4	4	0	0	0	0	0	0
FIRST	ISOLATION ROOM	TB 2046	2	2	1	0	0	0	0	1
FIRST	WARD(6 BED)	TB 2050	4	4	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2053	4	4	0	0	0	0	0	0
FIRST	ISOLATION ROOM	TB 2057	2	2	1	0	0	0	0	1
FIRST	WARD(6 BED)	TB 2058	4	4	0	0	0	0	0	0
FIRST	WARD(6 BED)	TB 2061	4	4	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2070	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2072	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2074	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2076	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2078	1	1	0	0	0	0	0	0
FIRST	VIP ROOM	TB 2079	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2080	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2082	1	1	0	0	0	0	0	0
FIRST	PRIVATE ROOM	TB 2084	1	1	0	0	0	0	0	0

LEVEL	DESCRIPTION OF ROOM	ROOM NO	NO OF OUTLETS						SINGLE ARM MOVABLE	BEDHEAD PANEL
			OXYGEN	VACUUM	AIR 4 bar	AIR 7 bar	N2O	AGSS		
FIRST	PRIVATE ROOM	TB 2085	1	1	0	0	0	0	0	0
SECOND	ICU&NURSE STATION(10 BED)	TB 3033	20	20	10	0	0	0	0	10
SECOND	ISOLATION ROOM	TB 3050	2	2	1	0	0	0	0	1
SECOND	RECOVERY ROOM(6 BED)	TB 3076	12	12	6	0	0	0	0	6
	TOTAL		122	122	26	0	1	1	1	24