

Corrigendum No: 1 Issued on dated 21/03/2014

Ref: Tender Enquiry No: HLL/PCD/RIMS-03/13-14 and Pre-Bid meeting held on 12/03/2014 at HLL Lifecare Ltd., B-14A, Sector-62, Noida - 201 307

The following amendments are issued with respect to the above tender Enquiry Document.

Section VII, Technical Specifications

Item No. 1

C-Arm Image Intensifier

1. For Existing Specification: X-RAY GENERATOR

- 5. mA in fluoroscopy : 0.1 to 4 mA or more in normal fluoroscopy and 12 mA or more in High Level Fluoroscopy
- 7. Should have Digital Spot for high quality single image, 16 mA or more

Read as:

- 5. mA in fluoroscopy : 0.1 to 4 mA or more in normal fluoroscopy and 10 mA or more in High Level Fluoroscopy
- 7. Should have Digital Spot for high quality single image, 12 mA or more

2. For Existing Specification: CONTROL PANEL (DIGITAL WORK STATION)

- 11. Input diameter 9" with Triple field (9/6/4)
- 17. Min 18 inch or more , black and white, flicker free, high resolution (1280x1024 pixels or better), medical grade flat screen TFT, automatic and manual control of brightness and contrast, mounted on mobile trolley with locking device

Read as:

- 11. Input diameter 9" with Triple field (9/6/4.5)
- 17. Min 18 inch or more ,flicker free, high resolution (1280x1024 pixels or better), medical grade flat screen TFT, automatic and manual control of brightness and contrast, mounted on mobile trolley with locking device

Item No. 3

Mobile X-Ray Machine 100 mA

1. For Existing Specification: CERTIFICATION:

1.System shall have valid AERB certificate of the quoted model.

Read as:

1.System shall have valid AERB & CE certificate of the quoted model.

Added Para: (Item No. 1 C-Arm Image Intensifier & Item No. 3 Mobile X-Ray Machine 100 mA)

All X-Ray equipment should have AERB type approval.

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.