

HLL LIFECARE LIMITED
(A Government of India Enterprise)
Projects Division
Corporate Head Office, Poojappura.P.O,
Thiruvananthapuram – 695012,
Kerala, India
Phn: 0471- 2354949, 2775588

INVITATION FOR BIDS (IFB)

IFB No : HLL/CHO/PROJ/HCS/MCG/ECG/2019-20
Date : 07-09-2019

HLL Lifecare Limited (HLL), a Government of India Enterprise invites bids from competent and experienced Suppliers/Dealers who are capable to do the following supply/work meeting their requirements as per our tender.

Sl. No	Brief Description of Item/Work	Qty
1	Supply, Installation, Testing & Commissioning of ECG-12 CHANNEL EQUIPMENT with accessories at Diagnostic Lab Facility, Gurugram	As per Specification

2. The scheduled date for issue, receipt and opening of bids is as follows.
- Last date and time for receipt of bids - 21-09-2019 up to 15.00 Hrs.
 - Date and time of opening of bids - 21-09-2019, 15.30 Hrs.
 - Address for receipt and place of opening of bids:

HLL Lifecare Limited,
Corporate Head Office, Poojappura.P.O,
Thiruvananthapuram – 695012, Kerala.
E-mail: harikrishnankp@lifecarehll.com, choprojects@lifecarehll.com

3. **Price:** The price quoted should be inclusive of all material cost, loading and unloading charges, installation & commissioning charges, Warranty charges, GST and all applicable taxes and other levies, labor charges, insurance etc. Price quoted should be firm without any escalation till the order is completely executed.
4. **Completion Time:** Supply and installation should be completed within 1 WEEK from the date of issuing the Letter Of Intent or Order.
5. **Payment Terms:** 100% payment will be paid after satisfactory Supply, Installation, Testing and Commissioning of the item as per the schedule at Diagnostic Lab Facility, Gurugram ,and submission of original warranty certificate.

Tax Deduction: All statutory deductions like Income Tax, Works Contract Tax, E.S.I., P.F. or any other government-imposed liabilities shall be borne by the Contractor (as applicable at the time of execution of job) and shall be deducted from each bill submitted by the Contractor.

6. **Guarantee/Warranty Period:** The Guarantee against any possible manufacturing defects shall be given for a period of 12 months from the date of supply and inspection of the item at HLL site.

The company reserves the right to club or split the items of works, change the qualifying criteria at their discretion and to reject or cancel the Invitation for bids without assigning any reason thereof.

SENIOR MANAGER (PROJECTS)

SCHEDULE OF ITEMS

Sub:	Supply, Installation, Testing & Commissioning of ECG-12 CHANNEL EQUIPMENT with accessories for Diagnostic Lab Facility, Gurugram. – reg.			
S No	Item Description	Qty	Unit	Amount in Rs.
1	<p>Supply, Installation, Testing & Commissioning of ECG-12 CHANNEL EQUIPMENT with accessories at Diagnostic Lab Facility, Gurugram</p> <ul style="list-style-type: none"> • Fully automatic, microprocessor controlled mains & battery operated ECG Unit, including: <ul style="list-style-type: none"> • built-in 8 inch or more color LCD • built-in Ethernet port for data transfer • printout on internal thermal printer. ❖ Resting ECG software with 12 simultaneous leads ❖ Preview of 12 lead ECG prior to printout ❖ ECG measurement software , incl. average complexes, measurement markers and detailed measurement results table ❖ Unit should have advisory software for placement of leads ❖ Unit should give lead quality assessment as to whether properly connected or improper connection ❖ Unit should have Pace maker detection ❖ Unit should have sampling rate of atleast 30,000 Hz ❖ Unit should have frequency range 0 Hz - 250 Hz ❖ Unit Should be able to print Right precordial lead with notation ❖ Should be able to print left posterior leads ❖ Unit should have high-resolution A4 size thermal printer ❖ Alphanumeric keyboard with rubber keys ❖ Rechargeable Battery backup for at least 7 hrs ❖ Built in Memory of 250 patient ECG or more ❖ Unit should store atleast 10 minute or more continuous ECG ❖ Should have QT correction by Bazett, Fredericia, Fremingham or Hodges ❖ Should be upgradeable to Culpit Artery criteria to detect probable site of occlusion ❖ Accessories to be supplied with each machine: <ul style="list-style-type: none"> ❖ 10 lead patient cable with banana plugs – 1 no. ❖ extremity electrodes – 1 set ❖ Valid IEC certificate should be submitted 	1	No.	
TOTAL				
APPLICABLE GST @18%				
TOTAL AMOUNT IN Rs.				

I agree to supply the scheduled items at the rates quoted by me as above.

SIGN & SEAL OF THE BIDDER