

MINUTES OF MEETING

MEETING TITLE	PRE-BID MEETING FOR SUPPLY, INSTALLATION AND COMMISSIONING OF DIGITAL RADIOGRAPHY SYSTEM. Dated: 02.08.2025
DATE	08.08.2025
TYPE OF MEETING	VIRTUAL (GOOGLE MEET)- https:// meet.google.com/hpk-bkwt-kbj
MEETING DURATION	12.00 PM – 01.00 PM

PARTICIPANTS DETAILS

HLL	Mr. Joseph Savy K J (AVPi/c & BH (HCS))
	Mr. Jayakumar A (DVP - HCS)
	Mr. Prashant V P, (DGM (F-HCS))
	Mr. Narendranath K R (DGM- Biomedical)
	Mr. Hariprasad A P (SM -HCS)
	Mrs. Aswathy Dinesh.T , (M - HCS)
Prospective bidders participants	Mr. Sujit, Allengers Medical Systems

Pre bid meeting against the Tender floated for Supply, Installation and Commissioning of Digital Radiography System at Hindlabs at Government Taluk hospital Fort Kochi, Kerala on 08.08.2025 at 12.00 AM. Prospective bidders were given the option to participate either in person or through the online link provided in the tender document. During the meeting, the bidders raised various queries and were advised to submit the same in writing via email to hcstenders@lifecarehll.com for further consideration. It was also informed that all queries received would be examined by the HLL team and clarifications/ responses would be published subsequently as the official Pre-Bid Minutes.

In continuation of the Pre-Bid meeting, the following queries have been received from prospective bidders:

SNO	NAME OF THE FIRM	RFP PAGE NUMBER	CLAUSE (TENDER REF.)	QUERY	HLL REPLY
1	Allengers Medical Systems Limited	SECTION F Technical Specification 1.a.TECHNICAL SPECIFICATIONS OF DIGITAL RADIOGRAPHY SYSTEM WITH SINGLE FLAT PANEL DETECTOR	D. X-RAY TABLE: 4. Should have cassette tray and should moved across the length of the X-ray images of various body parts.	We request you to delete the asked specification as Mobit type 4-way floating tabletop will be provided which don't include cassette tray.	No change in tender conditions
			E. FLAT PANEL DETECTOR: 6. Detector resolution should be more than 3.6 Ip /mm or less.	We request you to amend it as Detector resolution should be 3.5 lp/mm or more for healthier & wider participation in the tender.	Tender term is revised as "6. Detector resolution should be 3.5 lp/mm or more "
			We request kindly add the following certification for better product quality and standard:	<p>The offered Unit should be European CE Certified having Notified body number registered in European Commission</p> <p>Company should be ID 13485 Certified (Indian Certification for Medical Devices) Based on the international Harmonized Standard of "Quality Management System for Medical Devices plus additional requirements specified under the scheme.</p> <p>The X Ray Generator, Detector and Console Software should be from the same manufacturer for seamless connectivity and consistent image quality The company should be ISO 9001:2015 certified with manufacturing, design of Flat panel detectors & X-Ray Generators. The model of the detector mentioned on it.</p> <p>We request you to conduct physical demonstration of the same or similar type equipment installed nearby or online/virtual demonstration of the asked specifications before the technical bid opening this will ensure that all technically qualified bidders are meeting the asked tender specifications.</p> <p>The X-Ray Generator should be of indigenous make if the generator is</p>	No change in tender conditions

				imported from other Count complete specifications of data sheet from the principal manufacturer with an undertaking from then stating pares/parts of the generator would be made available for next 10 years to be attached with BID.	
--	--	--	--	---	--

This minute of pre-bid meeting shall form the part of bid document as Addendum/Amendment. All the bidders must attach a copy of this MoM, sealed and signed by authorized signatory, which indicates that all the points in this MoM is agreed by the bidding agency.