Amendment No. 7

Date: 27.02.2015

Sub: Amendment to Tender Enquiry Document for extension of due date for submission of bids.

Ref: NIT No.: HLL/PCD/GNCTD/27/GBPH/14-15 dated 30.12.2014 read with its amendment no. 1, 2, 3, 5, 6 and 7 dated 28.01.2015, 29.01.2015, 29.01.2015, 09.02.2015, 18.02.2015 & 18.02.2015 respectively.

The following changes are incorporated to the referred NIT for following Tender IDs only.

SECTION I NOTICE INVITING TENDERS (NIT)

Existing:

SI. No.	Tender ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	cubmission of	Date & time of opening of tender
6	2014_HFWD _74190_6	Portable colour Doppler Ultrasound	2	1,000	80,000	06-01-2015 11:00 AM	02-03-2015 06:00 PM	03-03-2015 02:00 PM	03-03-2015 02:30 PM
7	2014_HFWD _74190_7	Anesthesia Workstation with Multiparameter Monitor	6	3,000	3,60,000	06-01-2015 11:00 AM	02-03-2015 06:00 PM	03-03-2015 02:00 PM	03-03-20150 2:30 PM
8	2014_HFWD _74190_8	3 Tesla MRI Machine	1	5,000	28,00,000	06-01-2015 11:00 AM	02-03-2015 06:00 PM	03-03-2015 02:00 PM	03-03-20150 2:30 PM

Read as:

SI. No.	Tender ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	submission of	Date & time of
6	2014_HFWD _74190_6	Portable colour Doppler Ultrasound	2	1,000	80,000	06-01-2015 11:00 AM	09-03-2015 06:00 PM	10-03-2015 02:00 PM	10-03-2015 2:30 PM
7	2014_HFWD _74190_7	Anesthesia Workstation with Multiparameter Monitor	6	3,000	3,60,000	06-01-2015 11:00 AM	09-03-2015 06:00 PM	10-03-2015 02:00 PM	10-03-2015 2:30 PM
8	2014_HFWD _74190_8	3 Tesla MRI Machine	1	5,000	28,00,000	06-01-2015 11:00 AM	09-03-2015 06:00 PM	10-03-2015 02:00 PM	10-03-2015 2:30 PM

Section – VII Technical Specification

Schedule No. 7 Anesthesia Workstation with Multiparameter Monitor

Existing Specification:

Para 2.4: The system should have minimum of 60 min battery backup through UPS for the entire unit **Read as:**

Para 2.4: The system should have minimum 60 minutes power backup by **integrated rechargeable** battery or UPS

Existing Specification:

Para 6.1:- The work station should have integrated anaesthesia ventilator system for neonates to adult with display of parameters and control on integrated **touch screen** TFT display.

Read as:

The work station should have integrated anaesthesia ventilator system for neonates to adult with display of parameters on integrated colour TFT monitor.

Existing Specification:

Para 7.6: SPO2 measurement by nellcor or masimo or **equivalent**

Read as:

7.6 should read "SpO₂ measurement by Nellcore or Masimo technology".

The prospective bidders shall submit their online bids as per the revised schedules at https://govtprocurement.delhi.gov.in

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- 1. Prospective bidders are advised to ensure the validity of their EMD as per this revised schedule.
- 2. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of tenders.