# **AMENDMENT No.2**

Date:17.08.2015

Tender Enquiry No.: HLL/PCD/IMMU-06/15-16 dated 03.08.2015
Items of Procurement: Cold Chain Equipments under routine Immunisation

#### **Section IV: General Condition of Contract:**

1. Page no.28 Clause no. 11

# **Existing**

If the equipment is not commissioned and handed over to the consignee within 3 months, the insurance will be got extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the consignee. In case the delay in the installation and commissioning is due to handing over of the site to the supplier by the consignee, such extensions of the insurance will still be done by the supplier, but the insurance extension charges at actual will be reimbursed.

### Read As

### **Deleted**

2. Page no. 30 Clause no.15.5

## **Existing**

In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the warranty for the rectified/replaced goods shall be extended to a further period of twenty four (24) months from the date such rectified / replaced goods starts functioning to the satisfaction of the purchaser.

# **Read As**

In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the warranty for the rectified/replaced goods shall be extended to a further period of <u>Twelve (12)</u> months from the date such rectified / replaced goods starts functioning to the satisfaction of the purchaser.

#### **SECTION II: General Instructions to Tenderer**

3. Page no. 15, Clause 19.2

# **Existing**

The Tenderer who are currently registered and, also, will continue to remain registered during the tender validity period with Directorate General of Supplies & Disposals or with National Small Industries Corporation, **New Delhi** for the specific goods as per tender enquiry specification shall be eligible for exemption from EMD. Vague stipulations in the Registration Certificate such as "to customers' specification" etc. will not be acceptable for exemption from furnishing of earnest money. In case the Tenderer falls in these categories, it should furnish copy of its valid registration details (with DGS&D or NSIC, as the case may be).

### Read As

The Tenderer who are currently registered and, also, will continue to remain registered during the tender validity period with Directorate General of Supplies & Disposals or with National Small Industries Corporation, for the specific goods as per tender enquiry specification shall be eligible for exemption from EMD. Vague stipulations in the Registration Certificate such as "to customers' specification" etc. will not be acceptable for exemption from furnishing of earnest money. In case the Tenderer falls in these categories, it should furnish copy of its valid registration details (with DGS&D or NSIC, as the case may be).

**4.** Page no 15, Clause no. 19.3

## **Existing**

The earnest money shall be denominated in Indian Rupees <u>or equivalent</u> currencies as per GIT clause 12.2.

### Read As

The earnest money shall be denominated in Indian only.

#### **SECTION XV: Bank Guarantee Form**

**5.** Page 62

### **Existing**

This guarantee shall be valid up to 66 (Sixty Six) months from the date of Notification of Award i.e. up to ----- (indicate date)

#### **Read As**

This guarantee shall be valid up to <u>30 (Thirty) months for Stabilizers and 18 (Eighteen) months Cold Box, Vaccine Carrier, Thermometer and Ice Pack from the date of Notification of Award i.e. up to ------ (indicate date)</u>

#### **SECTION VII: Technical Specifications – Vaccine Carrier**

6. Page 49, Clause 3.8:

#### **Existing:**

Leakage Test: No leaks when 80 Kg lateral force is applied.

Features: Effective reinforcements to restrain walls against swelling. Removable cap for filling ;cap to have effective internal water seal to resist **80 kg** lateral force with no leakage.

#### Read As

Leakage Test: No leaks when **40 Kg** lateral force is applied.

Features: Effective reinforcements to restrain walls against swelling. Removable cap for filling ;cap to have effective internal water seal to resist **40 kg** lateral force with no leakage.

All other	Terms and	Conditions of	of Tender	<b>Enquiry</b>	<b>Document</b>	will remain
unchange	d.					

------XXXXXX------