

## Technical Amendment no. 10

Date: 18.06.2015

Subject: Amendment no. 10 to the Tender Enquiry Document

Ref: Tender Ref: HLL/PCD/PMSSY-II/NAGPUR/05/14-15 dated 26.03.2015

### Technical Specifications

#### Item Sl. No.15: Automated Hematology Analyzer —5 part differential —very high Throughput with accessories Bar code Reader and Software upgrades for 10 years

##### 1. Existing Specifications (after amendment)

Para 25: linearity of WBC :  $0.02-400 \times 10^3/\text{ul}$ , PLT- $5-3500 \times 10^3/\text{ul}$ , Retics: 0.2-24.5%, HB: 0.0-22.5 g/dl

Read as

Para: linearity of WBC :  $0-150 \times 10^3/\text{ul}$  & PLT- $0-1900 \times 10^3/\text{ul}$ , Retics: 0.2-24.5%, HB: 0.0-22.5 g/dl

##### 2. Existing Specifications (after amendment)

Para 28: Should be able to integrate with the optional slide make and strainer

Read as

Para 28: DELETED

##### 3. Existing Specifications (after amendment)

Para 27: Instrument must have the capability to run CSF & other body fluids & should be FDA approved

Read as

Para 27: Instrument must have the capability to run CSF & other body fluids. **Offered model should be USFDA or European CE certified.**

##### 4. Added Para: Bidder should quote price for all required consumables and reagents for 10000 samples all parameters and 1000 samples for reticulocyte to be supplied over a period of time. This price shall be considered for ranking purpose.