# Amendment No. 1

02.02.19

#### Sub: Amendment to the Bidding Document

#### Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/39/18-19 dated 09.01.2019

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

#### SECTION - VII

#### **TECHNICAL SPECIFICATION AND GENERAL POINTS**

#### A. TECHNICAL SPECIFICATION:

### Item No:- 1. (Rfx no. 3000003729)

#### Full HD Video Endoscope System (2 nos.) with Endoscopic Ultrasound System (1 no.)

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 5. Full	Automatic IRIS Control &	Automatic IRIS Control &
	HD video	automatic white balance	white balance
	processor		
	Page 49		

### Item no :- 2. (Rfx 3000003730).

### Full HD Video Bronchoscope System with Endobronchial Ultrasound System

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para A Page 52	The flexible bronchoscope should be supplied complete with light source and trolley and <b>minimum 15</b> " <b>LCD</b> monitor	The flexible bronchoscope should be supplied complete with light source, trolley and monitor
2	Under Para2 Linear probe Page 53	Depth of field : Approximately 2-50 mm	Depth of field : Approximately <b>3-50 mm</b>

### Item No:- 4. (Rfx 3000003732)

#### **<u>Rigid Cysto-Resectoscope</u>**

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 1. C Page 56	Should have Robust ceramic beak with guarantee during warranty period	Should have Robust ceramic beak to be covered under warranty. <b>Replacement if</b> needed to be done free of cost during the warranty period
2	Para 7 Page 57	ESU with 300W or more <b>(of</b> <b>any reputed make)</b> output power	ESU with 300 W or more output power. ESU should be of same make of cysto- resectoscope
3	Para 7 f Page 57	All compatible accessories for therapeutic & diagnostic hysteroscope work should be supplied along with ESU unit	All compatible accessories for cysto-resectoscopy work should be supplied along with ESU unit
4	Para 7.e.iv Page 57	Disposable patient plate paediatric – 10 nos.	Deleted
5	BOQ Sl.no.33 Page 58	Disposable patient plate pediatric	Deleted

### Item No:- 5. (Rfx 3000003733)

### Micro Vascular Instruments Set

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 5 Page 59	Dedicated Slotted sterilization container with lock and silicon mat	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.
2	Para 8 Page 59	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments	-

	should be manufacturer	from sa	me manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Declaration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
3	Added BOQ Ite at Sl.no. 22:	m	Sterilization container as per technical specifications - 1 set

# Item.No:- 6. (Rfx 3000003734)

## Major Vascular Instruments Set

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 7 Page 61	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Declaration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
2	BOQ Sl.no. 64 Page 63	Reynolds' vascular forceps small	<b>Reynolds' or similar</b> vascular forceps small
3	BOQ Sl.no. 65 Page 63	Reynolds' vascular forceps medium	<b>Reynolds' or similar</b> vascular forceps medium
4	BOQ Sl.no. 66 Page 63	Reynolds' vascular forceps long	<b>Reynolds' or similar</b> vascular forceps large

5	Added Para:	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.
6	Added BOQ Item:	Sterilization container as per technical specifications - <b>1 set</b>

# Item.No:- 7. (Rfx 3000003735)

### **Thoracic Vascular Instruments Set**

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 7 Page 64	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Declaration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
2		Added Para:	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.
3		Added BOQ Item:	Sterilization container as per technical specifications - 1 set

### Item .No:- 8. (Rfx 3000003736)

# **Plastic Surgical Instruments Set**

S1. No.	Reference to the bidding document	Existing Specification of the bidding document	Amended as
1	Para 7 Page 64	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer	Bidder should quote for all the instruments of the set. At least 80% of the quoted instruments should be from same manufacturer. If a bidder is offering 80% or more instruments from a particular Manufacturer, then Declaration from that manufacturer should be submitted stating that the rest of instruments are not being manufactured by that manufacturing firm.
2		Added Para:	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.
3		Added BOQ Item:	Sterilization container as per technical specifications - 1 set

All other contents of the Bidding Document including terms & conditions remain unaltered.