

MINUTES OF MEETING

MEETING TITLE	PRE-BID MEETING FOR SUPPLY, INSTALLATION AND COMMISSIONING OF X- RAY MACHINES (4 NOS.). (Tender no. : HLL/HCS/Tender-GeM/2025-26/16 DTD: 04.02.2026- GEM/2026/B/7191879Dated: 04-02-2026)
DATE	11.02.2026
TYPE OF MEETING	PHYSICAL/ VIRTUAL (GOOGLE MEET)- meet.google.com/yum-xcpz-fbx
MEETING DURATION	11.00 AM – 12.00 PM

PARTICIPANTS DETAILS

HLL Lifecare Limited	Mr. Binu Thomas (VP-PS & GH- HCS)
	Mr. Jayakumar A (DVP - HCS)
	Mr. Narendranath (DGM-Biomedical)
	Mr. Hariprasad A P (SM -HCS)
Prospective bidders participants	Mr. Viswadev, Dynamic ventures, Thiruvananthapuram

Prebid meeting against the Tender floated for pre-bid meeting for supply, installation and commissioning of X- RAY MACHINES (4 NOS.) at Hindlabs at 4 locations was conducted on 11.02.2026 at 11.00 AM at corporate and head office of HLL Lifecare Limited at Thiruvananthapuram. Prospective bidders were given the option to participate either in person or through the online link provided in the tender document. During the meeting, the bidders raised various queries and were advised to submit the same in writing via email to hcstenders@lifecarehll.com for further consideration. It was also informed that all queries received would be examined by the HLL team and clarifications/ responses would be published subsequently as the official Pre-Bid Minutes. In continuation to the Prebid meeting, the prospective bidders have also mailed their respective queries. Also we have representations were received through GeM platform also. Details of response to the queries are as below.

SNO	Representation by Bidder	HLL Reply
1	Epsilon Healthcare solutions Pvt Ltd- Clarify the type of tube stand along with machine for healthy and wider participation	As per clause V, X-ray tube stand shall be Floor self standing model

REPRESENTATIONS IN GEM		
1	Kindly Clarify tube stand column you required in specification	As per clause V, X-ray tube stand shall be Floor self standing model
2	Column rotation: 360°; from +180° to -180° in 90° increments. We requests you to amend it as Column rotation: +/-90 degree (180degree) for wider participation	Column rotation: 180°; from +90° to -90°
3	We request kindly add the following certification for better product quality and standard: • The unit should be European CE certified from notified body number. • We request you to conduct physical demonstration of the same or similar type equipment installed nearby or online/virtual demonstration of the asked specifications before the technical bid opening this will ensure that all technically qualified bidders are meeting the asked tender specifications. • The X-Ray Generator should be of indigenous make if the generator is imported from other Country, complete specifications of data sheet from the principal manufacturer with an undertaking from them stating spares/parts of the generator would be made available for next 10 years to be attached with the BID	No change
4	Delivery 20 days You are requested to amend the delivery as 60 Days after the receipt of procurement permission of e-LORA from the hospital authorities.	No Change
5	Payment as per the terms shall be released by the Government Authority upon satisfactory completion of the contract and fulfillment of all terms of the Purchase Order by the Supplier. If the payment due is not released within 90 days from the date it becomes due, interest shall be payable on the delayed amount at the prevailing bank rate, provided the delay is not attributable to the supplier.	No Change

This minutes of pre-bid meeting shall form part of bid document as Addendum/Amendment. All the bidders must attach a copy of this MoM, sealed and signed by authorized signatory, which indicates that all the points in this MoM is agreed by the bidder.