Amendment No. 1

Date: 14/03/2014

Subject: Amendment to the Tender Enquiry Document

Ref: NIT ref.: HLL/PCD/GNCTD/02/13-14 dated 12/02/2014

Tender ID in GNCTD e-Procurement portal: 2014_HFWD_52731

The following changes have been incorporated in the referred tender.

SECTION I

Existing:

| Description | Qty. | Tender Fees (Rs.) | EMD Amount (Rs.) | Date & time of Prebid meeting | Date & time of closing of online tender | Closing date & time for submission of physical Tender | Date & time of opening of tender |
|--------------------------|------|----------------------|------------------------|-------------------------------------|---|--|----------------------------------|
| Haemodialysis Machine | 4 | 1,000 | 96,000 | 20-02-2014 11:00 AM | 17-03-2014 6:00 PM | 18-03-2014 1:00 PM | 18-03-2014 2:00 PM |

Amended as:

| Description | Qty. | Tender Fees (Rs.) | EMD Amount (Rs.) | Date & time of Prebid meeting | Date & time of closing of online tender | Closing date & time for submission of physical Tender | Date & time of opening of tender |
|--------------------------|------|----------------------|------------------------|-------------------------------------|---|--|----------------------------------|
| Haemodialysis Machine | 4 | 1,000 | 96,000 | 20-02-2014 11:00 AM | 02-04-2014 6:00 PM | 03-04-2014 1:00 PM | 03-04-2014 2:00 PM |

Section – VII Technical Specifications Haemodialysis Machine

| Sl. No | Tender Page No. | Existing Specification | Read As |
|-----------|---|---|--|
| 1 | Page No: 43 Section – VII Technical Specification | Blood temperature and Volume monitors are desirable | Blood temperature and Volume monitors |
| 2 | Page No: 43 Section – VII Technical Specification | Battery Backup for power failure should be present for at least 30 minutes. | Internal Battery Backup for power failure should be present for at least 15 minutes. |

All other contents of the tender enquiry read with its earlier amendment including terms & conditions remain unaltered.