

21st January 2019**Amendment No. 03**

SUB: Amendment to the referred Tender Enquiry No. HITES/PCD/PMSSY-IV/AIIMS/10/18-19 dated 13.12.2018 read with Amendment No. 1 and Amendment No. 02 dated 03.01.2019 and 09.01.2019 respectively.

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION I
NOTICE INVITING TENDER

Existing:

Sl. No.	Description	Schedule
d.	Closing date & time for submission of online bids	22.01.2019, 12:00 PM
c.	Closing date & time for submission of tender processing fee and EMD in physical form*	22.01.2019, 02:00 PM
e.	Time and date of opening of online bids	22.01.2019, 02:30 PM

Read as:

Sl. No.	Description	Schedule
d.	Closing date & time for submission of online bids	05.02.2019, 12:00 PM
c.	Closing date & time for submission of tender processing fee and EMD in physical form*	05.02.2019, 02:00 PM
e.	Time and date of opening of online bids	05.02.2019, 02:30 PM

SECTION VII
TECHNICAL SPECIFICATIONS

The revised Technical Compliance Sheet **TC_3000003618_v1.1 of Sch. No. 2 (MRI 1.5 T)** & **TC_3000003624_v1.1 of Sch. No. 8 (Digital Mammography System)** is being uploaded with the following changes as under:

Name of Item :- MRI 1.5 T (Rfx No. 3000003618)		
Tender Page & Para	Existing	Read as
E-tender Portal	Existing Para of TC_3000003618_v1.0	Amended Para of TC_3000003618_v1.1
Para 8 j , Page 55	b. Flex coils (minimum 8 channel) – 2nos. (One large FOV and one small FOV)	b) Flex Coil(4 channel or more)- 2nos. (One large FOV and one small FOV)
Page 64 Para 7	BOQ	BOQ
	ACCESSORIES Multiparameter Vital Signs Patient Monitor and slave monitor both, Compatible with 1.5 T MRI . qty 1 set	ACCESSORIES MRI compatible Multiparameter Vital Signs Patient Monitor of 5000 Gauss/1.5m from isocentre Compliance and One Slave monitor -qty 1 set
		1.5 T MRI compatible syringe pump qty- 1no. to be supplied as per BOQ

Digital Mammography System(3000003624)		
Tender Page & Para	Existing	Read as
E-tender Portal	Existing Para of TC_3000003624_v1.0	Amended Para of TC_3000003624_v1.1
Para 6a, page 84	Beam filter: 1 No or more. (Mo/Ag/Al) with Rh	Beam filter: 1 No or more. (Mo/Ag/Al/Rh)
Page 12 K, page 85	Dry view camera: 600 DPI or more, with 300 films each (size 18x24cm and 24x30cm)	Dry view camera: 508 DPI or more, with 300 films each (size 18x24cm and 24x30cm)

It is strictly advise to upload the revised technical sheet.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids
- ii. Prospective bidders are advised to ensure the validity of their EMD as per this revised schedule.