

Amendment No. 18

Date: 19/02/2016

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/07/15-16 dated 28/10/2015 and subsequent amendments

The following changes have been incorporated in the referred NIT.

Section VII **Technical Specification**

Item Sl. No. 31 **VENTILATORS HIGH END**

1. Existing Para 4.e: Advanced mode like Pressure Regulated volume control **mode and volume support mode.**

Read as: Advanced mode like Pressure Regulated volume control mode

2. Existing Para 5.i: **Inspiratory and Expiratory flow and** pressure Trigger Sensitivity.

Read as: Should have flow and pressure trigger sensitivity.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids