

AMENDMENT NO.2 Dated 15.07.2026

Ref IFB No: HLL/PSD/2026-27/TENDER/26 Dt:11.06.2026

GeM-Tender ID: GEM/2026/B/7653104 DT:11.06.2026

The following amendments have been incorporated in the ATC document of the above mentioned tender invited for Supply of Digital Mammography Machine at Various District Hospitals across the country

SECTION F - TECHNICAL SPECIFICATION

Sl.No.	Existing Specification	Amended Specification
	1. X-RAY GENERATOR – High Frequency	
1.	c. mA = 150mA or more	c. mA = 100mA or more
2.	d. KV Range = 23KV to 49KV or higher in 1 KV increment.	d. KV range required is at least 23 KV lowest limit and at least 49 KV higher limit (1KV increment is preferred)
	3. FLAT PANEL DETECTOR	
3.	d) pixel pitch = Atleast 85µm	d) Pixel Pitch: 100µm (microns) or less in both 2D and 3D
	Added clause 13	
4.	The system should be onsite upgradable to Contrast Enhanced Mammography facility & CEDM Biopsy if required in the same quoted model in future. Undertaking to be submitted. The rate for the same shall be offered separately and the rate shall be valid for three years. This rate will not be part of financial ranking	The system should be onsite upgradable to Contrast Enhanced Mammography facility if required in the same quoted model in future. Undertaking to be submitted. The rate for the same shall be offered separately and the rate shall be valid for three years. This rate will not be part of financial ranking

All other technical specifications of the original tender document remain unchanged

COMMERCIAL TERMS & CONDITIONS

Sl.No.	Existing Specification	Amended Specification
	Section B; E.3	
1.	Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 20% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.	Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 20% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Private institutions. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

All relevant clauses of the tender document are to be read in accordance with the above changes and in case if any bids were already submitted, if required you may resubmit the documents in compliance of the above amendment.

All other terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & GH (HCS)