

AMENDMENT NO.1 Dated 17.09.2025

REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 13RE Dated 09th September 2025

Amendment made to the details given in Tender Document.

REF NO: HLL/CHO/PSD/HCS/2025-26/TENDER 13RE Dated 09th September 2025

CPP TENDER ID : 2025_HLL_244874_2

The following amendment has been incorporated in the above mentioned **E-TENDER FOR SUPPLY OF SEMI AUTOMATED COAGULATION ANALYSER FOR HINDLABS DIAGNOSTIC CENTER, GURUGRAM ON REAGENT RENTAL BASIS.**

The following amendment / Inclusion have been incorporated to the bid document for the above tender:

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 3 - Should use optical/mechanical detection principle and there should not be any interference from lipemic /icteric /hemolytic samples.

MAY BE READ AS

Point No.3 – Should use optical/ Mechanical detection principle

All relevant clauses of the Tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & Group Head (HCS)