

Technical Amendment- 05

TENDER REF: HLL/PCD/PMSSY-II/NAGPUR/09/15-16

Plasma Sterilizer

1. Existing Amendment:

Added Para of Amendment 4: Plasma sterilizer should be **square or rectangular** in shape

Read as: Plasma sterilizer should be **square or rectangular or circular** in shape

2. Existing specifications

Para: 3.4: The Sterilizer should have usable Volume of 100-120Litres

Read as

Para 3.4: The Sterilizer should have usable Volume of 100-120 Litres. **Undertaking from Principal manufacturer is to be enclosed along with supporting documents.**