

Amendment No. 1

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/10/18-19 DATED 13.12.2018.

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII
Technical Specification**

**Schedule No. 01
CT 128 Slice (Rfx No. 3000003617)**

Tender Page & Para	TENDER SPECIFICATION	READ AS
Para 3b , page 48	X-Ray Tube Anode Heat Storage Capacity of at least 7.0 MHU or direct cooling tube	X-Ray Tube Anode Heat Storage Capacity of at least 6.0 MHU or direct cooling tube.
Para 10, page 49	Workstations & Server: A multimodality client server architecture based solution with minimum concurrent 24000 slices rendering capacity, with 64GB RAM with storage of minimum 2TB and Additional online storage of 10 TB on the server. Client hardware specification- 2 no.s Workstations with two licences: dual quad core processor, 16 GB RAM, 1TB hard drive, DVD Writing with medical grade monitor of minimum 3 MP resolution & 3 button mouse	Workstations & Server: A multimodality client server architecture based solution with minimum concurrent 24000 slices rendering capacity, with 32GB or more RAM with storage of minimum 2TB and Additional online storage of 10 TB on the server. Client hardware specification- 2 no.s Workstations with two licences: dual quad core processor, 16 GB RAM, 1TB hard drive, DVD Writing with medical grade monitor of minimum 3 MP resolution & 3 button mouse
Para 10, Page 49	Whole organ (brain & body) perfusion CT.	Whole organ/ wide volume (brain & body) perfusion CT.
Para 13, page 50	Training for a period of Two Weeks.	Onsite Training for a period of Two Weeks.
Para f, page 51	Civil work Active and passive room shielding for magnetic, fringe field should be provided as per the requirement of the equipment.	Deleted
Para 1, page 51	Plumbing work All water pipes and fittings shall be of high density polythene of approved and standard make. The gratings shall be brass chrome plated. All plumbing accessories should be of standard make	Deleted
Para 2, Furniture .C page 52	Furniture: c.Cupboard with laminate door shutters for storage of spare parts and accessories and records as per requirement. – 3 NO.S	Furniture: c.Cupboard with laminate door shutters for storage of spare parts and accessories (approx size of 6'X3') and records as per requirement. – Qty 3 NO.S

Para 2, Furniture .h page 52	Furniture: h.Changing rooms should have change lockers and dressing table	Deleted
Para 2, Furniture. J page 52	Any other furniture item as per requirement	Deleted
Pg 49,Para 7a	a. Detector- Capable of acquiring 64 slices per 360 degree of rotation.	Deleted
		Added Para: Should have Continuous spiral length of 100cm or more
Pg 49,Para 7a	3-D laser lights for positioning.	3-D laser lights for positioning. (Inbuilt)
Page 51, para f	Patient waiting area	Deleted
	BOQ	BOQ
1	CT Scanner 128 SLICE CT ,as specified	CT Scanner 128 SLICE CT ,as specified qty 1 no.
2	Cardiac package (including Hardware & Software)	Cardiac package (including Hardware & Software) qty 1 no.
3	Servers: as specified	Servers: as specified qty -01 no
4	Workstation Nodes: as specified	Workstation Nodes: as specified qty - 2nos.
5	Dry Chemistry camera : as specified	Dry Chemistry camera : as specified -qty 1 no.
6	Lead Glass of 120 x 90 cm.	Lead Glass of 120 x 90 cm. qty 1 no
7	UPS with Maintenance free batteries	UPS with Maintenance free batteries qty 1 set
8	Dual Head Pressure Injector	Dual Head Pressure Injector qty 1 no.
9	Pressure Injector Syringes	Pressure Injector Syringes qty-50nos.
10	Tubings for Pressure Injector	Tubings for Pressure Injector qty - 200sets
11	Multi Para monitor: as specified	Multi Para monitor: as specified - qty 1no.
12	ULTRA LIGHT WEIGHT lead aprons	ULTRA LIGHT WEIGHT lead aprons with hangers qty-4nos.
14	Lead apron stand	Lead apron stand qty 1no.
15	Thyroid Shields	Thyroid Shields qty - 2 no
16	Gonadal Shields	Gonadal Shields qty - 2 nos.
	Components of Site Modification Work	Components of Site Modification Work :
1	Civil works	Civil works 1200sqft
2	Electrical work	Electrical work 1200sqft
3	Public health (plumbing and sanitary fittings).	Public health (plumbing and sanitary fittings). 1200sqft
4	Air Conditioning	Air Conditioning 15 TR
5	Interior Furnishing & Furniture	Interior Furnishing & Furniture 1200 sqft
	Miscellaneous items	Miscellaneous items
	Furniture:	Furniture:
1	Revolving chairs height adjustable, medium-back with hand-rest	Revolving chairs height adjustable, medium-back with hand-rest qty - 4 Nos.
2	Chairs for patient waiting area – Three seater (chrome	Chairs for patient waiting area – Three seater (chrome

	plated). -	plated). - 10 NO.S
3	Cupboard with laminate door shutters	Cupboard with laminate door shutters for storage of spare parts and accessories (approx size of 6'X3') and records as per requirement. – qty 3 Nos.
4	Drug trolleys for patient preparation area.	Drug trolleys for patient preparation area. Qty -1 no.
5	Patient trolley with rubber foam mattress	Patient trolley with rubber foam mattress qty 2no.
6	Tables for Workstation and Radiologist .	Tables for Workstation and Radiologist . - qty-2 nos.
7	Changing rooms (with change lockers and dressing table).	deleted
8	Dustbins	Dustbins qty-10 nos.
9	Room Signage	Room Signage 1 LS
10	Venetian Blinds	Venetian Blinds 1 LS
	Miscellaneous:	Miscellaneous:
1	LED X-ray Film viewer	LED X-ray Film viewer qty - 2Nos.
2	Cabling of Network (LAN) connectivity for camera system, console system, workstation and computers etc	Cabling of Network (LAN) connectivity for camera system, console system, workstation and computers etc 1 LS
3	Dry chemical power type fire extinguisher of 5kgs capacity	Dry chemical power type fire extinguisher of 5kgs capacity qty 3 Nos.
		Added para Broad band Connection : for remote service of CT System qty -1 nos

Schedule No. 02
MRI 1.5 T (Rfx No. 3000003618)

Tender Page & Para	TENDER SPECIFICATION	READ AS
Para 8f, page 55	f.Body Array/Matrix coil with 18 – 32 channels with at least 38 cm Z- axis coverage for imaging of abdomen, angiograms and heart	f.Body Array/Matrix coil with 28 channels or more with at least 38 cm Z- axis coverage for imaging of abdomen, angiograms and heart with one or two body coils
Para 8 i , Page 55	Bilateral Breast Coil with at least 16 channels.	One Bilateral Breast Coil with at least 16 channels. OR Two Bilateral Breast Coil with at least 8 channels.
Para 8 j , Page 55	SHOULDER coils a. Dedicated Shoulder coil – Multi channel - 1 No b. Flex coils – 2nos. (One large and one small)	SHOULDER coils a. Dedicated Rigid Shoulder coil – Multi channel - 1 No b. Flex coils (minimum 8 channel) – 2nos. (One large FOV and one small FOV)
Para 9r , page 56	Perfusion imaging of brain (both contrast and T1 perfusion) Iron quantification software for liver heart, etc .T1 and T2 relaxation values	Perfusion imaging of brain (both contrast and T1 perfusion) Iron and fat quantification software for liver heart, etc .T1 and T2 relaxation values
Para 9s , page 56	Susceptibility weighted imaging (i.e.SWI) /FMRI / Venous BOLD imaging.	Susceptibility weighted imaging (i.e.SWI) /Venous BOLD imaging.

<p>Para 8 , page 57</p>	<p>Workstation One server with 2 node with concurrent licenses to be supplied with the system. Licenses: Two Concurrent license here implies the capability to process all the loaded software to be accessible and usable on all the systems simultaneously without any processing delay. The software should also include a reputed antivirus software of a perpetual type or renewed by the supplier. Hardware: Node: The vendor has to supply the hardware in the form of CPU and Medical grade monitor 18" or more of 2MP resolution. Hardware Server: The server (single/dual configuration) should have image storage capacity of at least 4 Tera bytes, minimum 20,000 concurrent slice processing power and at least 64GB RAM. The server hardware to be included with 21" or more TFT/LCD monitor with dual processor. DICOM 3.0 compatibility and interfacing with other modalities must be possible. The workstation shall have the resolution, software and all functionality of a stand-alone workstation</p>	<p>Workstation One server with 2 node with 2 concurrent licenses to be supplied with the system for all the application. Licenses: two Concurrent license here implies the capability to process all the loaded software to be accessible and usable on all the systems simultaneously without any processing delay. The software should also include a reputed antivirus software of a perpetual type or renewed by the supplier. Hardware: Node: The vendor has to supply the hardware in the form of CPU and Medical grade monitor 18" or more of 2MP resolution. Hardware Server: The server (single/dual configuration) should have image storage capacity of at least 4 Tera bytes, minimum 20,000 concurrent slice processing power and at least 32GB RAM. The server hardware to be included with 21" or more TFT/LCD monitor with dual processor. DICOM 3.0 compatibility and interfacing with other modalities must be possible. The workstation shall have the resolution, software and all functionality of a stand-alone workstation</p>
<p>Para 16.2.i page 58</p>	<p>2. (i) MRI Compatible Dual Syringe Pressure injector : Independent dual-Syringe Pressure injector with following Features; Non-ferrous, automatic syringe size detection, performs single and dual phase contrast injections, provides Saline flush delivery and allows timed contrast delivery Must be compatible with 5, 7.5 &10ml pre-filled contrast syringes and 50 ml syringes for both saline & contrast (20 Nos of 50 ml Syringes with 100 nos. of tube connectors should be provided) Must be able to observe progress of injection and view injection result</p>	<p>2. (i) MRI Compatible Dual Syringe Pressure injector : Independent dual-Syringe Pressure injector with following Features; Non-ferrous, automatic syringe size detection, performs single and dual phase contrast injections, provides Saline flush delivery and allows timed contrast delivery Must be compatible with 5, 7.5 &10ml pre-filled contrast syringes and 50 ml syringes for both saline & contrast (100 Nos of 50 ml Syringes with 100 nos. of tube connectors should be provided) Must be able to observe progress of injection and view injection result</p>
<p>Para 16.2.ii page 58</p>	<p>(ii) (Optional) - MRI Compatible Dual Syringeless Pressure injectors with pump hose : Independent Dual Syringeless Pressure injector with following Features; 5000 gauss Compliant, Non-ferrous, performs single and dual phase contrast injections, provides Saline flush delivery and allows timed contrast delivery. Must be able to observe progress of injection and view injection result. 100 nos. of tube connectors should be provided</p>	<p>Deleted</p>
<p>Para d. I ,Page 61</p>	<p>Plumbing Work All water pipes and fittings shall be of high density polythene of approved and standard make. The gratings shall be brass chrome plated. All plumbing accessories should be of standard make.</p>	<p>Deleted</p>
<p>Para h.iii, page 63</p>	<p>Furniture: Cupboard with laminate door shutters for storage of spare parts and accessories and records as per requirement. – 3 NO.S</p>	<p>Furniture: h.iii.Cupboard with laminate door shutters for storage of spare parts and accessories (approx size of 6'X3') and records as per requirement. – 3 NO.S</p>

Para h.vii , page 63	Furniture : Changing rooms should have change lockers and dressing table	Changing room with 6 lockers and one dressing table to be supplied
Para i) .3, page 63	Miscellaneous: 4 Dry chemical power type fire extinguisher of 5kgs capacity, with initial filling in brand new cylinder with power coated finish, fitted with Gun metal union, high pressure CO2 gas cartridge, discharge hose, wall mounting bracket etc. complete, confirming t IS:2171 of approved make & complete as directed by EIC.	Amended as: 4 Dry chemical power type MRI compatible fire extinguisher of 5kgs capacity, with initial filling in brand new cylinder with power coated finish, fitted with Gun metal union, high pressure CO2 gas cartridge, discharge hose, wall mounting bracket etc. complete, confirming t IS:2171 of approved make & complete as directed by EIC.
Page 58 para 5	5. One MRI compatible Multiparmeter Vital Signs Patient Monitor of 5000 Gauss Compliance in MRI Room and One Slave monitor in console room with following modules provision to monitor the following	5. One MRI compatible Multiparmeter Vital Signs Patient Monitor of 5000 Gauss/ 1.5m from isocentre Compliance in MRI Room and One Slave monitor in console room with following modules provision to monitor the following; Monitor should be atleast 12" display
		Added para: 1 IBP
	h. Temperature (adult and pediatric	Deleted
Page 59 para 7	MRI compatible syringe pump – 2 Nos	1.5T MRI compatible syringe pump – 2 Nos
		Added para: MRI Room oxygen deficiency level monitor - 1 no MRI compatible WHEEL CHAIR - 2 NOS
page 60, pare f.	patient waiting area	Deleted
		Added para: Cartilage Assesment software to be included as standard .
	BOQ	BOQ
1	Whole body 1.5 Tesla Magnetic Resonance Imaging system - 32 channels RF system ; as specified	Whole body 1.5 Tesla Magnetic Resonance Imaging system - 32 channels RF system ; as specified - Qty 1 no.
2	System Body Coil - Quadrature	System Body Coil - Quadrature - Qty 1no
3	Neuro-vascular Coil with 16 or more channels OR Head / Neck neuro-vascular imaging Coil	Neuro-vascular Coil with 16 or more channels OR Head / Neck neuro-vascular imaging Coil - qty 1 no
4	Spine Array/Matrix Coils with atleast 32 channels	Spine Array/Matrix Coils with atleast 32 channels Qty - 1no
5	Body Array/Matrix coil with 18 – 32 channels	Body Array/Matrix coil with 18 – 32 channels Qty- 1no
6	Dedicated 32 channel Peripheral Angio Coil or 32 channel whole body coil with coverage of minimum 80 cm with max combination of 2 coils - 1 no	Dedicated 32 channel Peripheral Angio Coil or 32 channel whole body coil with coverage of minimum 80 cm with max combination of 2 coils - qty 1 no
7	Bilateral Breast Coil with at least 16 channels	(One Bilateral Breast Coil with at least 16 channels. OR Two Bilateral Breast Coil with at least 8 channels) - 1 set
8	Multi channel (minimum 8 channel) flex loop or rigid type - Large FOV	Multi channel (minimum 8 channel) flex loop Large FOV qty-1no
9	Multi channel (minimum 8 channel) flex loop or rigid	Multi channel (minimum 8 channel) flex loop SMALL

	type - SMALL FOV	FOV qty-1no
10	Shoulder coil: Dedicated Shoulder coil – Multi channel	Shoulder coil: Dedicated Rigid Shoulder coil – Multi channel qty-1 no
11	Dedicated Knee Coil with atleast 15 channels.	Dedicated Knee Coil with atleast 15 channels. qty-1 no.
12	High resolution foot / ankle coil – minimum 8 channel	High resolution foot / ankle coil – minimum 8 channel qty-1 no.
13	Coil Storage Cart	Coil Storage Cart qty-1 no.
14	Server : Thin-client server as per specification	Server : Thin-client server as per specification qty 1no
15	Licenses: Concurrent licenses for Server.	Licenses: Concurrent licenses for Server. Qty-2 nos.
16	Node Hardware: CPU and Medical grade monitor (18 or more; 2 megapixels or more resolution).	Node Hardware: CPU and Medical grade monitor (18 inch or more; 2 megapixels or more resolution). Qty-2nos.
17	Antivirus software of reputed make (perpetual type or license to be renewed by the supplier).	Antivirus software of reputed make for two licenses software (perpetual type or license to be renewed by the supplier) as per specification Qty-1 .nos.
18	Cardiac Package - License	Cardiac Package - License qty-2 nos.
19		Cartilage Assesment software to be included . Qty - 1no.
20		Voice recognition software qty 2 nos.
	ACCESSORIES	ACCESSORIES
1	Storage box for all coils	Storage box for all coils qty 1 no
2	Dual Syringe Pressure injector	Dual Syringe Pressure injector qty-1 no.
3	Dual Syringe Pressure injector syringes	syringes (50 ml) -Dual Syringe Pressure injector qty - 100
4	Dual Syringe Pressure injector syringe connector	connector -Dual Syringe Pressure injector syringe qty - 100
5	MRI Compatible ECG electrodes (disposable)	MRI Compatible ECG electrodes (disposable) - qty-100
6	MRI Compatible Anaesthesia Machine with integrated Ventilator, 2 vaporiser, circle absorber	MRI Compatible Anaesthesia Machine with integrated Ventilator, 2 vaporiser, circle absorber qty-1no.
7	MRI Compatible Multiparmeter Vital Signs Patient Monitor of 5000 Gauss Compliance & Slave monitor	Multiparmeter Vital Signs Patient Monitor and slave monitor both, Compatiple with 1.5 T MRI . qty 1 set
8	MRI compatible syringe pump	1.5 T MRI compatible syringe pump qty- 1no.
9	MRI Compatible sets of Laryngoscope : 4 sizes blades- Neonatal, paediatrics, adult, extra large	1.5 T MRI Compatible sets of Laryngoscope : 4 sizes blades- Neonatal, paediatrics, adult, extra large - qty 1 set
10	MRI compatible Magill forceps : Adult size-	1.5 T MRI compatible Magill forceps : Adult size- qty2 nos.
11	MRI compatible Magill forceps : Paediatric size-	1.5 T MRI compatible Magill forceps : Paediatric size- qty 2 nos.
12	Stylet for endotracheal tube : Adult size	Stylet for endotracheal tube : Adult size qty 3. nos.
13	Stylet for endotracheal tube : Paediatric size	Stylet for endotracheal tube : Paediatric size qty 3nos.
14	MRI compatible Clamps : Either towel clip or artery forceps.	1.5 T MRI compatible Clamps : Either towel clip or artery forceps qty - 2 nos.
15	MRI Compatible IV stands	1.5 T MRI Compatible IV stands qty 2 No
16	MRI compatible suction apparatus	1.5 T MRI compatible suction apparatus qty 2 no.

17	Non-magnetic patient transfer trolleys	Non-magnetic patient transfer trolleys qty 2 no.
18	Metal detectors : Handheld	Metal detectors : Handheld qty 2 nos.
19	Metal detector: Walk-through	Metal detector: Walk-through qty 1 no.
20	Phantoms to be provided for regular QA studies.	Phantoms to be provided for regular QA studies. Qty 1
21	MRI Compatible Dual Syringeless Pressure injectors (Optional)	Deleted
22	Tube connectors for Syringeless Pressure injector (Optional)	Deleted
23	Walk through Metal detector with multiple sensor and multiple location LED (Zone III type) - (Optional)	Walk through Metal detector with multiple sensor and multiple location LED (Zone III type) - (Optional) qty 1 no.
24	Dry Chemistry laser camera as specified	24 a Dry Chemistry laser camera as specified qty -1 no b. films (14 x 17)inch , (11x 14) inch , (10 x 8)inch- 500 Nos each size
		Added Para :- Anaesthesia bed/trolley for recovery room -qty -2 nos.
	Components of Site Modification Work:	Components of Site Modification Work: for 1500 sqft
1	Civil works	Civil works
2	Electrical work	Electrical work
3	Public health (plumbing and sanitary fittings).	Public health (plumbing and sanitary fittings).
4	Air Conditioning	Air Conditioning -20 Tr
	Furniture:	Furniture:
1	Revolving chairs height adjustable, medium-back with hand-rest	Revolving chairs height adjustable, medium-back with hand-rest - qty 6 nos.
2	Chairs for patient waiting area – Three seater (chrome plated). -	Chairs for patient waiting area – Three seater (chrome plated). Qty 10 Nos.
3	Cupboard with laminate door shutters	Cupboard with laminate door shutters for storage of spare parts and accessories (approx size of 6'X3') and records as per requirement. – 3 NO.S
4	Drug trolleys for patient preparation area.	Drug trolleys for patient preparation area. Qty- 1 Nos.
5	Patient trolley with rubber foam mattress	Patient trolley with rubber foam mattress qty- 2 nos.
6	Tables for Workstation nodes.	Tables for Workstation nodes. Qty -2 nos.
7	Changing rooms (with change lockers and dressing table).	Changing room with 6 lockers and one dressing table to be supplied qty 1 set
8	Dustbins (plastic with lid) to be provided as required.	Dustbins (plastic with lid) to be provided as required. Qty- 10nos.
9	Room Signage	Room Signage 1 LS
10	Venetian Blinds	Venetian Blinds 1 LS
	Miscellaneous:	Miscellaneous:
11	LED X-ray Film viewer with adjustable brightness; capable of holding 3 films of 14 x17 size.	LED X-ray Film viewer with adjustable brightness; capable of holding 3 films of 14"x17" size. Qty 2 nos.
12	Cabling of Network (LAN) connectivity for camera system, console system, workstation and computers etc	Cabling of Network (LAN) connectivity for camera system, console system, workstation and computers etc 1 LS
14	Dry chemical powder type fire extinguisher of 5kgs capacity	Dry chemical powder type fire extinguisher of 5kgs capacity as per specification 3 Nos.

		Added para Broad band Connection : for remote service of MRI System qty -1 nos
		Added para :- UPS qty 1 no.
		Added para: MRI Room oxygen deficiency level monitor - 1 no MRI compatible WHEEL CHAIR - 2 NOS

Schedule No. 03
Digital X-Ray 1000 mA (Rfx No. 300003619)

Tender Page & Para	TENDER SPECIFICATION	Read as
Para 3 c, page 66	KVP range 40 kV - 150 kV with 1 kV steps	KVP range 40 kV - 150 kV
Para 9g, page 68	Added Para: Automatic stitching should be available.	Added Para: Automatic stitching should be available. (Both vertical and horizontal)
Para 1. a.i, page 69	Flooring 600 x 600 mm vitrified tiles with 100mm tile skirting to match in console room, lobby and patient preparation areas, Radiologist room etc.	Flooring 600 x 600 mm vitrified tiles with 100mm tile skirting to match in console room, lobby and patient preparation areas, Radiologist room for entire turnkey area only.
Para 4c, page 70	AIR CONDITIONING: Ventilation is required in toilet.	Deleted
Page 6c, page 70	Furniture: Cupboard with laminate door shutters for storage of spare parts and accessories and records as per requirement. – 3 NO.S	Furniture: Cupboard with laminate door shutters for storage of spare parts and accessories and records as per requirement. – 3 NO.S (Size approx 6'X 3')
Page 6j, page 71	Any other essential furniture item as per requirement	Deleted
Para e, page 69	Patient Waiting area	Deleted

Schedule No. 04
Mobile X-Ray (High Frequency) (Rfx No. 300003620)

Tender Page & Para	Tender Specification	Read As
pg 73 para 4	The tube stand should be fully counterbalanced/ Spring Balanced with rotation in all directions.	The tube stand should be fully counterbalanced/ Spring Balanced with Tube Head rotation in all directions.
		Added para The machine should be European CE with 4 digit notified body number or US FDA or BIS certification.

Schedule No. 05
Colour Doppler (2D & 3D) (Rfx No. 3000003621)

Tender Page & Para	TENDER SPECIFICATION	Read as
Para 1.2 , page 75	The system shall include at least a 17" LCD monitor to allow for both excellent images viewing as well as providing for workflow and productivity features	The system shall include at least 19" LCD monitor to allow for both excellent images viewing as well as providing for workflow and productivity features
Para 3, page 75	Tissue Harmonic imaging with contrast should be available as standard feature	Tissue Harmonic imaging should be available as standard feature
Para 6 , page 75	System should have minimum of 17,000 Digital Channels for better resolution.	System should have minimum of 60,000 or more Digital Channels for better resolution.
Para 10.2, page 76	The system shall include at least 100 GB of dedicated hard drive for large local storage capacity, with 20000 image storage capacity or more	The system shall include at least 300 GB of dedicated hard drive for large local storage capacity, with 40000 image storage capacity or more

Schedule No. 06
Portable Colour Doppler (Rfx No. 3000003622)

Tender Page & Para	TENDER SPECIFICATION	Read as
BOQ ,Point 7	Thermal Printer Paper (Rolls)	Thermal paper rolls: 50 nos
		Added para: Sector Probe for Cardiac application 2-4MHz - 1 no

Schedule No. 07
CR System (Rfx No. 3000003623)

Tender Page & Para	TENDER SPECIFICATION	Read as
Para 2c, page 80	It should have a resolution of 6 pixels/mm (minimum) for standard resolution cassettes & 10 pixel/mm (minimum) for high resolution cassette reading	It should have a resolution of 5 pixels/mm (minimum) for standard resolution cassettes & 10 pixel/mm (minimum) for high resolution cassette reading
Para 4d, para 81	The system must deliver its first film within 80 seconds from the request sent	The system must deliver its first film within 85 seconds from the request sent

Schedule No. 08
Digital Mammography System (Rfx No. 3000003624)

Tender Page & Para	TENDER SPECIFICATION	READ AS
Para 3q, Page 84	Regular (19 x 23)+/- 2 cm sliding paddle	Regular (19 x 23)+/- 2 cm paddle.
Para 6a, page 84	Beam filter: 2 Nos. (Mo/Ag/Al) & Rh	Beam filter: 1 No or more. (Mo/Ag/Al) with Rh
Para 7d, page 84	d. DQE at 0 LP/mm: 60% OR DQE at 5LP/mm : 29%	d. DQE at 0 LP/mm: 60% OR DQE at 5LP/mm : 25% or more
Page 12 f, page 85	Spot sliding compression paddle	Spot compression paddle
Page 12 K, page 85	Dry view camera: 508 DPI or more, with 300 films each (size 18x24cm and 24x30cm)	Dry view camera: 600 DPI or more, with 300 films each (size 18x24cm and 24x30cm)
Para 4 a, page 88	AIR CONDITIONING: a. Package air conditioners units and split AC units may be used according to room requirement and suitability. Humidity control should be effective to eliminate moisture condensation on equipment surface. The Air conditioning should be designed with standby provision to function 24 hours a day.	AIR CONDITIONING: a. Package air conditioners units/Split AC units may be used according to room requirement and suitability. Humidity control should be effective to eliminate moisture condensation on equipment surface.
Para 4c , page 88	Ventilation is required in toilet.	Deleted
BOQ point 11	Dry view camera: 508 DPI or more, with 300 films each (size 18 X 24cm and 24X30cm)	Dry view camera: 600 DPI or more, with 300 films each (size 18 X 24cm and 24X30cm) qty 1

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids