Prebid queries/clarifications asked by Firms and its recommendations/replies (ref: prebid meeting dated: 07/08/2018) NIB ref: HITES/PCD/NCI-AIIMS/24/18-19 dated 26.07.2018

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|---|--|--|
| | | | TECHNIC | AL ISSUES | |
| | | | Rfx No. 3000003190, | Chemport Insertion Set | |
| 1 | page 46, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. I.G. Instrument LLP | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. It should be fda/ce/bis / iso 13485+declaration of c onformity. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified |
| 2 | page 46, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Page 46, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | Added para :- Items mentioned at SNo 7,8,are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 4 | Page 46, BOQ point 1 | SS tray medium | Cure surgical | Delete | No Change |
| 5 | Page 46, BOQ point 4 | Mosquito forceps curved | Cure surgical | 125mm | To be Amended as: Mosquito forceps curved 12-14cm |
| 6 | Page 46, BOQ point 5 | Artery forceps curved | Cure surgical | 150mm | To be Amended as: Artery forceps curved 14-16cm |
| 7 | Page 46, BOQ point 6 | Allis forceps | Cure surgical | 150mm | To be Amended as: Allis forceps 14-16cm |
| 8 | Page 46, BOQ point | Eyelid retractor | Cure surgical | 140mm | To be Amended as: Eyelid retractor |

Page 1 of 10

I that our of

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|------------------------------|--|--|
| | 9 | | | | 14-16cm |
| 9 | Page 46, BOQ point 14 | Metzenbaum scissors | Cure surgical | medium | To be Amended as: Metzenbaum scissors 15-17cm |
| 10 | Page 46, BOQ point 15 | Mayo scissors | Cure surgical | medium | To be Amended as: Mayo scissors 15-17cm |
| 11 | Page 46, BOQ point 17 | Sponge holder | Cure surgical | 240mm | To be Amended as: Sponge holder 22-24cm |
| | | | Rfx No. 300000319 | 1, Trachestomy Set | |
| 1 | page 48, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified |
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | |
| 2 | page 48, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | To be Amended as: Deleted |
| 3 | Page 48, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | Added para :- items mentioned at SNo 2,7,9,18, are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 4 | Page 48, BOQ point 6 | Iris scissor curved | Cure surgical | 11.5cm | To be Amended as: Iris scissor curved 11-13cm |

Page 2 of 10 Dec

TA

on E

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|--|---------------|---------------------------------------|---|
| 5 | Page 48, BOQ point 7 | Iris Scissor straight | Cure surgical | 11.5cm | To be Amended as: Iris Scissor straight 11-13cm |
| 6 | Page 48, BOQ point 8 | Mayo Scissors | Cure surgical | medium | To be Amended as: Mayo Scissors 15-17cm |
| 7 | Page 48, BOQ point 10 | Non-toothed forceps | Cure surgical | medium | To be Amended as: Non-toothed forceps 15-17cm |
| 8 | Page 48, BOQ point 11 | S.S Bowl (small) | Cure surgical | medium | To be Amended as: Deleted |
| 9 | Page 48, BOQ point 12 | S.S tray (Medium) | Cure surgical | Delete | To be Amended as: Deleted |
| 10 | Page 48, BOQ point 13 | Sponge holder | Cure surgical | 240mm | To be Amended as: Sponge holder 24-26cm |
| 11 | Page 48, BOQ point 14 | Toothed dissections forceps | Cure surgical | medium | To be Amended as: Toothed dissections forceps 15-17cm |
| 12 | Page 48, BOQ point 17 | Tracheal dilator | Cure surgical | 14cm | To be Amended as: Tracheal dilator 14-16cm |
| 13 | Page 48, BOQ point 18 | Tracheal retractor 1 Prong (Tracheal hook) | Cure surgical | 160mm | To be Amended as: Tracheal retractor 1 Prong (Tracheal hook) 16- 18cm |
| 14 | Page 48, BOQ point 19 | Tracheal retractor 2 Prong | Cure surgical | 160mm | To be Amended as: Tracheal retractor 2 Prong 16-18cm |
| 15 | Page 48, BOQ | Tracheal retractor double | Cure surgical | 160mm | To be Amended as: Tracheal retractor double prong 16-18cm |

Here or K

Page 3 of 10

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|------------------------------|--|--|
| | point 20 | prong | | | |
| | aleren Tulks | | Rfx No. 300000319 | 92, Chest Tube Set | |
| 1 | page 49, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified |
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | |
| 2 | pag∈ 49, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Pag∈ 49, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | Added Para :- Items mentioned at SNo .3,6,14,15 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 4 | Pag∈ 49, BOQ point 1 | SS tray medium | Cure surgical | Delete | No change |
| 5 | Pag∈ 49, BOQ point8 | Sponge Holder | Cure surgical | 240mm | To be Amended as: Sponge Holder 24-26cm |

Joe on R Mars A

Page 4 of 10

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|------------------------------|--|--|
| 6 | Page 49, BOQ point 9 | Needle holder | Cure surgical | medium | To be Amended as: Needle holder 15-17cm |
| 7 | Page 49, BOQ point 13 | Small Bowl | | | Amended as: S.S. Bowl Small |
| | | | Rfx No. 3000003193 | 3, Minor Surgical Set | |
| 1 | page 50, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified |
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | |
| 2 | page 50, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Page 50, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | . Added Para :- items mentioned at SNo 4,6,14,19,23a,23c,23g are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |

Page 5 of 10

Jee om K

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|--|---------------|---------------------------------------|---|
| 4 | Page 50, BOQ point 1 | SS tray medium – 1 no. | Cure surgical | Delete | Amended as: SS tray medium – 3 no. |
| | Page 50, BOQ point 2 | Towel clips large – 1 no. | | | Amended as: Towel clips large – 13 no. |
| | Page 50, BOQ point 3 | Towel clips medium – 4 no. | | | Amended as: Towel clips medium – 23 no. |
| | Page 51, BOQ point 20 | SS Bowl medium-1no. | | | Amended as: SS Bowl medium-10nos. |
| | Page 51, BOQ point 22 | Sponge holder – 1no. | | | Amended as: Sponge holder -10nos. |
| | Added Para: BOQ s.no.24 | | | | Added BOQ para: Towel clips small – 19 nos. |
| | Added Para: BOQ s.no.25 | | | | Added BOQ para: SS Tray Large – 8 nos. |
| | Added Para: BOQ s.no.26 | | | | Added BOQ para: SS Bowl large – 6 nos. |
| | Added Para: BOQ s.no.27 | | | | Added BOQ para: Kidney tray medium – 1no. |
| | Added Para: BOQ s.no.28 | | | | Added BOQ para: Kidney tray big – 4 nos. |

Page 6 of 10

fre on h

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|------------------------------|--|---|
| 1 | page 52, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified |
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | |
| 2 | page 52, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Page 52, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | Added Para :- items mentioned at SNo 2,8 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| | Page 52, BOQ point 4 | Small bowl | | | To be Amended as: SS Bowl Small |
| 4 | Page 52, BOQ point 9 | Stitch scissors | Cure surgical | 121mm | To be Amended as: Stitch scissors 12-14cm |
| | | | Rfx No. 3000003195, E | pidural Set, Qty. 30 Set | |
| 1 | page 53, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 |

Page 7 of 10

Where the

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|------------------------------|--|---|
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | certified |
| 2 | page 53, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Page 53, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | No Change |
| 4 | Page 53, BOQ point 2 | Sponge Holder | Cure surgical | 240mm | To be Amended as: Sponge Holder 24-26cm |
| 5 | Page 53, BOQ point 3 | S.S. Bowl | Cure surgical | medium | To be Amended as: S.S. Bowl Medium |
| | | | Rfx No. 300000 | 3196, CVP Tray | |
| 1 | page 54, point 2 | The Instruments Should be FDA or CE or BIS certified | Volksman n India Pvt Ltd. | ISO 13485 with the declaration of conformity should be mentioned in the specification in addition to FDA/CE/BIS. | To be Amended as: The Instruments Should be FDA or CE or BIS certified or ISO 13485 |
| | | | I.G. Instrument LLP | It should be fda/ce/bis / iso 13485+declaration of c onformity . | certified |

Com K S AD Gere A OV

Page 8 of 10

| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee |
|------------|---------------------------------|---|---------------|---------------------------------------|--|
| 2 | page 54, point 3 | The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free | Cure surgical | delete | Deleted |
| 3 | Page 54, point 7 | Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. | Cure surgical | six week time | To be Amended as: litems mentioned at SNo .4,5,7 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval. |
| 4 | Page 54, BOQ point 3 | Sponge holder | Cure surgical | 240mm | To be Amended as: Sponge holder 24-26cm |
| 5 | Page 54, BOQ point 5 | One tooth forceps | Cure surgical | 150mm | To be Amended as: One tooth forceps 15-17cm |
| 6 | Page 54, BOQ point 6 | B P Hande No. 3 | Cure surgical | 1 | No Change |
| 7 | Page 54, BOQ point 7 | Curved scissors | Cure surgical | 150mm | To be Amended as: Curved scissors 15-17cm |

Adhert F web-

Page 9 of 10

| | Commercial issues | | | | | | |
|------------|------------------------------------|---|---------------------|---|--|--|--|
| SI. No. | Ref. to the Bidding Document | Specification/clauses in the Bidding Document | Name of Firms | Queries/clarifications asked by Firms | Recommendations/ replies by the Prebid Committee | | |
| | Page no. 22, | | Volksman n India | Suggested by bidder: - EMD should be waived | It is advised to refer GIB clause | | |
| | Point No. 39.2 | | Pvt Ltd. | off for MSME & Start Up's. | no.19.2 (pg.16) of the bidding document for MSME | | |
| | Page no. 33, Point No. 15.1 | | I.G. Instrument LLP | Suggested by bidder :- 2. Qualification criteria | No change considered | | |
| | | | | relaxation for msme and startup who manufactur | | | |
| | | | | e instruments , as per the procurement policy | | | |
| | | | | This should be waived off who meet quality parameters . | | | |
| | Page no. 33, | | Volksman n India | Suggested by bidder :- 3. Qualification Criteria | No change considered | | |
| | Point No. 15.4 | | Pvt Ltd. | should be waived off for start ups | | | |

5

A.C.



Page 10 of 10