

AMENDMENT NO.2 Dated 07.10.2025

REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

Amendment made to the details given in Tender Document.

REF NO: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

CPP TENDER ID : 2025_HLL_250885_1

The following amendment has been incorporated in the above mentioned tender.

The following amendment / Inclusion have been incorporated to the bid document for the above tender:

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 1 - Should be brand new automated, integrated system, dedicated to HbA1c testing based on HPLC technology for measurement of HbA1c and Hemoglobinopathies

MAY BE READ AS

Point No.1 – Should be brand new automated, integrated system, dedicated to HbA1c testing based on HPLC technology for measurement of HbA1c.

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 2 - Complete ready to use kit should be provided with Buffers in transparent plastic tanks to view the level of buffers; columns, primers, calibrators & sample vials / similar pack with facility to know the quantity of buffers; columns, primers, calibrators & sample vials.

MAY BE READ AS

Point No.2 – Complete ready to use kit should be provided with Buffers in transparent plastic tanks to view the level of buffers; columns, primers, calibrators & sample vials **OR** The system should have a real time inventory management system for monitoring of all reagents and buffers where all the reagent levels can be viewed from a single screen aiding to operator convenience.

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 8 - The system should have a feature of rack & sample position identification to avoid error in case of bad/fault barcode reading.

AMENDMENT NO.2 Dated 07.10.2025

REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

MAY BE READ AS

Point No: 8 - The system should have a feature of rack & sample position identification **OR** automatic onboard barcode reading facility and sample position identification to avoid error in case of bad/fault barcode reading.

All relevant clauses of the Tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & Group Head (HCS)