

**Amendment No. 1****Date: 22/01/2019****Sub: Amendment to the Tender Enquiry Document.****Ref: NIT No.: HLL/PCD/CDSCO-04-RT/18-19 Dated 26.12.2018**

The following changes are being incorporated in the referred tender enquiry.

**SECTION I  
NOTICE INVITING TENDERS (NIT)****1.) EXISTING:**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
c	Closing date & time for submission of <b>tender processing fee and EMD in physical form</b>	<b>25.01.2019, 1400 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender processing fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	<b>25.01.2019, 1300 hrs IST</b>
e	Time and date of opening of online bids	<b>25.01.2019, 1430 hrs IST</b>

**READ AS:**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
c	Closing date & time for submission of <b>tender processing fee and EMD in physical form</b>	<b>05.02.2019, 1400 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender processing fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	<b>05.02.2019, 1300 hrs IST</b>
e	Time and date of opening of online bids	<b>05.02.2019, 1430 hrs IST</b>

**SECTION – XII  
QUESTIONNAIRE****ADDED AS:****Fill up the Techno-Commercial Compliance Sheet Bid provided in spreadsheet (Excel file) and upload in the C-Folder.**

## SECTION-VII TECHNICAL SPECIFICATIONS

Schedule No. 01 -UV Spectrophotometer		
Para	TENDER SPECIFICATION	READ AS
Para 12 Pg 44	Additional spare - 1 pair quartz cuvette (1-cm,5ml) - 1 no.	Additional spare - 1 pair quartz cuvette (1-cm, <b>3.5ml</b> ) - 1 no.
Para 13 Pg 44	Software with P.C. and printer: -Suitable software 21 CFR part II compliance, laser printer and computer as per general specification	a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Pg 44	Suitable ups with 01 hour backup and surge protector should be supplied with the system	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
Schedule No. 02 Fourier Transform Infra Red Spectrophotometer (FT-IR)		
Pg 45	S/N Ratio: 10000:1 or better Peak-Peak, 1 minute Scan	S/N Ratio: <b>30000:1</b> or better Peak-Peak, 1 minute Scan
Pg 46	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, coloured Laser Printer and computer as per General Specification	a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Pg 46	Suitable ups with 01 hour backup and surge protector should be supplied with the system	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
Schedule No. 03 Digital Melting Point Apparatus		
Para 12 Pg 47	Compatible PC, Printer and software to be quoted	Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Pg 47	Suitable ups with 01 hour backup and surge protector should be supplied	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
Schedule No. 04 Digital Polarimeter		
Pg 48	Optical Cell: Stainless steel cell of 100mm and 200mm length	Optical Cell: Stainless steel cell of <b>100mm (5ml) and 200mm (10ml) length - 1 each</b>
Pg 48	Measurement Accuracy	Deleted
Pg 48	Resolution 0.01	Deleted
Pg 48	Temperature Controlled Cell Compartment	Temperature controlled cell compartment by <b>automatic Thermostat/Peltier</b>
Pg 48	Accessories CELL: 1 dm (5 ml), 2 dm (10 ml)	Deleted

Pg 48	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, Laser Printer and computer as per General Specification.	a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer. C. Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
<b>Schedule No. 06 Gas Chromatography with Head Space</b>		
Pg 52	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, Laser Printer and computer as per General Specification	a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer. C. Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
<b>Schedule No. 07 Tablet Dissolution Apparatus with Auto sampler</b>		
Page 54	Data Handling System Suitable Software CFR – 21, Part -11 compliance. Suitable computer as per specification mentioned in General Specification Suitable Laser Printer	<b>Data Handling System</b> a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Page 54	Suitable UPS with 01 hour backup and surge protector should be supplied with the system	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
<b>Schedule No. 08 HPTLC SYSTEM</b>		
Page 56	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, Laser Printer and computer as per General Specification	SOFTWARE WITH P.C. AND PRINTER a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Page 57	Suitable UPS with 01 hour backup and surge protector should be supplied with the system.	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
<b>Schedule No. 11 Atomic Absorption Spectrophotometer</b>		
Page 60	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, Laser Printer and computer as per General Specification	SOFTWARE WITH P.C. AND PRINTER a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
Page 60	Suitable UPS with 01 hour backup and surge protector should be supplied with the system.	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.

<b>Schedule No. 12 Rotational viscometer</b>		
pg 62	SOFTWARE WITH P.C. AND PRINTER Suitable software 21 CFR PART II compliance, Laser Printer and computer as per General Specification	SOFTWARE WITH P.C. AND PRINTER a. Suitable software 21 CFR part II compliance b. Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer.
pg 62	Suitable UPS with 01 hour backup and surge protector should be supplied with the system.	Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**