

**Amendment No. 8**

01.03.2019

**Sub: Amendment to the Bidding Document**

**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/36/18-19 dated 26.09.2018 read with its Amendment No. 01, 02, 03, 04, 05, 06 & 07 dated 19.11.18, 06.12.18, 11.12.18, 27.12.18, 07.01.19, 11.02.19 & 26.02.19 respectively**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - VII****TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 2 (Rfx/ Event number 3000003420)****IORT Machine (Electron Based)**

***The following para is added in the existing specification:***

<b>Sr. no</b>	<b>Added Para</b>
1	Vendor should provide the TPS for 3D treatment planning simulation software for IORT (electron based) for pre-, intra- and post-treatment operations. TPS should offer the possibility of import export images, contouring, planning and dose computation with the Monte Carlo algorithm to correct tissue heterogeneity and plan evaluation and enables easy documentation.
2	TPS should be DICOM compatible and to be connected with existing CT/ MRI/PET-CT/PET-MRI imaging system.
3	The TPS software shall be DICOM RT/3.0 compliant and HL 7 compliant.

**Note:**

Technical Compliance Sheet in HLL e-bid portal 'for RFX / Event number 3000003420' has been replaced by a new version i.e. 'Ver1.1'. The same need to be downloaded for submission.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**