

10-05-2018

**Amendment No. 3****Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/27/CTVS/18-19 dated 10/04/2018****Amendment No. 01 dated 27-04-2018****Amendment No. 02 dated 08-05-2018**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION VII  
TECHNICAL SPECIFICATION**

<b>Schedule No. 03 -Battery Operated Sternum System (Rfx no. 3000002708)</b>		
<b>Para</b>	<b>CURRENT TENDER SPECIFICATION</b>	<b>READ AS</b>
	BATTERY POWERED VERTICAL (RECIPROCATING)STERNUM SAW& REDO (OSCILLATING) STERNUM SAW.	BATTERY POWERED VERTICAL (RECIPROCATING)STERNUM SAW& REDO ( <b>OSCILLATING/ SAGITTAL</b> ) STERNUM SAW.
2	Redo (Oscillating) Sternum Saw (Qty: 1)	Redo ( <b>Oscillating/Sagittal</b> ) Sternum Saw (Qty: 1)
BOQ (2)	Redo(Oscillating) Sternum Saw	Redo( <b>Oscillating/ Sagittal</b> ) Sternum Saw
<b>Schedule No. 12- Thoracic Videoscopy System (Rfx no. 3000002717)</b>		
<b>Para</b>	<b>CURRENT TENDER SPECIFICATION</b>	<b>READ AS</b>
3.1.7	Optical System- Field of view 90-100 degrees or more; Directional of view- <b>30 degree (forward viewing)</b> ; Depth of field- 18 to 100 mm; <b>3 Chip CCD FHD Camera</b>	Optical System- Field of view <b>65 degrees or more</b> ; Directional of view- <b>30 degree (forward viewing)</b> ; <b>3 Chip CCD FHD Camera with Depth of field- 18 to 100 mm</b>

**All other contents of the tender enquiry including terms & conditions remain unaltered.****Note:**

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**