Amendment No. 2

Date: 08/07/2014

Subject: Amendment to the Tender Enquiry Document

Ref: NIT ref.: HLL/PCD/GNCTD/09/LNH/14-15 dated 04/06/2014 read with its Amendment No. 1 dated 23.06.2014

Tender ID in GNCTD e-Procurement portal: 2014_HFWD_58813_27

The following changes have been incorporated in the referred tender.

SECTION VII

Item no. 27

3.0 Tesla MRI System

1. Existing specifications-Para 8 ix: Dedicated Knee Coil - at least 12 Channels or more transmit & receive coil.

Read as: Dedicated Knee Coil - at least 8 Channels.

2. Existing specifications-Para 8 xii: Small flex coil – 8 Channel or more coil for Neonatal Head and Neck imaging.

Read as: Small flex coil – 8 Channel or more coil for Neonatal Head and Neck imaging. **Or any coil capable of neonatal head and Neck imaging.**

3. Existing specifications-Para a, Neuro Applications 4 (pg. 88): Susceptibility weighted imaging with phase information SWI/SWIp/eSWAN

Read as: Susceptibility weighted imaging with phase information SWI/SWIp/SWAN2

4. Existing specifications-Para e, Vascular Imaging 4 (pg. 89): Time resolved angiography with contrast kinetics like 4D TRAKS/TWIST/TRICKS 15

Read as: Time resolved angiography with contrast kinetics like 4D TRAKS/ TWIST/ TRICKS

Note added under the head "GENERAL TECHNICAL SPECIFICATIONS" (ref. pg.101)

- **Note 5:** OPTIONAL ITEMS: Bidders are requested to quote for all the optional items as asked in the bidding document with reasonable pricing. However, the pricing for optional items will not be considered for price comparison and ranking purpose. If the firm has not quoted for any optional item (except the items of turnkey) their offer will be treated as TECHNICALLY RESPONSIVE if otherwise meeting the specification.
- Note 6: Separately quoted items (which are not optional) are to be considered for evaluation and ranking purpose. Therefore, offer should clearly specify if these separately quoted items are already included in the main offer or not.

All other contents of the tender enquiry including terms & conditions remain unaltered.