Amendment No.1 to Tender Enquiry Document

Sub: Amendment to Tender Enquiry Document.

Ref: NIT No. HLL/PCD/RIMS-05/15-16 Dated 05.10.2015

The following amendments are issued with respect to the above tender Enquiry Document.

Existing:

Section – VI "LIST OF REQUIREMENTS"

Part I:

Item No.	Description	Qty.	warranty	CMC
		(Nos)	period	
1	Dental chairs and unit	52	3 Years	2 Years

Read As:

Part I:

Item No.	Description	Qty. (Nos)	warranty period	CMC
1	Dental chairs and unit	54	3 Years	2 Years

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.

Existing:

Part II: Required Delivery Schedule:

c) For MOBILE DENTAL VAN:

Within **60 days** from the date of Notification of Award, as followed by CMVR compliance as per GCC clause no. 8.9

Read As:

Part II: Required Delivery Schedule:

c) For MOBILE DENTAL VAN:

Within **90 days** from the date of Notification of Award, as followed by CMVR compliance as per GCC clause no. 8.9

Existing:

Section – IX Qualification Criteria

02. (b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, **at least 50% of the quoted quantity** of similar equipments which is functioning satisfactorily, anywhere in India of the same manufacturer.

Read As:

Section – IX Qualification Criteria

02. (b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria 02 (a) should have supplied and installed in last **Five** years from the date of Tender Opening, **at least 25% of the quoted quantity** of similar equipments which is functioning satisfactorily, anywhere in India of the same manufacturer.

Section - VII

Technical Specification

Item no-1: Dental Chair

1. Existing Specification:

Para 3.7: It should have LED light cure unit on unit sides (Min. Intensity 1200 mW/cm2).

Read as:

Para 3.7: It should have LED light cure unit on unit sides (with cord or cordless) (Min. Intensity 1200 mW/cm2).

2. Existing Specification:

Para 3.14: It should have latest foot operated minimum dual intensity LED Light (30,000 to 40,000 LUX).

Read as:

Para 3.14: It should have latest foot operated/sensor based minimum dual intensity LED Light (30,000 to 40,000 LUX).

Item no-3: C-ARM SYSTEM (LOCAL)

1. Existing Specification:

Para: Generator - Microprocessor controlled generator with 2.0Kw or More with integrated beam filters to reduce patient skin radiation dose.

Read as:

Para: Generator - Microprocessor controlled generator **with 1.4Kw or More** with integrated beam filters to reduce patient skin radiation dose.

2. Existing Specification:

Para: Fluoroscopy- Fluoroscopy should not exceed 5 mA.

Read as:

Para: Fluoroscopy- Fluoroscopy should not exceed 9 mA.

3. Existing Specification:

Accessories:

a) Wrap around light weight vinyl Lead Aprons with 0.5 mm lead equivalence certified by BARC or AERB **or ISO**: 6 (Six Nos.)

Read as:

Accessories:

a)Wrap around light weight vinyl Lead Aprons with 0.5 mm lead equivalence certified by BARC or AERB: 6 (Six Nos.)

4. Added Para:

- i) The CCD camera must have at least 1Kx1K resolution with rotation facility.
- ii) The machine should have image storage of at least 50,000 & more images.
- iii) Model should be AERB type approved.

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.