

Amendment No. 02

Date: 17.02.2016

Subject: Amendment no. 02 to the Tender Enquiry Document

Ref: Tender Ref: HLL/PCD/PMSSY-II/NAGPUR/06-RT/15-16 dated 29.01.2016 for CT Scan

SECTION I

NOTICE INVITING e-TENDERS (NIT)

The schedule for submission of tenders against the referred NIT has been revised as follows.

Sl.	Description	Schedule
	i) Closing date & time for submission of online bids	03-March-2016 , 1800 hrs IST
d	ii) Closing date & time for tender fee, EMD and bid documents in physical form (Price bid shall only be uploaded online; All other technical and commercial documents uploaded in the e-portal are to be submitted)	04-March-2016 , 1200 hrs IST Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
e	Time and date of opening of bids	04-March-2016 , 1500 hrs IST

Note:

- (i) If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e., upto 16 August 2016.
- (ii) Tender fee (Rs.5,000/-) and EMD (As applicable) should be deposited in the Tender Box located at HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh on or before 04-March-2016, 1200 hrs IST. Submission beyond stipulated date & time would result in REJECTION of BID.

Section – VII

Technical Specifications

1. Existing Specifications :

Para 7.a: System should have minimum 64 rows of detector capable of generating 128 slices through latest flying focal spot technology or equivalent (mandatory requirement).

Read as :

Para 7.a: System should have minimum 64 rows of detector capable of generating 128 slices.

2. Existing Specifications :

Para 8.a: Real time reconstruction speed: 20 images per second or more at 512 x 512 matrix

Read as :

Para 8.a: Real time reconstruction speed: **18** images per second or more at 512 x 512 matrix.

3. Existing Specifications :

Para 11.a: ASIR / iDose4 Premium / SAFIRE to be quoted as standard.

Read as :

Para 11.a: Latest Iterative Reconstruction Technique to be quoted as standard.

4. Existing Specifications :

Para 15.b:- Suitable Dual Head pressure injector, with complete accessories & 500 sets (Each set having 2 syringes), tubing and connector. **Kindly quote the rate of syringes, tubing and connector for future supply.**

Read as :

Para 15.b:- Suitable Dual Head pressure injector, with complete accessories & 500 sets (Each set having 2 syringes), tubing and connector.

5. Existing Specifications :

Para 15.P (Pg. no. 47):- Bidder should be responsible for supply, installation, commissioning and testing of A Diesel Generator of suitable capacity to run the CT Scanner, the associated equipment, air conditioning and lighting (Price to be quoted separately)

Read as:

Para 15.P (Pg. no. 47):- Bidder should be responsible for supply, installation, commissioning and testing of A Diesel Generator of suitable capacity to run the CT Scanner, the associated equipment, air conditioning and lighting. (Price to be quoted separately). **Laying of power cable from the generator control panel room to the CT scan room will be the responsibility of the vendor**
(This is as suggested GMC, Nagpur)

All other terms and conditions of the tender enquiry remain unaltered.