27-04-2018

Amendment No. 1

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-III/28/NEPH/18-19 dated 10/04/2018

The following changes are being incorporated in the above referred Tender Enquiry Document.

Section VII Technical Specification

Schedule No. 02 - CRRT Equipment RFX No. 3000002720			
Para	TENDER SPECIFICATION	READ AS	
3.13	Should have Effluent Flow Rate: 60-2000 mL/Hr	Should have Effluent Flow Rate: 0-1500 mL/Hr	

Schedule No. 03 - Dialysis Chair RFX No. 3000002721			
Para	TENDER SPECIFICATION	READ AS	
1	Auto CPR Function It should be safe and sturdy power seat with height adjustment (vertical lift of 150 - 300mm) which should reduce the risk of bend injuries.	Auto CPR Function It should be safe and sturdy power seat with height adjustment (vertical lift of 100 - 300mm) which should reduce the risk of bend injuries.	
3	It should have heavy duty washable vinyl.	It should have heavy duty washable vinyl/Rexine	
4	Zoned seat cushioning to accommodate patient upto 200Kgs.	Zoned seat cushioning to accommodate patient of at least 150Kgs.	
5	Adjustable back rest to provide maximum comfort. It should have two spring back and seat construction for comfort.	Adjustable back rest to provide maximum comfort.	
10	Seat Width (mm) 570 OR MORE.	Seat Width (mm) 540 OR MORE.	
14	It should have swing away arms up to 180 to make entry and exit easy of patient.	It should have swing away arms up to 180 deg to make entry and exit easy of patient.	
17	It should have 4" twin wheel thermoplastic casters with option of front locking caster.	It should have 4" twin wheel thermoplastic casters with option of locking caster.	
19	It should have full vinyl & foam covered arm rest. Vinyl seams it should be turned away for infections control & full medical grade vinyl bacterial resistant & anti-microbial.	It should have full vinyl/Rexine & foam covered arm rest. Vinyl/rexine seams it should be turned away for infections control & full medical grade vinyl bacterial resistant & anti- microbial.	
20	It Should also have one/two rear castors, It should have directional lock for easy transportation.	It Should also have one/two rear castors, It should have lock for easy transportation.	

	Schedule No. 05 - Hemodialysis Machine (RFX No. 3000002723)			
Para	TENDER SPECIFICATION	READ AS		
2.1	Machine should have facility for Bicarbonate, Dry powder, Sequential dialysis (Isolated UF)	Machine should have facility for Bicarbonate, Sequential dialysis (Isolated UF)		
3.3	Battery back-up for 20-30 minutes to run extra corporeal circuit	Battery back-up for 15-20 minutes to run extra corporeal circuit		
4.4	UPS of suitable rating with voltage regulation and spike protection for 60 minutes back up.	Online UPS of suitable rating with voltage regulation and spike protection for 60 minutes back up for 15 machines to be quoted and price to be quoted separately		
	BOQ			
4	RO Machine - 1000 LPM - 1 no	RO Machine - 1000 LPH - 30 nos(for price evaluation) and unit price to be quoted separately		
5	RO Machine - 1500 LPM - 1 no	RO Machine - 1500 LPH - 15 nos(for price evaluation) and unit price to be quoted separately		
		Added para: Online UPS for 15 machines - 1 no (price to be quoted separately)		
	RO plant 1500 & 1000 LPH			
В.З	Raw Water Tank having food grade quality for at least 1000 liters capacity to store Raw Water	Raw Water Tank having food grade quality for at least 750 liters capacity to store Raw Water		
D.1	Should have food grade material and conical shape Permeate Storage Tank of at least 1000 Liters capacity with level control system.	Should have food grade material and conical shape Permeate Storage Tank of at least 750 Liters capacity with level control system.		

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids