

Amendment No. 21

Date: 23/03/2016

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/07/15-16 dated 28/10/2015

The following changes have been incorporated in the referred NIT.

Section I
Notice Inviting Tender

Sch no	Event Number	Name of Item	Qty.	EMD	Due date & time for submission of online bids	Due date & time for submission of Tender fee and EMD in physical form	Date and time of opening of online bids
14	3000000617	ENT Treatment Unit (IMPORTED)	2	₹ 60,000.00	06-04-2016 1800 Hrs	07-04-2016 1130 Hrs	07-04-2016 1230 Hrs

Section VII
Technical Specification**Item Sl. No. 14**
ENT Treatment Unit (Imported)

1. **Existing Para:** Compressor air system, adjustable.
Read as: Compressor air system, adjustable **should be integrated into the unit.**
2. **Existing Para:** Warm water/ irrigation system with temperature display.
Read as: Warm water/ irrigation system with temperature display **should be integrated in main unit, water heated up to 37deg, indicator available, irrigation with the meter guage.**
3. **Existing Para: Integrated X Ray viewer.**
Read as: X Ray viewer.
4. **Existing Para:** Integrated holder for Endoscopes.
Read as: Should have endoscope management facility.
 - (a) Heated quivers for 3 rigid & one flexible endoscope.
 - (b) Disinfection time monitoring tanks for 3 rigid & one flexible endoscope.
5. **Added Para:** The model offered should be USFDA or European CE approved.

All other contents of the tender enquiry including terms & conditions remain unaltered.**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**