

AMENDMENT NO.1 Dated 10.07.2025
REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 08 Dated 02nd JULY 2025

Amendment made to the details given in Tender Document.

REF NO: HLL/CHO/PSD/HCS/2025-26/TENDER 08 Dated 02nd JULY 2025

CPP TENDER ID : 2025_HLL_240079_1

The following amendment has been incorporated in the above mentioned **E-TENDER FOR SUPPLY OF MICROSCOPES FOR HINDLABS DIAGNOSTIC CENTRE AT GOVT. TALUK HEAD QUARTERS HOSPITAL FORT KOCHI.**

The following amendment / Inclusion have been incorporated to the bid document for the above tender:

1. Notice Inviting Tender

FOR

LAST DATE AND TIME FOR RECEIPT OF TENDER – 16.07.2025, 15:30 Hrs

MAY BE READ AS

LAST DATE AND TIME FOR RECEIPT OF TENDER - 29.07.2025, 15:30 Hrs

2. Notice Inviting Tender

FOR

TIME AND DATE OF OPENING OF THE TENDER - 17.07.2025, 15:30 Hrs

MAY BE READ AS

TIME AND DATE OF OPENING OF THE TENDER – 30.07.2025, 15:30 Hrs

**3. ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED –
ITEM NO.01 – MICROSCOPE – MONOCULAR**

FOR

- a. Point No: 1 - Monocular microscope with stand type single mould aluminium die-cast with ergonomic handrests.

MAY BE READ AS

Point No: 1 - Monocular microscope with stand type aluminum die-cast with ergonomic handrests.

FOR

- b. Point No: 11 - Should have CDSCO approval.

MAY BE READ AS

Point No: 11 - Should have CDSCO approval/Registration.

AMENDMENT NO.1 Dated 10.07.2025
REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 08 Dated 02nd JULY 2025

FOR

- c. **ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED - NEW POINT ADDED**

MAY BE READ AS

Point No.12 – ISO 13485:2016 Certified.

4. ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED – ITEM NO.02 – MICROSCOPE - BINOCULAR

FOR

- a. Point No: 2 - Eyepieces - Highest quality 10 X wide field anti-fungus field eyepiece, FOV 20mm.

MAY BE READ AS

Point No: 2 - Eyepieces - Highest quality 10 X wide field anti-fungus field eyepiece, FOV 22mm.

FOR

- b. Point No: 3 - Objectives - Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with plan achromatic correction for FOV of 20 mm.

MAY BE READ AS

Point No: 3 - Objectives - Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with plan achromatic correction for FOV of 22 mm.

FOR

- c. Point No: 12 - Should have CDSCO approval.

MAY BE READ AS

Point No: 12 - Should have CDSCO approval/Registration.

FOR

- d. **ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED - NEW POINT ADDED**

MAY BE READ AS

Point No.14 – ISO 13485:2016 Certified.

5. ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED – ITEM NO.03 – FLUORESCENT MICROSCOPE WITH DIGITAL CAMERA

FOR

- a. Point No: 2 - Microscope must have a dual focusing and refocusing mechanism.

AMENDMENT NO.1 Dated 10.07.2025
REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 08 Dated 02nd JULY 2025

MAY BE READ AS

Point No: 2 - Microscope must have refocusing mechanism.

FOR

- b. Point No: 7 - Field of view should be > 20 mm

MAY BE READ AS

Point No: 7 - Field of view should be ≥ 22 mm

FOR

- c. Point No: 8 - Capable for brightfield, phase contrast, darkfield and fluorescence microscopy.

MAY BE READ AS

Point No: 8 - Capable for brightfield and fluorescence microscopy.

FOR

- d. Point No: 11 - Microscope must support fluorescence LED illumination system with DAPI, FITC, and TRITC filters.

MAY BE READ AS

Point No: 11 - Microscope must support fluorescence LED illumination system with Blue, Green, Violet and UV Filter sets.

FOR

- e. ***ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED - NEW POINT ADDED***

MAY BE READ AS

Point No.18 – ISO 13485:2016 Certified.

All relevant clauses of the Tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & Group Head (HCS)