

Amendment No. 3**Date: 23/03/2018****Subject: Amendment to the Tender Enquiry Document.****Ref: Tender no. HLL/HMC/Hospital Equipment/18-19/001 dated 01.03.2018**

The following changes have been incorporated in the referred tender.

**Section-I
NOTICE INVITING TENDER (NIT)****List of items to be read as:**

| SI no. | Equipment | Group | Qty for single Hospital | Qty for 20 hospital | EMD |
|--------|---|-------|-------------------------|---------------------|-----------|
| 2 | Hydraulic OT Table | O | 4 | 80 | 42,30,000 |
| 3 | Delivery Table with Footstep | | 4 | 80 | |
| 4 | Ceiling shadow less lights for OT | | 2 | 40 | |
| 5 | Portable shadow less lights for Labour Room | | 6 | 120 | |
| 6 | Examination Table SS with footstep | | 5 | 100 | |
| 7 | X-ray view box | | 5 | 100 | |
| 8 | Transfer Trolley for OT Sterile Zone | | 2 | 40 | |
| 9 | Gynaec electric cautery | | 2 | 40 | |
| 10 | Suction Apparatus high Vacuum | | 4 | 80 | |
| 11 | Suction Apparatus Electrical | | 4 | 80 | |
| 12 | Mucus Extractor with suction tube and foot operated suction machine | | 4 | 80 | |
| 22 | Multi Para Pulse Oxymeter | | A | 3 | |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre (15 monitor to be connected with one main Nursing Station) | 15 | | 300 | |
| 25 | Pediatric Ventilator for ICU | 2 | | 40 | |
| 27 | Transport Ventilator | 1 | | 20 | |
| 28 | Transport Ventilator Neonatal | 1 | | 20 | |
| 29 | Defibrillator | 2 | | 40 | |
| 43 | ECG Machine | 2 | | 40 | |
| 50 | Infusion Pump | 6 | | 120 | |
| 59 | Ambu Bag | 10 | | 200 | |
| 60 | Oxygen Concentrator | 1 | | 20 | |

| SI no. | Equipment | Group | Qty for single Hospital | Qty for 20 hospital | EMD |
|--------|---|-------|-------------------------|---------------------|-----------|
| 1 | Syringe & needle cutter | GNL | 20 | 400 | 8,60,400 |
| 13 | Adult resuscitation kit | | 4 | 80 | |
| 14 | Neonatal resuscitation kit | | 4 | 80 | |
| 15 | Oxygen Cylinder B Type | | 25 | 500 | |
| 16 | Nitrous Oxide Cylinder B Type | | 5 | 100 | |
| 17 | Regulator & Flow Meter for Medical Gas | | 25 | 500 | |
| 18 | Vento use (for vacuum extraction delivery) | | 2 | 40 | |
| 19 | MTP set with karmon suction canula and MVA syringe | | 02 set | 40 set | |
| 20 | Laryngoscope | | 1 | 20 | |
| 23 | Nebulizer | | 10 | 200 | |
| 36 | Stethoscope for Pediatric as well as Adult | | 10 | 200 | |
| 37 | BP Instrument stand type | | 5 | 100 | |
| 38 | BP Instrument table top model | | 5 | 100 | |
| 39 | Infant BP Instrument | | 5 | 100 | |
| 40 | Digital Weighing Machine Adult | | 5 | 100 | |
| 41 | Digital Weighing Machine Infant | | 5 | 100 | |
| 49 | Digital Thermometer | | 20 | 400 | |
| 56 | Glucometer | | 2 | 40 | |
| 58 | Height measuring stand | | 4 | 80 | |
| 62 | Air curtain at entrance for O.T, SNCU, ISU, Labour Room | | 8 | 160 | |
| 33 | Constant Temp Water Bath with Thermometer for Lab | L | 1 | 20 | 7,59,400 |
| 55 | Microscope binocular | | 2 | 40 | |
| 57 | Serum bilirubino meter | | 2 | 40 | |
| 66 | Bacteriological Incubator | | 1 | 20 | |
| 72 | Semi auto biochemistry analyzer | | 1 | 20 | |
| 73 | Blood Storage Refrigerator | | 1 | 20 | |
| 74 | Centrifuge- 12 Tubes | | 1 | 20 | |
| 75 | Blood Gas Analyzer | | 1 | 20 | |
| 76 | Hematology Analyzer | | 1 | 20 | |
| 77 | Blood Cell Counter (Manual) | | 1 | 20 | |
| 21 | Foetal Monitor | P | 3 | 60 | 17,32,000 |
| 34 | CPAP Machine with Heated Humidifier | | 2 | 40 | |
| 35 | Foetal Doppler | | 6 | 120 | |
| 63 | Radiant Warmer | | 6 | 120 | |
| 64 | Phototherapy Unit | | 4 | 80 | |
| 65 | Microprocessor Based Incubator (Transport type) | | 1 | 20 | |
| 67 | Baby Bassinet | | 10 | 200 | |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | R | 2 | 40 | 35,60,000 |
| 30 | Portable Ultrasound | | 1 | 20 | |
| 31 | 3-D Ultrasound- High End for Diagnostics | | 1 | 20 | |
| 32 | X-ray | | 1 | 20 | |

| SI no. | Equipment | Group | Qty for single Hospital | Qty for 20 hospital | EMD |
|--------|---|-------|-------------------------|---------------------|-----------|
| 42 | Dressing Drum SS of different sizes | X | 3 | 60 | 32,80,000 |
| 44 | Delivery Instrument Set | | 10 set | 200 set | |
| 45 | Episiotomy Instrument Set | | 10 set | 200 set | |
| 46 | MVA/EVA Instrument Set | | 08 set | 160 set | |
| 47 | PPIUCD Instrument Set | | 06 set | 120 set | |
| 48 | Outlets forceps | | 2 | 40 | |
| 51 | Caesarean Section | | 04 set | 80 set | |
| 52 | Suture removal | | 10 set | 200 set | |
| 53 | Suturing Tray | | 10 | 200 | |
| 54 | Anesthesia trolley and General Anesthesia kit | | 04 set | 80 set | |
| 61 | Instrument Tray Different Sizes | | 30 | 600 | |
| 68 | High Pressure Sterilizer Horizontal | | 1 | 20 | |
| 69 | High Pressure Sterilizer Vertical | | 2 | 40 | |
| 70 | Instrument Sterilizer | | 8 | 160 | |
| 71 | PN Sterilization | | 04 set | 80 set | |

Existing:

2. Tenderer would be required to submit as Interest free Earnest Money Deposit (EMD) in the form of Demand Draft in favor of 'HLL Mother & Child Care Hospitals Ltd.' payable at Noida in the envelope containing Technical Bid (as per GIT clause no.12.1).

Read as:

Tenderer would be required to submit as Interest free Earnest Money Deposit (EMD) in the form of Demand Draft or **Bank Guarantee or FDR (Fixed Deposit Receipt)** in favor of 'HLL Mother & Child Care Hospitals Ltd.' payable at Noida in the envelope containing Technical Bid (as per GIT clause no.12.1).

Existing:

8. The rate quoted by the bidder should be valid for a period of one year from the date of award of purchase order as per details mentioned in Financial Bid document.

Read as:

The tenders shall remain valid for acceptance for a period of 15 Months after the date of tender opening prescribed in the TE document. Any tender valid for a shorter period shall be treated as unresponsive and rejected.

**SECTION -II
GENERAL INSTRUCTIONS TO TENDERERS (GIT)**

Existing:

- 12.1** The tenderer shall furnish along with its tender, earnest money as mentioned in the section-1 in form of Account Payee demand draft to be drawn on any scheduled commercial bank in India in favour of the "**HLL Mother & Child Care Hospitals Limited**" payable at Noida.

Read as:

The tenderer shall furnish along with its tender, earnest money as mentioned in the section-1 in form of Account Payee demand draft **or bank guarantee or or FDR (Fixed Deposit Receipt)** to be drawn on any scheduled commercial bank in India in favour of the "**HLL Mother & Child Care Hospitals Limited**" payable at Noida.

SECTION-III SPECIAL INSTRUCTIONS TO TENDERERS (SIT)

Existing:

- Bidders have to submit the Manufacturer authorization form in the prescribed format mentioned at section XII for the items in the table below against their offered group of items.

| NIT S. No | Equipment | Group |
|-----------|--|-------|
| 2 | Hydraulic OT Table | A |
| 4 | Ceiling shadow less lights for OT | A |
| 5 | Portable shadow less lights for Labour Room | A |
| 9 | Gynaec electric cautery | A |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre | C |
| 25 | Pediatric Ventilator for ICU | C |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | C |
| 27 | Transport Ventilator | C |
| 28 | Transport Ventilator Neonatal | C |
| 30 | Portable Ultrasound | C |
| 31 | 3-D Ultrasound- High End for Diagnostics | C |
| 54 | Anesthesia trolley and General Anesthesia kit | D |
| 63 | Radiant Warmer | E |
| 65 | Microprocessor Based Incubator (Transport type) | E |
| 72 | Semi auto biochemistry analyzer | F |
| 73 | Blood Storage Refrigerator | F |
| 74 | Centrifuge- 12 Tubes | F |
| 75 | Blood Gas Analyzer | F |
| 76 | Hematology Analyzer | F |
| 77 | Blood Cell Counter (Manual) | F |

Read as:

- Bidders have to submit the Manufacturer authorization form in the prescribed format mentioned at section XII for the items in the table below against their offered group of items

| SI no. | Equipment | Group | Qty for 20 hospital |
|--------|---|-------|---------------------|
| 2 | Hydraulic OT Table | 0 | 80 |
| 4 | Ceiling shadow less lights for OT | 0 | 40 |
| 5 | Portable shadow less lights for Labour Room | 0 | 120 |

| Sl no. | Equipment | Group | Qty for 20 hospital |
|--------|--|-------|---------------------|
| 9 | Gynaec electric cautery | 0 | 40 |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre | A | 20 |
| 25 | Pediatric Ventilator for ICU | A | 40 |
| 27 | Transport Ventilator | A | 20 |
| 28 | Transport Ventilator Neonatal | A | 20 |
| 29 | Defibrillator | A | 40 |
| 50 | Infusion Pump | A | 120 |
| 72 | Semi auto biochemistry analyzer | L | 20 |
| 73 | Blood Storage Refrigerator | L | 20 |
| 75 | Blood Gas Analyzer | L | 20 |
| 76 | Hematology Analyzer | L | 20 |
| 77 | Blood Cell Counter (Manual) | L | 20 |
| 63 | Radiant Warmer | P | 120 |
| 65 | Microprocessor Based Incubator (Transport type) | P | 20 |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | R | 40 |
| 30 | Portable Ultrasound | R | 20 |
| 31 | 3-D Ultrasound- High End for Diagnostics | R | 20 |
| 32 | X-ray | R | 20 |
| 68 | High Pressure Sterilizer Horizontal | X | 20 |
| 69 | High Pressure Sterilizer Vertical | X | 40 |

SECTION-IV
General Conditions of Contract

Existing:

9.4 In case of any claim arising out of this warranty, the Purchaser shall promptly notify the same in writing to the supplier. Upon receipt of such notice, the supplier shall, within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis respond to take action to repair or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim, whatsoever shall lie on the purchaser for such replaced parts/goods thereafter. The penalty clause for non-rectification will be applicable as per tender conditions.

Read as:

In case of any claim arising out of this warranty, the Purchaser shall promptly notify the same in writing to the supplier. Upon receipt of such notice, the supplier shall, within **24** hours on a 24(hrs) X 7 (days) X 365 (days) basis respond to take action to repair or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim, whatsoever shall lie on the purchaser for such replaced parts/goods thereafter. The penalty clause for non-rectification will be applicable as per tender conditions.

20. Terms of Payment

Existing:

20.1 On successful installation & commissioning of equipment, designated Committee appointed by State shall conduct an inspection to inspect & certify that the commissioned equipment are as per

specification given in tender document. The payment to the supplier shall be released after 30 days of submission of report by the State Level Committee subject to submission of following documents:-

- Four copies of supplier's invoice showing contract number, goods description, quantity, unit price, total amount, Permanent Account Number and GSTIN ;
- Delivery Challan duly signed & stamped by the authorized representative of purchaser as per format.
- Documentary proof for supply, installation & commissioning of equipment.
- Crossed/cancelled cheque of Bank Account for making payment through NEFT/RTGS.

Read as:

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

100% payment shall be made on successful installation & commissioning of equipment, designated Committee appointed by State shall conduct an inspection to inspect & certify that the commissioned equipment are as per specification given in tender document. The payment to the supplier shall be released after 30 days of submission of report by the State Level Committee subject to submission of following documents:-

- Four copies of supplier's invoice showing contract number, goods description, quantity, unit price, total amount, Permanent Account Number and GSTIN ;
- Delivery Challan duly signed & stamped by the authorized representative of purchaser as per format.
- Documentary proof for supply, installation & commissioning of equipment.
- Crossed/cancelled cheque of Bank Account for making payment through NEFT/RTGS.
- Performance Bank Guarantee

Existing:

20.4 Irrevocable & non – transferable LC shall be opened by the Purchaser. However, if the supplier requests specifically to open confirmed LC, the extra charges would be borne by the supplier. If LC is required to be extended and/or amended for reasons not attributable to the purchaser/consignee, the charges thereof shall be borne by the supplier.

Read as: Deleted

SECTION-VI

LIST OF REQUIREMENT to be read as:

Part-1

| SI no. | Equipment | Group | Warranty period | CMC Period |
|--------|---|-------|-----------------|--------------|
| 2 | Hydraulic OT Table | 0 | 5 Years | 5 Years |
| 3 | Delivery Table with Footstep | 0 | 5 Years | 5 Years |
| 4 | Ceiling shadow less lights for OT | 0 | 5 Years | 5 Years |
| 5 | Portable shadow less lights for Labour Room | 0 | 5 Years | 5 Years |
| 6 | Examination Table SS with footstep | 0 | 5 Years | not required |
| 7 | X-ray view box | 0 | 5 Years | not required |
| 8 | Transfer Trolley for OT Sterile Zone | 0 | 5 Years | not required |
| 9 | Gynaec electric cautery | 0 | 5 Years | 5 Years |
| 10 | Suction Apparatus high Vacuum | 0 | 5 Years | 5 Years |
| 11 | Suction Apparatus Electrical | 0 | 5 Years | 5 Years |

| SI no. | Equipment | Group | Warranty period | CMC Period |
|--------|--|-------|-----------------|--------------|
| 12 | Mucus Extractor with suction tube and foot operated suction machine | 0 | 5 Years | 5 Years |
| 22 | Multi Para Pulse Oxymeter | A | 5 Years | 5 Years |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre | A | 5 Years | 5 Years |
| 25 | Pediatric Ventilator for ICU | A | 5 Years | 5 Years |
| 27 | Transport Ventilator | A | 5 Years | 5 Years |
| 28 | Transport Ventilator Neonatal | A | 5 Years | 5 Years |
| 29 | Defibrillator | A | 5 Years | 5 Years |
| 43 | ECG Machine | A | 5 Years | 5 Years |
| 50 | Infusion Pump | A | 5 Years | 5 Years |
| 59 | Ambu Bag | A | 1 year | not required |
| 60 | Oxygen Concentrator | A | 5 Years | 5 Years |
| 1 | Syringe & needle cutter | GNL | 1 year | not required |
| 13 | Adult resuscitation kit | GNL | 1 year | not required |
| 14 | Neonatal resuscitation kit | GNL | 1 year | not required |
| 15 | Oxygen Cylinder B Type | GNL | 1 year | not required |
| 16 | Nitrous Oxide Cylinder B Type | GNL | 1 year | not required |
| 17 | Regulator & Flow Meter for Medical Gas | GNL | 1 year | not required |
| 18 | Vento use (for vacuum extraction delivery) | GNL | 1 year | not required |
| 19 | MTP set with karmon suction canula and MVA syringe | GNL | 5 years | not required |
| 20 | Laryngoscope | GNL | 1 year | not required |
| 23 | Nebulizer | GNL | 1 year | not required |
| 36 | Stethoscope for Pediatric as well as Adult | GNL | 1 year | not required |
| 37 | BP Instrument stand type | GNL | 1 year | not required |
| 38 | BP Instrument table top model | GNL | 1 year | not required |
| 39 | Infant BP Instrument | GNL | 1 year | not required |
| 40 | Digital Weighing Machine Adult | GNL | 1 year | not required |
| 41 | Digital Weighing Machine Infant | GNL | 1 year | not required |
| 49 | Digital Thermometer | GNL | 1 year | not required |
| 56 | Glucometer | GNL | 1 year | not required |
| 58 | Height measuring stand | GNL | 1 year | not required |
| 62 | Air curtain at entrance for O.T, SNCU, ISU, Labour Room | GNL | 1 year | not required |
| 33 | Constant Temp Water Bath with Thermometer for Lab | L | 5 Years | 5 Years |
| 55 | Microscope binocular | L | 5 Years | 5 Years |
| 57 | Serum bilirubino meter | L | 1 year | not required |
| 66 | Bacteriological Incubator | L | 5 Years | 5 Years |
| 72 | Semi auto biochemistry analyzer | L | 5 Years | 5 Years |
| 73 | Blood Storage Refrigerator | L | 5 Years | 5 Years |
| 74 | Centrifuge- 12 Tubes | L | 5 Years | 5 Years |
| 75 | Blood Gas Analyzer | L | 5 Years | 5 Years |
| 76 | Hematology Analyzer | L | 5 Years | 5 Years |
| 77 | Blood Cell Counter (Manual) | L | 5 Years | 5 Years |

| Sl no. | Equipment | Group | Warranty period | CMC Period |
|--------|---|-------|-----------------|--------------|
| 21 | Foetal Monitor | P | 5 Years | 5 Years |
| 34 | CPAP Machine with Heated Humidifier | P | 5 Years | 5 Years |
| 35 | Foetal Doppler | P | 1 year | not required |
| 63 | Radiant Warmer | P | 5 Years | 5 Years |
| 64 | Phototherapy Unit | P | 5 Years | 5 Years |
| 65 | Microprocessor Based Incubator (Transport type) | P | 5 Years | 5 Years |
| 67 | Baby Bassinet | P | 5 Years | 5 Years |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | R | 5 Years | 5 Years |
| 30 | Portable Ultrasound | R | 5 Years | 5 Years |
| 31 | 3-D Ultrasound- High End for Diagnostics | R | 5 Years | 5 Years |
| 32 | X-ray | R | 5 Years | 5 Years |
| 42 | Dressing Drum SS of different sizes | X | 1 year | not required |
| 44 | Delivery Instrument Set | X | 1 year | not required |
| 45 | Episiotomy Instrument Set | X | 1 year | not required |
| 46 | MVA/EVA Instrument Set | X | 1 year | not required |
| 47 | PPIUCD Instrument Set | X | 1 year | not required |
| 48 | Outlets forceps | X | 1 year | not required |
| 51 | Caesarean Section | X | 1 year | not required |
| 52 | Suture removal | X | 1 year | not required |
| 53 | Suturing Tray | X | 1 year | not required |
| 54 | Anesthesia trolley and General Anesthesia kit | X | 1 year | not required |
| 61 | Instrument Tray Different Sizes | X | 1 year | not required |
| 68 | High Pressure Sterilizer Horizontal | X | 5 Years | 5 Years |
| 69 | High Pressure Sterilizer Vertical | X | 5 Years | 5 Years |
| 70 | Instrument Sterilizer | X | 5 Years | 5 Years |
| 71 | PN Sterilization | X | 5 Years | 5 Years |

**SECTION-VII
TECHNICAL SPECIFICATIONS**

| Schedule No. 2.Hydraulic OT Table | |
|---|---|
| TENDER SPECIFICATION | READ AS |
| Hydraulic OT Table | Electro Hydraulic OT table. |
| Leg Section: 90 degree down | Leg Section: 45 degree down |
| | |
| Schedule No. 4.Ceiling Shadowless Light for OT | |
| TENDER SPECIFICATION | READ AS |
| Doube Dome | Doube Dome with Major dome should have Illumination of 1,60,000 Lux and minor dome 1,20,000 Lux or more |
| Intensity control in 9 steps for individual domes | Intensity control in 5 or more variable steps for individual domes. |
| Height Adjustment: 600mm | Height Adjustment should be possible |
| Should have over charging cut off with visual symbol | Deleted |
| | |

| Schedule No. 5.Portable shadow less lights for Labour Room | |
|---|--|
| TENDER SPECIFICATION | READ AS |
| LED lights-2 nos | LED Light |
| CR± 95000 | CR± 95 |
| | |
| Schedule No. 10.Suction Apparatus High Vacuum | |
| TENDER SPECIFICATION | READ AS |
| Dimensions to be Max: 43 x 30 x 68 cms | Deleted |
| Should be fitted with wide mouthed 2 x 2 Ltrs. (Polycarbonate) with self sealing bungs and mechanical over flow safety device. | Should be fitted with wide mouthed 2 x 2-3 Ltrs. (Polycarbonate) with self sealing bungs and mechanical over flow safety device. |
| | |
| Schedule No. 21.Foetal Monitor | |
| TENDER SPECIFICATION | READ AS |
| Foetal Dopler, table model with digital display of FHR. The unit shall have S.S tube, Penhold transducer for O.P.D and elastic flat transducer for continuous monitoring. The unit shall be fitted with hi/lo alarms for F.H.R. | Foetal Dopler, table model with digital display of FHR. The unit shall have Penhold transducer for O.P.D and elastic flat transducer for continuous monitoring. The unit shall be fitted with hi/lo alarms for F.H.R. |
| Should have 10" or more high resolution color TFT/LCD display with tilt able screen for better viewing | Should have 6" or more high resolution color TFT/LCD Touch Screen Display with tilt able screen for better viewing |
| Should display the monitoring information of last 24 hours is essential | Deleted |
| | |
| Schedule No. 22.Multi Para Pulse Oximeter | |
| TENDER SPECIFICATION | READ AS |
| Mains electricity (AC-powered) photoelectric device intended for the continuous transcutaneous measurement and display of haemoglobin oxygen saturation (SpO2). The signals, typically produced by light-emitting diodes (LEDs) and a receiving detector in a probe, or directly built-in, are used to make the measurements using the principle of spectrophotometry. The Oxymeter displays the SpO2 values and may calculate/display other parameters, e.g. pulse rate, electrocardiogram (ECG). The device is typically used at the bedside. | Mains electricity (AC-powered) photoelectric device intended for the continuous transcutaneous measurement and display of haemoglobin oxygen saturation (SpO2). The signals, typically produced by light-emitting diodes (LEDs) and a receiving detector in a probe, or directly built-in, are used to make the measurements using the principle of spectrophotometry. The Oxymeter displays the SpO2 values and may calculate/display other parameters, e.g. pulse rate. The device is typically used at the bedside. |
| B) Accuracy of SpO2 better oxymeter** than + 1% for range 40-70 and better than + 3% for range 70-99 | Accuracy of SpO2 better oxymeter** ±4% for Range 40-70 and better than ±3% for Range 70-99 |
| ISO 80601-2-61-2011: Medical Electrical equipment - part 2-61: Particular requirements for the basic safety | Deleted |
| | |
| Schedule No. 37.BP instrument Stand Type | |
| TENDER SPECIFICATION | READ AS |
| Should be portable mercurial type, stand model. | It should be Dial type |
| | |
| Schedule No. 38.BP Instrument Table Top Model | |

| TENDER SPECIFICATION | READ AS |
|--|--|
| Should be Portable mercurial type. | It should be Dial type |
| Schedule No. 39. Infant BP Instrument | |
| TENDER SPECIFICATION | READ AS |
| | Same as table top |
| Schedule No. 40. Digital Weighing Machine Adult | |
| TENDER SPECIFICATION | READ AS |
| Should have Zero adjustment | Deleted |
| Schedule No. 63. Radiant Warmer | |
| TENDER SPECIFICATION | READ AS |
| 50. Baby contact material should be biocompatible as per ISO 10993 standard requirement. | Baby Contact Material should be biocompatible |
| Schedule No. 24. Multi Para Bed Side Monitor (6 parameter) with Central nursing station | |
| TENDER SPECIFICATION | READ AS |
| FDA (US) and EU CE notified body approved and BIS/ISO 13485:2003; IEC-60601-1-2:2007; IEC 60601- | FDA (US) or EU CE notified body approved and BIS/ISO 13485:2003; IEC-60601-1-2:2007; IEC 60601 |
| Heat Dissipation - Should maintain nominal Temp and the heat should be disbursed through an exhaust cooling fan | Heat Dissipation - Should maintain nominal Temp and the heat should be disbursed through an exhaust cooling fan/ Self cooling technology |
| 2. Should have facility for charging from both 12V DC & 220V AC. | Deleted |
| Optional item to be quoted: invasive blood pressure-monitoring module complete with reusable transducer. | Optional item to be quoted: invasive blood pressure-monitoring module complete with disposable transducer. |
| | Added para: 15 monitor to be connected with one CNS Nursing Station. |
| Schedule No. 25. Pediatric Ventilator for ICU | |
| TENDER SPECIFICATION | READ AS |
| Should be convenient and quick USB interface | Should be convenient and quick USB /RS 232/ Memory card Interface |
| Should have compatible hanged arm for holding the circuit | Should have ventilator trolley and hinged arm for holding the circuit of the same Principal company |
| Internal, replaceable, rechargeable battery allows operation for at least four hour in the event of power failure. | Battery back up should be at least 2 hours. |
| Microprocessor control with external compressor based ventilator suitable for neonatal and paediatric ventilation. | Microprocessor control with external/inbuilt/same make compressor based ventilator suitable for neonatal and pediatric ventilation |
| Pressure support | pressure support, volume support |
| | Added Para: Machine should be supplied with Proximal sensor with necessary hardware & software and Servo Controlled Humidifier with Pediatric Chamber (2 Nos), Heater Wire (1No.) & Temperature Sensor (1 No) |

| Schedule No.27.Transport Ventilator | |
|---|---|
| TENDER SPECIFICATION | READ AS |
| a. Respiratory rate - 0 - 60 BPM | Respiratory rate - 5-60 BPM |
| g. Shall have NIV in all modes | Shall have NIV modes |
| a) should be FDA (US) and CE (EU) notified body approved | Should be FDA (US)/CE (EU) notified body approved |
| Schedule No.28.Transport Ventilator Neonatal | |
| TENDER SPECIFICATION | READ AS |
| The device should be capable of operation in various environments such as Emergency, Ambulance, Aircraft, Hospital and MRI | The device should be capable of operation in various environments such as Emergency, Ambulance, Aircraft, Hospital. |
| The device should be MRI conditioned upto 3 Tesla, 430 G/cm | Deleted |
| Standards & Safety: Should be FDA (US) and CE(EU) notified body approved | a) Should be FDA (US)/CE (EU) notified body approved |
| Schedule No.09.Gynaec electric cautery | |
| TENDER SPECIFICATION | READ AS |
| Should have maximum bipolar coagulation power of at least 50 in steps | Should have maximum bipolar coagulation power of at least 50W in steps |
| Schedule No.26.Colour Ultrasound Machine Mobile for Screening Purpose | |
| TENDER SPECIFICATION | READ AS |
| Should have 15" or more high resolution TFT/LCD monitor with tilt and swivel facility and should be able to view in all angles and all light conditions | Should have 15" or more high resolution TFT/LCD/LED monitor with tilt and swivel facility and should be able to view in all angles and all light conditions |
| Should have 4-9 MHz broadband side firing Trans-Rectal Probe | Should have 4-9 MHz broadband side firing Trans-Rectal Probe (Optional, Price to be quoted seperately) |
| Voltage corrector/stabilizer to allow operation at +/- 15% of local rated voltage. Use of SMPS to correct Voltage | Deleted |
| Schedule No. 44,45,46,47,48,19,71,51,52: Delivery Instrument set ,Episotomy Instrument set,MVA/EVA instrument set,PPIUCD Instrument set,Outlet forceps,MTP set with karmons suction canula and MVA syringe,PN sterilisation,Caesarean section,Suture removal | |
| TENDER SPECIFICATION | READ AS |
| • ISI certification | ISI/CE/FDA/ISO 13485 |
| Schedule No.60.Oxygen Concentrator | |
| TENDER SPECIFICATION | READ AS |
| Sound Level < 15dB | Sound Level < 50dB |
| Schedule No.65.Microprocessor Based Incubator (Transport Type) | |
| TENDER SPECIFICATION | READ AS |
| Tolerance (to variations, Voltage corrector/ stabilizer/UPS to allow operation at +/- 30% of local shutdowns) rates voltage | Machine should have internal voltage corrector to allow operation at +/- 30% of local rates voltage |

| Schedule No.75.Blood Gas Analyzer | |
|---|--|
| TENDER SPECIFICATION | READ AS |
| | Added Para Cost of all consumables (including reagents, Cartridge, paper, electrode if applicable) to be quoted for comparative evaluation. Consumables for two year @ at least 10 samples/day for all tests should be quoted and it will be taken for price comparison. |
| | |
| Schedule No.76.Hematology Analyzer | |
| TENDER SPECIFICATION | READ AS |
| Principles : Electrical impedance method with advanced SRV technology for better accuracy and precision | Principles: Electrical impedance method with advanced SRV technology/ liquied valves/Syringe Based for better accuracy and precision. |

Section – IX

Qualification Criteria to be read as:

- Bidder or manufacturer** should submit end user certificate(s) for having executed at least 25% of the tendered quantity during the last Five years from the date of tender opening of similar equipment meeting major parameters of technical specification which is functioning satisfactorily, anywhere in India **of the same manufacturer** for items mentioned at **table-1** below with respect to the relevant group(s) for their submitted bid.

Table-1

| Sl no. | Equipment | Group | Qty for 20 hospital |
|--------|--|-------|---------------------|
| 2 | Hydraulic OT Table | 0 | 80 |
| 4 | Ceiling shadow less lights for OT | 0 | 40 |
| 5 | Portable shadow less lights for Labour Room | 0 | 120 |
| 9 | Gynaec electric cautery | 0 | 40 |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre | A | 20 |
| 25 | Pediatric Ventilator for ICU | A | 40 |
| 27 | Transport Ventilator | A | 20 |
| 28 | Transport Ventilator Neonatal | A | 20 |
| 29 | Defibrillator | A | 40 |
| 50 | Infusion Pump | A | 120 |
| 72 | Semi auto biochemistry analyzer | L | 20 |
| 73 | Blood Storage Refrigerator | L | 20 |
| 75 | Blood Gas Analyzer | L | 20 |
| 76 | Hematology Analyzer | L | 20 |
| 77 | Blood Cell Counter (Manual) | L | 20 |
| 63 | Radiant Warmer | P | 120 |
| 65 | Microprocessor Based Incubator (Transport type) | P | 20 |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | R | 40 |
| 30 | Portable Ultrasound | R | 20 |
| 31 | 3-D Ultrasound- High End for Diagnostics | R | 20 |
| 32 | X-ray | R | 20 |
| 68 | High Pressure Sterilizer Horizontal | X | 20 |
| 69 | High Pressure Sterilizer Vertical | X | 40 |

2. Bidder should have an annual average turnover during the last 3 years for an a minimum amount as mentioned in the **table 2** below against the group of items there bidding for. The cumulative turnover of the groups shall be considered for qualifying in case the bidder is quoting for more than one group.

Table-2

| Group | Minimum Average Turn Over Criteria (In Rs.) |
|-------|---|
| O | 528,75,000 |
| A | 601,25,000 |
| GNL | 107,55,000 |
| L | 94,92,500 |
| P | 216,50,000 |
| R | 445,00,000 |
| X | 410,00,000 |

3. Bidders have to submit the Manufacturer authorization form in the prescribed format mentioned at section XII for the items in the table below against their offered group of items.

| Sl no. | Equipment | Group | Qty for 20 hospital |
|--------|--|-------|---------------------|
| 2 | Hydraulic OT Table | 0 | 80 |
| 4 | Ceiling shadow less lights for OT | 0 | 40 |
| 5 | Portable shadow less lights for Labour Room | 0 | 120 |
| 9 | Gynaec electric cautery | 0 | 40 |
| 24 | Multipara bed side monitor (6 parameter) with central Nursing Centre | A | 20 |
| 25 | Pediatric Ventilator for ICU | A | 40 |
| 27 | Transport Ventilator | A | 20 |
| 28 | Transport Ventilator Neonatal | A | 20 |
| 29 | Defibrillator | A | 40 |
| 50 | Infusion Pump | A | 120 |
| 72 | Semi auto biochemistry analyzer | L | 20 |
| 73 | Blood Storage Refrigerator | L | 20 |
| 75 | Blood Gas Analyzer | L | 20 |
| 76 | Hematology Analyzer | L | 20 |
| 77 | Blood Cell Counter (Manual) | L | 20 |
| 63 | Radiant Warmer | P | 120 |
| 65 | Microprocessor Based Incubator (Transport type) | P | 20 |
| 26 | Colour Ultrasound Machine Mobile- For Screening Purpose | R | 40 |
| 30 | Portable Ultrasound | R | 20 |
| 31 | 3-D Ultrasound- High End for Diagnostics | R | 20 |
| 32 | X-ray | R | 20 |
| 68 | High Pressure Sterilizer Horizontal | X | 20 |
| 69 | High Pressure Sterilizer Vertical | X | 40 |

4. The start-ups claiming exemption on the required prior experience, and complying the condition of GIT Clause 23, should furnish along with the bid
- I. All necessary documents in support of the claim regarding exemption on prior experience as mandated by concerned Ministry/ Board of Govt. of India.

Notwithstanding anything stated above, the Purchaser reserves the right to verify/ consider, whether the firm/ entity is eligible for exemption regarding prior experience requirement.

NOTE:

1. The tenderer shall give an affidavit as under:
“We hereby certify that if at any time, information furnished by us is proved to be false or incorrect, we are liable for any action as deemed fit by the purchaser in addition to forfeiture of the earnest money.”
2. In support of 1 , the Tenderer shall furnish Performance statement in the enclosed Proforma ‘A’.
The bidder shall furnish Satisfactory Performance Certificate from the end user in respect of above, duly translated in English and duly notarized in the country of origin, alongwith the tender.
3. The Tenderer shall furnish a brief write-up, packed with adequate data explaining and establishing his available capacity/capability (both technical and financial) to perform the Contract (if awarded) within the stipulated time period, after meeting all its current/present commitments. The Tenderer shall also furnish details of Equipment and Quality Control in the enclosed Section VIII.
4. Notwithstanding anything stated above, the Purchaser reserves the right to assess the Tenderer’s capability and capacity to perform the contract satisfactorily before deciding on award of Contract, should circumstances warrant such an assessment in the overall interest of the Purchaser.
5. The Purchaser reserves the right to ask for a free demonstration of the quoted equipment at a pre determined place acceptable to the purchaser for technical acceptability as per the tender specifications, before the opening of the Price Tender.

SECTION - XI

Price Schedule to be read as:

PRICE SCHEDULE

| | |
|--------------------|--|
| NIB Ref./ RFx No. | |
| Name of the bidder | |

| Price Format - A (in case of goods and/or services offered in INR) | | | | | | | | | |
|--|--|-------------|-------------------|----------|--|--------------------|------------------------------------|------------------------------------|---------------------------------------|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 |
| SI No | Description of goods as per specification and/or BOQ | Make/ Model | Country of Origin | Quantity | Unit Price at Consignee Site (excluding GST) | Applicable GST (%) | Applicable GST value/ unit (6 x 7) | Unit Price at Consignee Site (6+8) | Total Price at Consignee Site (5 x 9) |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | Sum Total of Column 10 | |

Total value of above offer in INR consideration: Rs. _____

(Total value in words):

- Note:**
- All the information must be entered in the relevant columns
 - The bidder may participate and submit their bids for any one or more groups in this tender. However, bidder has to quote for all the items (providing individual price of each line item) in the group(s) quoted for. In case any line item is not offered while quoting for a group of items the bid for that group shall be rejected without evaluating further. The group wise comparison/ranking of bids shall be determined for consideration of award based on the total value (including all taxes & duties and any other charges incurred till Consignee site) of a group of items.**
 - All the items under this tender must be quoted in Indian Rupees only. There will not be any CDEC issued against these items.

| Price Format - B (price schedule for Comprehensive Annual Maintenance Contract after warranty period) | | | | | | | | | |
|---|------------------------------------|-------------------------------------|---|----------|----------|----------|-----------|--|---------------------------------|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 |
| Sl. No. | Short description of the equipment | Quantity as per List of Requirement | Year wise charges per unit for annual comprehensive maintenance contract in INR | | | | | Charges for 5 years per unit (4+5+6+7+8) | Total Charges for 5 years 3 x 9 |
| | | | 6th year | 7th year | 8th year | 9th year | 10th year | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |

GST: (included or extra to be shown here)

(Total charges in words):

Note:

1. All the information must be entered in the relevant columns
2. The cost of CAMC may be quoted along with GST applicable on the date of Tender Opening. The GST to be paid extra, to be specifically stated. In the absence of any later such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained
3. Cost of CMC offered will be added in its Net Present Value (at a discounted rate of 10% per year) for Ranking/Evaluation purpose.
4. Items wise CMC cost must be quoted

SECTION - XII
MANUFACTURER'S AUTHORISATION FORM to be read as:

Team Leader- Admin & Commercial
HLL Mother & Child Care Hospitals Limited

Dear Sir,

Ref: Your TE document No _____ dated _____

We, _____ who are proven and reputable manufacturers of _____ (*name and description of the goods offered in the tender*) having factories at _____, hereby authorise Messrs _____ (*name and address of the agent*) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):

We also hereby extend our full warranty, CMC as applicable as per clause 15 of the General Conditions of Contract for the goods and services offered for supply by the above firm against this TE document.

We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorized agent.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly"

Yours faithfully,

[*Signature with date, name, designation and Email*] for
and on behalf of Messrs _____

[*Name & address of the manufacturers*]

Note:

- (1) *This letter of authorisation should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.*
- (2) *Original letter may be sent.*
- (3) *The purchaser reserves the right to verify this document with its signatory.*

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the website regularly prior to the closing date and time of submission of tenders.