

Amendment No: 1

Ref: Tender Enquiry No: HLL/PCD/ICCONS-02/12-13 Dtd. 23.01.2013

Pre-Bid meeting against the referred tender was held on 31/01/2013 at HLL Lifecare Ltd., Thycaud, Trivandrum-695 014

The following amendments are issued with respect to the above Tender Enquiry Document.

The address for submitting the tenders have been changed to:

HLL LIFECARE LTD.
TC 24/606, TENRA - 22
PALATHINKARA, THYCAUD
TRIVANDRUM - 695 014, KERALA.
PH: 0471-2330447

Technical Specification

Item No. 1

ICU VENTILATOR

Existing para:

2.15 (a) Stand-alone Medical Air compressor

Amended as:

2.15 (a) **Stand-alone / integrated** Medical Air compressors

Item No. 2

MULTIPARAMETER MONITOR

Existing para:

2.8) Should provide hemodynamic, oxygenation, Ventilation calculation package.

2.9) Should have drug calculation package.

2.12) Automatic Zoom in Facility in the monitor display.

Amended as:

2.8) clause deleted

2.9) clause deleted

2.12) clause deleted

Item No. 3

PULSE OXIMETER

Existing para:

2.10) Battery back-up operating time 5 hours.

Amended as:

2.10) Battery back-up operating time of **at least 2 hours**.

Item No. 5

DESKTOP NEXT GENERATION SEQUENCER

Existing para:

Chemistry: Proven chemistry required; no Homopolymer errors, ability to Detect Variants. Reagents should be premixed and preloaded in a reagent cartridge.

Amended as:

Chemistry: Proven chemistry required; no Homopolymer errors, ability to Detect Variants. **Specify type of reagent packaging.**

“Enclose separate quote for various possible reagents along with the price bid. This offer should be valid for 2 years after warranty.” (*However, these prices shall not be taken into consideration for comparing the price bids*)

Existing para:

System should use sequencing by synthesis technology.

Amended as:

System should elaborate the sequencing technique to allow detailed specific evaluation (*Provide necessary documents*).

Existing para:

Routine consumables for approximately 1 year with an average of 100 samples should be supplied free of cost.

Amended as:

Routine consumables for approximately 1 year with an average of 100 samples should be supplied free of cost (*Specify various possible combination*).

All other contents, terms & conditions of the tender enquiry remain unchanged.