27.09-2018

Amendment No. 2

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/03/18-19 dated 14/09//2018

The following changes are being incorporated in the above referred Tender Enquiry Document.

Section I Notice Inviting Tender

Existing:

<u>SI.</u> <u>No</u>	Name of the Item	<u>Department</u>	<u>Unit</u>	<u>Tendered</u> <u>Qty</u>	<u>Tender Fee</u> <u>in INR</u>	EMD in INR
1	Defibrillator with Monitor	General OPD	No.	5	590	45,000
2	Spirometer	General OPD	No.	5	590	30,000
3	Basic ENT examination units with motorised patient chairs and microscope	ENT OPD	No.	10	2,360	2,00,000
4	ENT Workstation	ENT OPD	No.	5	3,540	5,00,000
5	Endoscopes Set	ENT OPD	No.	15	3,540	7,50,000
6	Farnsworth Munsel 100 Hue Test	Ophthalmology OPD	No.	5	590	20,000
7	Slit lamp	Ophthalmology OPD	No.	10	1,180	60,000
8	Photo Slit lamp	Ophthalmology OPD	No.	5	1,180	1,20,000
9	Non Contact Tonometer with Pachymetry	Ophthalmology OPD	No.	10	1,180	1,00,000
10	Digital Fundus Camera	Ophthalmology OPD	No.	5	2,360	1,50,000
11	Automated Perimeter	Ophthalmology OPD	No.	5	1,180	80,000
12	Indirect Ophthalmoscope	Ophthalmology OPD	No.	10	590	50,000
13	USG A+B Scan	Ophthalmology OPD	No.	5	1,180	1,00,000
14	Auto Ref Keratometer	Ophthalmology OPD	No.	5	590	50,000
15	Refraction Chair unit	Ophthalmology OPD	No.	10	590	40,000
16	ND Yag Laser(1064nm)	Ophthalmology OPD	No.	5	2,360	1,50,000
17	ND Yag Laser(532nm)	Ophthalmology OPD	No.	5	3,540	3,00,000
				TOTAL	27,140	27,45,000

Read as:

<u>SI.</u> <u>No</u>	RFX No.	Name of the Item	<u>Department</u>	<u>Unit</u>	<u>Tendered</u> <u>Qty</u>	<u>Tender Fee</u> <u>in INR</u>	EMD in INR
1	3000003424	Defibrillator with Monitor	General OPD	No.	5	590	45,000
2	3000003425	Spirometer	General OPD	No.	5	590	30,000
3	3000003426	Basic ENT examination units with motorised patient chairs and microscope	ENT OPD	No.	10	2,360	2,00,000
4	3000003427	ENT Workstation	ENT OPD	No.	5	3,540	5,00,000
5	3000003428	Endoscopes Set	ENT OPD	No.	15	3,540	7,50,000
6	3000003429	Farnsworth Munsel 100 Hue Test	Ophthalmology OPD	No.	5	590	20,000
7	3000003430	Slit lamp	Ophthalmology OPD	No.	10	1,180	60,000
8	3000003431	Photo Slit lamp	Ophthalmology OPD	No.	5	1,180	1,20,000
9	3000003432	Non Contact Tonometer with	Ophthalmology OPD	No.	10	1,180	1,00,000

<u>SI.</u> <u>No</u>	RFX No.	Name of the Item	<u>Department</u>	<u>Unit</u>	<u>Tendered</u> <u>Qty</u>	<u>Tender Fee</u> <u>in INR</u>	EMD in INR
		Pachymetry					
10	3000003433	Digital Fundus Camera	Ophthalmology OPD	No.	5	2,360	1,50,000
11	3000003434	Automated Perimeter	Ophthalmology OPD	No.	5	1,180	80,000
12	3000003435	Indirect Ophthalmoscope	Ophthalmology OPD	No.	10	590	50,000
13	3000003436	USG A+B Scan	Ophthalmology OPD	No.	5	1,180	1,00,000
14	3000003437	Auto Ref Keratometer	Ophthalmology OPD	No.	5	590	50,000
15	3000003438	Refraction Chair unit	Ophthalmology OPD	No.	10	590	40,000
16	3000003439	ND Yag Laser(1064nm)	Ophthalmology OPD	No.	5	2,360	1,50,000
17	3000003440	ND Yag Laser(532nm)	Ophthalmology OPD	No.	5	3,540	3,00,000
			TOTAL		_	27,140	27,45,000

Section VI List of Requirement

Existing:

<u>SI.</u> <u>No</u>	Name of the Item	<u>Department</u>	<u>Unit</u>	<u>Tendered</u> <u>Qty</u>	<u>Warranty</u>	<u>CMC</u> <u>Required or</u> <u>Not</u>
1	Defibrillator with Monitor	General OPD	No.	5	5 years	Yes
2	Spirometer	General OPD	No.	5	5 years	Yes
3	Basic ENT examination units with motorised patient chairs and microscope	ENT OPD	No.	10	5 years	Yes
4	ENT Workstation	ENT OPD	No.	5	5 years	Yes
5	Endoscopes Set	ENT OPD	No.	15	2 years	No
6	Farnsworth Munsel 100 Hue Test	Ophthalmology OPD	No.	5	2 years	No
7	Slit lamp	Ophthalmology OPD	No.	10	2 years	Yes
8	Photo Slit lamp	Ophthalmology OPD	No.	5	5 years	Yes
9	Non Contact Tonometer with Pachymetry	Ophthalmology OPD	No.	10	5 years	Yes
10	Digital Fundus Camera	Ophthalmology OPD	No.	5	5 years	Yes
11	Automated Perimeter	Ophthalmology OPD	No.	5	5 years	Yes
12	Indirect Ophthalmoscope	Ophthalmology OPD	No.	10	5 years	Yes
13	USG A+B Scan	Ophthalmology OPD	No.	5	5 years	Yes
14	Auto Ref Keratometer	Ophthalmology OPD	No.	5	5 years	Yes
15	Refraction Chair unit	Ophthalmology OPD	No.	10	2 years	Yes
16	ND Yag Laser(1064nm)	Ophthalmology OPD	No.	5	2 years	Yes
17	ND Yag Laser(532nm)	Ophthalmology OPD	No.	5	2 years	Yes

Read as:

SI. No	RFX No.	Name of the Item	<u>Department</u>	<u>Unit</u>	Tendered Qty	<u>Warranty</u>	CMC Required or Not
1	3000003424	Defibrillator with Monitor	General OPD	No.	5	5 years	Yes
2	3000003425	Spirometer	General OPD	No.	5	5 years	Yes
3	3000003426	Basic ENT examination units with motorised patient chairs and microscope	ENT OPD	No.	10	5 years	Yes
4	3000003427	ENT Workstation	ENT OPD	No.	5	5 years	Yes
5	3000003428	Endoscopes Set	ENT OPD	No.	15	2 years	No
6	3000003429	Farnsworth Munsel 100 Hue Test	Ophthalmology OPD	No.	5	2 years	No
7	3000003430	Slit lamp	Ophthalmology OPD	No.	10	2 years	Yes

SI. No	RFX No.	Name of the Item	<u>Department</u>	<u>Unit</u>	Tendered Qty	<u>Warranty</u>	CMC Required or Not
8	3000003431	Photo Slit lamp	Ophthalmology OPD	No.	5	5 years	Yes
9	3000003432	Non Contact Tonometer with Pachymetry	Ophthalmology OPD	No.	10	5 years	Yes
10	3000003433	Digital Fundus Camera	Ophthalmology OPD	No.	5	5 years	Yes
11	3000003434	Automated Perimeter	Ophthalmology OPD	No.	5	5 years	Yes
12	3000003435	Indirect Ophthalmoscope	Ophthalmology OPD	No.	10	5 years	Yes
13	3000003436	USG A+B Scan	Ophthalmology OPD	No.	5	5 years	Yes
14	3000003437	Auto Ref Keratometer	Ophthalmology OPD	No.	5	5 years	Yes
15	3000003438	Refraction Chair unit	Ophthalmology OPD	No.	10	2 years	Yes
16	3000003439	ND Yag Laser(1064nm)	Ophthalmology OPD	No.	5	2 years	Yes
17	3000003440	ND Yag Laser(532nm)	Ophthalmology OPD	No.	5	2 years	Yes

Section VII Technical Specification

Sch. No. 03 ENT examination units with motorised patient chairs and microscope

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 45 Para 1	ENT Treatment Unit with Binocular Microscope made of High Quality steel casing with polyurethane covering for two or three storied large instrument surface with aluminum trays, unbreakable glass dust cover and built in waste container. Condensation discharge and liquid bottles should be easily accessible	ENT Treatment Unit with Binocular Microscope made of High Quality steel casing with polyurethane covering for two or three storied large instrument surface with aluminum trays/ SS 304 trays, unbreakable glass/ Transparent unbreakable acrylic dust cover and built in waste container. Condensation discharge and liquid bottles should be easily accessible
Pg 45 Para 7	Suction pump: 3000cc	Suction bottle: 2000cc or more
Pg 45 Para 8	Light source: 150W X 24V with minimum 4 outlets.	Light source LED with 4 light source
Pg 45 Para 9	Head light with adjustable head band (halogen) with suitable fibro-optic cable	Head light with adjustable head band (LED) with suitable fibro-optic cable
Pg 46 Para 19	Height: from upper edge of the back rest to the lowest position 1230cms (approx)	Height: from upper edge of the back rest to the lowest position 900mm (approx)

Sch. No. 04 ENT Workstation

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 46 Para 2	Minimum 02 large pull-out trays with dividers located in compartment under hinged cover	Minimum 02 large pull-out trays with dividers located in compartment.
Pg 47 Para 14 b	Capacity: 55 ltr/min to 70 ltr/min, vacuum: -80 to -98kpa, with secretion collection of 1ltr/1.5ltr with automatic secretion canister evacuator system.	Capacity: 40 ltr/min to 60 ltr/min, vacuum: -80 to -98kpa, with secretion collection of 1ltr/1.5ltr with automatic secretion canister evacuator system.
Pg 47 Para 15 g	Max. pressure: 0.1 to 4 bar	Max. pressure: 0.2 to 2.5 bar
Pg 47	Suitable fibroptic chords or light	Suitable fibroptic chords or light carriers for endoscopes.

Tender Page No. & Para	TENDER SPECIFICATION	Read As
	carriers for endoscopes. Disinfection and warming quivers for endoscopes preferably placed in extension arm (preferably for 5 endoscopes). Preheating quivers disinfection quivers, disinfection quivers, disinfection time control, removable and disinfectable quivers for flexible scopes. Both USFDA and European CE approved and inclusion subject to demonstration if asked for by experts.	Disinfection and warming quivers for endoscopes preferably placed in extension arm (preferably for 5 endoscopes). Preheating quivers disinfection quivers, disinfection quivers, disinfection time control, removable and disinfectable quivers for flexible scopes. USFDA or BIS or European CE with 4 digit no. issued from a notified body approved and inclusion subject to demonstration if asked for by experts.
Pg 47 Para 17 d	Should preferably have a separate stainless steel tank.	Should preferably have a separate stainless steel / plastic tank.
Pg 48 Para 20 p	16" TFT monitor, LCD flat screen. Not same as in point number 12.	Atleast 16" TFT monitor, LCD flat screen. Not same as in point number 12.
Pg 48 Para 21	Both USFDA and European CE approved and inclusion subject to demonstration if asked for by experts.	USFDA or BIS or European CE with 4 digit no. issued from a notified body
Pg 48 Para 23 b	Approximately 300 cm fibroptic chord.	180 cm or more fibroptic chord.
Pg 49 Para 25 i	Should confirm US FDA and European standards.	USFDA or BIS or European CE with 4 digit no. issued from a notified body

Sch. No. 05 Endoscopes Set

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 49	Oto Endoscope Strait forward Telescope	Oto Endoscope Strait forward Telescope 0 ° 1.9 mm 10cm-12cm
Para 1	0 ° 1.9 mm 10cm length	length - 1 No.
Pg 49	Nasal Endoscope Rigid Telescope 0 ° 4	Nasal Endoscope Rigid Telescope 0 ° 4 mm 16cm-18cm length -
Para 2	mm 18cm length	1 No.
Pg 49	Nasal Endoscope Rigid Telescope 70 ° 4	Nasal Endoscope Rigid Telescope 70 ° 4 mm 16cm-18cm length
Para 3	mm 18cm length	- 1 No.
Pg 49	Nasal Endoscope Rigid Telescope 30 ° 4	Nasal Endoscope Rigid Telescope 30 ° 4 mm 16cm-18cm length
Para 4	mm 18cm length	- 1 No.
Pg 49	Florible Nasanharungaal Larunga Coope	Flexible Nasopharyngeal-Laryngo Scope dia 3.5-4mm and
Para 5	Flexible Nasopharyngeal-Laryngo Scope	working length 300-350mm
Pg 49	Cold light fountain 150 watt	Cold light fountain minimum 100 watt LED
Para 7	halogen/LED/Xenon	Cold light lountain minimum 100 watt LED
Pg 49 Para 9	Compatible endoscope HD (720p) camera and LED monitor of 20"or more for image display.	Compatible endoscope atleast HD (720p) camera and LED monitor of 15" or more for image display.

Sch. No. 06 Farnsworth Munsel 100 Hue Test

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 50 Para 1.4	PC to be included with following configuration: 4 GB RAM,500GB HDD, 19" LCD monitor, DVD Drive, latest OS	PC to be included with following minimum configuration: 4 GB RAM,500GB HDD, 19" LCD monitor, DVD Drive, latest OS and online UPS with atleast 30 minutes backup.

Tender Page No. & Para	TENDER SPECIFICATION	Read As
	and UPS	

Sch. No. 07 Slit lamp

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 50	Motorized imported table for slit lamp,	Motorized table for slit lamp, wheel chair friendly from same
Para 14	wheel chair friendly	principal company.
		Added Para: Online UPS with atleast 30 minutes backup.
		Added Para: USFDA or BIS or European CE with 4 digit no.
		issued from a notified body

Sch. No. 08 Photo Slit lamp

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 50	Motorized imported table for slit lamp,	Motorized table for slit lamp, wheel chair friendly from same
Para 14	wheel chair friendly	principal company.
Pg 50 Para 16	Should be supplied with integrated HD camera(Camera and Slit lamp should be from same make) with suitable software for image processing and storage along with branded desktop with table	Should be supplied with integrated HD camera (Any reputed DSLR camera) with suitable software for image processing and storage along with branded desktop with table
		Added Para: USFDA or BIS or European CE with 4 digit no. issued from a notified body

Sch. No. 09 Noncontact Tonometer with Pachymetry

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 51	To measure intraocular pressure without	To measure intraocular pressure and center corneal thickness
Para 2	actual eye contact	without actual eye contact
Pg 51	Digital display of intraocular pressure	Digital display of intraocular pressure and center corneal
Para 3		thickness
		Added Para: Online UPS with atleast 30 minutes backup.

Sch. No. 10 Digital Fundus Camera

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 51 Para 4	Monitor 15 inches LCD for direct display	Monitor atleast 15 inches LCD for direct display
		Added Para: Online UPS with atleast 30 minutes backup.

Sch. No. 11 Automated Perimeter

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 51	C.D. drive, internal hard disk drive with	USB / C.D. drive, internal hard disk drive with Magneto Optical
Para 3	Magneto Optical Disk (MOD) drive.	Disk (MOD) drive.
Pg 52	Should be USFDA or European CE	USFDA or BIS or European CE with 4 digit no. issued from a
Para 19	approved product.	notified body
		Added Para: Online UPS with atleast 30 minutes backup.

Sch. No. 12 Indirect Ophthalmoscope (Wireless)

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 52	Should be US FDA/ European CE issued	USFDA or BIS or European CE with 4 digit no. issued from a
Para 19	by four digit notified body/ BIS approved	notified body

Sch. No. 13 USG A+B Scan

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 52	B-Scan:	B-Scan:
Para 1	Probe 10 MHz, Focused Transducer, 30 Frames/Sec.	Probe 10 MHz, Focused Transducer, 25 / 30 Frames/Sec.

Sch. No. 14 Auto Ref-Keratometer

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 53	Should be US FDA/ European CE issued by four digit notified body/ BIS approved	USFDA or BIS or European CE with 4 digit no. issued from a notified body
		Added Para: Online UPS with atleast 30 minutes backup.

Sch. No. 16 ND Yag Laser (1064nm)

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 54 Para 7	Energy levels: 22 steps	Energy levels: at least 22 steps or continuous variable
Pg 54 Para 12	Aiming beam focus offset +/- 150 µm posterior & anterior focus shift.	Aiming beam focus offset atleast +/- 150 µm posterior & anterior focus shift.
Pg 54 Para 16	Illumination : Halogen12V/30W;	Illumination : Halogen
Pg 54 Para 19	Should be USFDA or European CE	USFDA or BIS or European CE with 4 digit no. issued from a
Faid 19	approved product	notified body Added Para: Online UPS with atleast 30 minutes backup.

Sch. No. 17 ND Yag Laser (532nm)

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 55 Para 18	Mainster and PRP lenses	Mainster PRP and focal laser lenses
Pg 55 Para 19	UPS 1000 VA	Online UPS with atleast 30 minutes backup.
Pg 55 Para 21	Equipment should be European CE with 4 digit Notified body number /US FDA/BIS approved	USFDA or BIS or European CE with 4 digit no. issued from a notified body

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids