## **Ammendment No: 1**

Date: 16/01/2013

Dated: 16/01/2013

### Ref: Tender No.: HLL/PCD/DMER-06/12-13 dated 28.12.2012

With reference to the pre-bid meeting held on 04.01.2013 at HLL Lifecare Limited, Noida the following changes have been incorporated in the referred tender:

## SECTION I: NOTICE INVITING TENDERS (NIT); Para 2: Important events and dates:

### Existing:

-0.			
vi.	Closing date & time for receipt of Tender	25.01.2013 at 2:00pm (IST)	
vii.	Time and date of opening of Techno – Commercial tenders	25.01.2013 at 2:30pm (IST)	

### Amended as:

vi.	Closing date & time for receipt of Tender	28.01.2013 at 2:00pm (IST)
vii.	Time and date of opening of Techno – Commercial tenders	28.01.2013 at 2:30pm (IST)

## **SECTION – VII: TECHNICAL SPECIFICATIONS**

### For item no. 06: Modular Multipara Patient Monitor with Central Monitoring Station

### Existing:

3.1 It should have a high resolution large display of 12" flat colour LCD/TFT with numeric/graphic display of parameters.

#### Amended as:

3.1 It should have a high resolution large display of **15" Touch Screen** flat colour LCD/TFT with numeric/graphic display of parameters.

### Existing:

3.3 It should have monitoring capability of ECG, SPo2, NIBP, respiration rate and temperature (skin/rectal)

# Amended as:

3.3 It should have monitoring capability of ECG, **SPO2** (with Nelcore/Masimo technology), NIBP, respiration rate and temperature (skin/rectal).

## Existing:

- 3.12 It should have option of multiple operating modes
  - a) Monitor mode
  - b) Diagnostic mode
  - c) Surgery mode

#### Amended as:

- 3.12 It should have option of multiple operating modes
  - a) Monitor mode and b) Diagnostic mode & Surgery mode or its equivalent.

# Existing:

3.14 It should have 72 hours of trend storage of all parameters.

#### Amended as:

3.14 It should have **24 hours (or more)** of trend storage of all parameters.

## Existing:

4.3 It should have transmission by Ethernet (wired), Wi-Fi (wireless) and facility for integration into hospital management system.

## Amended as:

4.5 It should have transmission by Ethernet (wired) and facility for integration into hospital management system.

### Existing:

4.5 It should have simultaneous visualization of data received from 10 bed side patient monitors on LCD.

### Amended as:

4.5 It should have simultaneous visualization of data received from **8** bed side patient monitors (**upgradable to 16** bedside monitors) on the Central Station Monitor.

All the other contents of the tender enquiry shall remain unchanged.

-- End --

Dated: 16/01/2013