Amendment No. 2

06.09.2018

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/24/18-19 dated 26.07.2018 read with its Amendment No. 1 dated 28.08.18

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION - VII

TECHNICAL SPECIFICATION AND GENERAL POINTS

A. TECHNICAL SPECIFICATION:

Item No. 1 (Rfx/ Event number 3000003190)

Chemport Insertion Set

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 46, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified
2	page 46, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 46, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ SI. No 7, 8 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
5	Page 46, BOQ point 4	Mosquito forceps curved	Mosquito forceps curved 12-14 cm.
6	Page 46, BOQ point 5	Artery forceps curved	Artery forceps curved 14-16 cm
7	Page 46, BOQ point 6	Allis forceps	Allis forceps 14-16 cm
8	Page 46, BOQ point 9	Eyelid retractor	Eyelid retractor 14-16 cm
9	Page 46, BOQ point 14	Metzenbaum scissors	Metzenbaum scissors 15-17 cm

10	Page 46, BOQ point 15	Mayo scissors	Mayo scissors 15-17 cm
11	Page 46, BOQ point 17	Sponge holder	Sponge holder 22-24 cm

Item No. 2 (Rfx/ Event number 3000003191)

Trachestomy Set

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 48, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified
2	page 48, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 48, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ Sl. No 2,7,9,18, are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
4	Page 48, BOQ point 6	Iris scissor curved	Iris scissor curved 11-13cm
5	Page 48, BOQ point 7	Iris Scissor straight	: Iris Scissor straight 11-13cm
6	Page 48, BOQ point 8	Mayo Scissors	Mayo Scissors 15-17cm
7	Page 48, BOQ point 10	Non-toothed forceps	Non-toothed forceps 15-17cm
8	Page 48, BOQ point 11	S.S Bowl (small)	Deleted
9	Page 48, BOQ point 12	S.S tray (Medium)	Deleted
10	Page 48, BOQ point 13	Sponge holder	Sponge holder 24-26cm
11	Page 48, BOQ point 14	Toothed dissections forceps	Toothed dissections forceps 15-17cm
12	Page 48, BOQ point 17	Tracheal dilator	Tracheal dilator 14-16cm
13	Page 48, BOQ point 18	Tracheal retractor 1 Prong (Tracheal hook)	Tracheal retractor 1 Prong (Tracheal hook) 16-18cm
14	Page 48, BOQ point 19	Tracheal retractor 2 Prong	Tracheal retractor 2 Prong 16- 18cm
15	Page 48, BOQ point 20	Tracheal retractor double prong	Tracheal retractor double prong 16-18cm

Item No. 3 (Rfx/ Event number 3000003191)

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 48, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified
2	page 48, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 48, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ SNo 2,7,9,18, are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
4	Page 48, BOQ point 6	Iris scissor curved	Iris scissor curved 11-13cm
5	Page 48, BOQ point 7	Iris Scissor straight	Iris Scissor straight 11-13cm
6	Page 48, BOQ point 8	Mayo Scissors	Mayo Scissors 15-17cm
7	Page 48, BOQ point 10	Non-toothed forceps	Non-toothed forceps 15-17cm
8	Page 48, BOQ point 11	S.S Bowl (small)	Deleted
9	Page 48, BOQ point 12	S.S tray (Medium)	Deleted
10	Page 48, BOQ point 13	Sponge holder	Sponge holder 24-26cm
11	Page 48, BOQ point 14	Toothed dissections forceps	Toothed dissections forceps 15-17cm
12	Page 48, BOQ point 17	Tracheal dilator	Tracheal dilator 14-16cm
13	Page 48, BOQ point 18	Tracheal retractor 1 Prong (Tracheal hook)	Tracheal retractor 1 Prong (Tracheal hook) 16-18cm
14	Page 48, BOQ point 19	Tracheal retractor 2 Prong	Tracheal retractor 2 Prong 16- 18cm
15	Page 48, BOQ point 20	Tracheal retractor double prong	Tracheal retractor double prong 16-18cm

Item No. 4 (Rfx/ Event number 3000003192)

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 49, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO
	point 2	FDA 61 CE 61 BIS certified	13485 certified
2	page 49, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 49, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ SNo .3,6,14,15 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
4	Page 49, BOQ point 8	Sponge Holder	Sponge Holder 24-26cm
5	Page 49, BOQ point 9	Needle holder	Needle holder 15-17cm
6	Page 49, BOQ point 13	Small Bowl	S.S. Bowl Small

Item No. 5 (Rfx/ Event number 3000003193)

Minor Surgical Set

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 50,	The Instruments Should be	The Instruments Should be FDA
	point 2	FDA or CE or BIS certified	or CE or BIS certified or ISO 13485 certified
2	page 50,	The Instruments should be of	
	point 3	medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 50,	Demonstration of sample	Items mentioned at BOQ Sl .No
	point 7	Instruments is mandatory and	4, 6, 14, 19, 23a, 23c, 23g are to
		feedback of evaluation	be demonstrated and feedback
		committee regarding	of evaluation committee
		satisfactory performance will	regarding satisfactory
		be taken into account for	performance will be taken into
		technical approval.	account for technical approval.
4	Page 50,	SS tray medium – 1 no.	SS tray medium – 3 no.
	BOQ point 1		

5	Page 50,	Towel clips large – 1 no.	Towel clips large – 13 no.
	BOQ point 2		
6	Page 50,	Towel clips medium – 4 no.	Towel clips medium – 23 no.
	BOQ point 3	_	
7	Page 51,	SS Bowl medium-1no.	SS Bowl medium- 10nos.
	BOQ point 20		
8	Page 51,		Sponge holder - 10nos.
	BOQ point 22	Sponge holder – 1no.	
9	Added Para:		Added BOQ para: Towel clips
	BOQ s.no.24		small – 19 nos.
10	Added Para:		Added BOQ para:SS Tray Large
	BOQ s.no.25		– 8 nos.
11	Added Para:		Added BOQ para: SS Bowl large
	BOQ s.no.26		– 6 nos.
12	Added Para:		Added BOQ para: Kidney tray
	BOQ s.no.27		medium – 1no.
13	Added Para:		Added BOQ para: Kidney tray
	BOQ s.no.28		big – 4 nos.

Item No. 6 (Rfx/ Event number 3000003194)

Dressing Set

Sl. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 52, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified
2	page 52, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 52, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ SNo 2,8 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
	Page 52, BOQ point 4	Small bowl	SS Bowl Small
4	Page 52, BOQ point 9	Stitch scissors	Stitch scissors 12-14cm

Item No. 7 (Rfx/ Event number 3000003195)

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 53, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO
	_		13485 certified
2	page 53, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
4	Page 53, BOQ point 2	Sponge Holder	Sponge Holder 24-26cm
5	Page 53, BOQ point 3	S.S. Bowl	S.S. Bowl Medium

Epidural Set

Item No. 8 (Rfx/ Event number 3000003196)

S1. No.	Ref. to the Bidding Document	Existing specification/clauses in the Bidding Document	Amended as
1	page 54, point 2	The Instruments Should be FDA or CE or BIS certified	The Instruments Should be FDA or CE or BIS certified or ISO 13485 certified
2	page 54, point 3	The Instruments should be of medical grade stainless steel (SS 304/SS 316), rust free	Deleted
3	Page 54, point 7	Demonstration of sample Instruments is mandatory and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.	Items mentioned at BOQ Sl. No. 4, 5, 7 are to be demonstrated and feedback of evaluation committee regarding satisfactory performance will be taken into account for technical approval.
4	Page 54, BOQ point 3	Sponge holder	Sponge holder 24-26 cm
5	Page 54, BOQ point 5	One tooth forceps	One tooth forceps 15-17 cm
7	Page 54, BOQ point 7	Curved scissors	Curved scissors 15-17 cm

CVP Tray

All other contents of the Bidding Document including terms & conditions remain unaltered.