Tender No. HLL/PCD/PMSSY-II/NAGPUR/11/15-16 Dated 06.11.2015									
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non- Responsiveness	Technical Responsiveness	Reason for Technical Non- Responsiveness	Overall Responsiveness		
	C-arm 3D/ CT for ortho	M/s Siemens Limited	Responsive	NA	Responsive	NA	Responsive		
1		M/s BET Medical (P) Ltd.	Responsive	NA	Responsive	NA	Responsive		
2	C-arm for general surgery	M/s Allengers Medical Systems Ltd.	Responsive	NA	Responsive	NA	Responsive		
	OT Table - Surgery	M/s Jolly Technocrats.	Responsive	NA	Non-Responsive	Valid IEC Certificate not submitted	Non-Responsive		
4		M/s Indian Surgical Equipment Co. Pvt Ltd.	Responsive	NA	Responsive	NA	Responsive		
5	OT Table - Bariatric	M/s Jolly Technocrats.	Responsive	NA	Non-Responsive	European CE with notified body number not submitted & certificate for confirmity to IEC 60601 not submitted	Non-Responsive		
		M/s Indian Surgical Equipment Co. Pvt Ltd.	Responsive	NA	Responsive	NA	Responsive		

	Tender No. HLL/PCD/PMSSY-II/NAGPUR/11/15-16 Dated 06.11.2015									
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non- Responsiveness	Technical Responsiveness	Reason for Technical Non- Responsiveness	Overall Responsiveness			
6	Ultrasonic cutting and coagulation device	M/s BNS Health Aids	Responsive	NA	Responsive	NA	Responsive			
		M/s Johnson & Johnson Limited	Responsive	NA	Responsive	NA	Responsive			
7	Advanced Laproscopy system	M/s B. Braun Medical India Pvt. Ltd.	Non-Responsive	End User Certificate/ PO not submitted	Non-Responsive	Technical specifications not complied	Non-Responsive			
		M/s Eagle Medical System Pvt. Ltd.	Non-Responsive	non acceptance of terms and conditions w.r.t. warranty and CMC clause	Non-Responsive	Technical specifications not complied	Non-Responsive			
		M/s Karl Storz Endoscopy India Pvt. Ltd.	Responsive	NA	Responsive	NA	Responsive			

	For Tender No.HLL/PCD/Rohtak/02/14-15 Dated 03.01.2015									
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non- Responsiveness	Technical Responsiveness	Reason for Technical Non- Responsiveness	Overall Responsiveness			
	Pneumatic duct system	Srishty Medical Pvt. Ltd.	Responsive		Responsive		Responsive			
11		Narula Udyog (India) Pvt. Ltd.	Responsive		Responsive		Responsive			
		Ace Orthocare Pvt Ltd	Non-Responsive	EMD not valid as per terms and conditions of the tender enquiry	Not Evaluated		Non-Responsive			

For Tender No.HLL/AMU/PMSSY-II/06RT-02/15-16 Dated 06.02.2016								
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non- Responsiveness	Technical Responsiveness	Reason for Technical Non- Responsiveness	Overall Responsiveness	
1	Kitechen Items- 1 Set	Bharati Refregeration Works	Responsive	NA	Responsive	NA	Responsive	

	HLL / PCD / PMSSY-II/07/15-16 Dtd. 28.10.2015								
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non- Responsiveness	Technical Responsiveness	Reason for Technical Non- Responsiveness	Overall Responsiveness		
1	Endoscopic Ultrasound	M/s SISCO	Not evaluated	-	Non Responsive	3D Imaging, Real time spatial compound Imaging not available in the offered machine.	Non Responsive		
		M/s DSS Imagetech Pvt. Ltd	Responsive	-	Responsive	1-	Responsive		
2	C-Arm for ERCP	M/s. Allengers Medical Systems Limited	Responsive	-	Responsive	-	Responsive		