

**MINUTES OF MEETING**

<b>MEETING TITLE</b>	Pre-Bid Meeting For Supply, Installation and Commissioning of EMG Machines (4nos.) (Tender no. : HLL/HCS/Tender-GeM/2025-26/29 DTD: 18.02.2026 ) GEM/2026/B/7254427
<b>DATE</b>	21.02.2026
<b>TYPE OF MEETING</b>	PHYSICAL/ VIRTUAL (GOOGLE MEET) meet.google.com/gyf-eobm-ypd
<b>MEETING DURATION</b>	2.00 PM – 3.00 PM

**PARTICIPANTS DETAILS**

<b>HLL Lifecare Limited</b>	Mr. Binu Thomas (VP-PS & GH- HCS)
	Mr. Jayakumar A (DVP - HCS)
	Mr. Narendranath (DGM-Biomedical)
	Mr. Hariprasad A P (SM -HCS)
<b>Prospective bidders participants</b>	Mr. Puneet Sharma- M/s. Allengers

Prebid meeting against the Tender floated for pre-bid meeting for supply, installation and commissioning of EMG Machines (4 nos.) at Hindlabs at 4 locations was conducted on 21.02.2026 at 02.00 PM at corporate and head office of HLL Lifecare Limited at Thiruvananthapuram. Prospective bidders were given the option to participate either in person or through the online link provided in the tender document. During the meeting, the bidders raised various queries and were advised to submit the same in writing via email to hcstenders@lifecarehll.com for further consideration. It was also informed that all queries received would be examined by the HLL team and clarifications/ responses would be published subsequently as the official Pre-Bid Minutes. In continuation to the Prebid meeting, the prospective bidders have also mailed their respective queries. Also representations were received through GeM platform also. Details of response to the queries are as below.

SNO	Representation by Bidder	HLL Reply
<b>ALLENGERS GLOBAL HEALTHCARE(P) LTD, NEWDELHI</b>		
<b>1</b>	<p><b>Point No. 4</b> Noise Level: 0.4 micro volts RMS <b>Request for Amendment:</b> Noise Level: 0.4 micro volts RMS TO 0.5 micro Volts RMS <b>Justification:</b> A noise level of <math>\leq 0.5 \mu\text{V}</math> RMS is clinically acceptable and widely adopted in standard EMG systems for accurate neurophysiological recordings. The marginal</p>	Noise Level: 0.5 micro volts RMS
<b>2</b>	<p><b>Point No. 29</b> The quoted model should have CE/FDA/BIS certification <b>Request for Amendment:</b> The quoted model should have CE/FDA/BIS/CDSCO certification <b>Justification:</b> As per Govt. Of India where Indian Standards are 2available there compulsion of CE or USFDA is not mandatory. So request to ask for CDSCO certificate. As this is the only Indian Standards available in EMG NCV EP Machine</p>	NO change

This minute of pre-bid meeting shall form part of bid document as Addendum/Amendment. All the bidders must attach a copy of this MoM, sealed and signed by authorized signatory, which indicates that all the points in this MoM is agreed by the bidder.