

/KIK-06/LE/2016-17

25.07.2016

Amendment No: 3

to

Tender ref. no. HITES/IDS/16/08/JIPMERII/KIK-06/LE/2016-17

Dated: 24.06.2016

for

Procurement of various Lab equipment's

at

**JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION &
RESEARCH , PUDUCHERRY CAMPUS AT KARAIKAL.**

The following amendments are incorporated in the referred tender enquiry document:

In Technical specification:

6. Binocular Microscope

Page No.	Para No.	Tender Specification	Read as
46	7	Abbe full plan Achromat condenser, Iris diaphragm	Abbe condenser, Iris diaphragm
46	8	Revolving Quadruple (5) nose piece (for objectives).	Revolving Quintuple (5) nose piece (for objectives).
46	11	Eye piece 12X (FOV 20) ó Magnification (wide angle eye piece).	Eye piece 10X (FOV 20) ó Magnification (wide angle eye piece).
46	12	Antifungal treatment should be applied to the observation tube , eyepiece and objective.	Antifungal treatment should be applied to the eyepiece and objective
46	14	Eye pieces pointers ó 10 nos.	Eye pieces pointers ó 1 no per unit

with Camera

No.	No.	Specification	Read as
47	5	Objectives - RP Series Infinity Corrected Plan 4X, 10X,40X(Spring Loaded), 100X (Spring Loaded, Oil Immersion)	Objectives - Infinity Corrected Plan 4X, 10X,40X(Spring Loaded), 100X (Spring Loaded, Oil Immersion)

16. Deep Freezer (-20 °C)

Page No.	Para No.	Tender Specification	Read as
54	B.5	Polyurethane Insulation around 150mm for better thermal insulation and sample safety in case or power failure.	Polyurethane Insulation around 75mm to 100 mm for better thermal insulation and sample safety in case or power failure.
54	C.2	Cascade cooling system with two hermetic compressors.	Cooling system with hermetic compressors.
55	D.5	Key Board lockable Battery Powered	Should have lockable keyboard

34. Laboratory Refrigerator – Type 1

Page No.	Para No.	Tender Specification	Read as
69		For storing blood plasma and other blood products, vaccines, other medical or pharmaceutical supplies. Also to cool samples or specimens for preservation. For faster pull-down and recovery times, it should have bypass refrigeration and microprocessor-based controls.	For storing blood plasma and other blood products, vaccines, other medical or pharmaceutical supplies. Also to cool samples or specimens for preservation. For faster pull-down and recovery times, it should have microprocessor-based controls.
69	9	Control panel with thermometer, main switch and temperature selection.	Control panel with digital display with temperature controller.

tor – Type 2

Page No.	Para No.	Tender Specification	Read as
69	8	Control panel with temperature alarm, on/off switch and digital thermometer.	Deleted
70	12	Control panel with temperature alarm, ON /OFF switch with power on indicator, digital thermometer, temperature display.	Control panel with temperature alarm, ON /OFF switch with power on indicator, digital display and temperature controller.

42. pH meter

Page No.	Para No.	Tender Specification	Read as
79	6	Refillable Triode 3-in-1 epoxy body combination pH electrode	Refillable Triode 3-in-1 epoxy /glass body combination pH electrode

44.Rotary Microtome with Knife / Blade

Page No.	Para No.	Tender Specification	Read as
81	10	Spare low and high profile blades in dispenser pack of 50 blades: 6 packets each.	Spare low profile blades in dispenser pack of 50 blades: 6 packets per unit (ie 300 low profile blades per unit) Spare high profile blades in dispenser pack of 50 blades: 6 packets per unit (ie 300 high profile blades per unit)
81	11	Microtome knives 6 02 nos.	Microtome knives 6 02 nos per unit.

45.Rotary Microtome for paraffin sections

Page No.	Para No.	Tender Specification	Read as
81	11	Disposal blades 5,000 number low profile.	Low profile Disposal blades 6 5000 nos per unit

Protome with disposable Blades

No.		ification	Read as
82	10	Spare low and high profile blades in dispenser pack of 50 blades: 6 packets each.	<p>Spare low profile blades in dispenser pack of 50 blades: 6 packets per unit (ie 300 low profile blades per unit)</p> <p>Spare high profile blades in dispenser pack of 50 blades: 6 packets per unit (ie 300 high profile blades per unit)</p>
82	11	Disposable blades ó 5000 nos.	Deleted

All other terms and condition remains unaltered.