

MINUTES OF MEETING

MEETING TITLE	Pre-Bid Meeting For Supply, Installation and Commissioning of ECG Machines (4 nos.) (Tender no. : HLL/HCS/Tender-GeM/2025-26/22 DTD: 09.02.2026) GEM/2026/B/7211844
DATE	12.02.2026
TYPE OF MEETING	PHYSICAL/ VIRTUAL (GOOGLE MEET)- meet.google.com/ivf-ystw-iex
MEETING DURATION	05:00 PM – 06:00 PM

PARTICIPANTS DETAILS

HLL Lifecare Limited	Mr. Binu Thomas (VP-PS & GH- HCS)
	Mr. Jayakumar A (DVP - HCS)
	Mr. Narendranath (DGM-Biomedical)
	Mr. Hariprasad A P (SM -HCS)
Prospective bidders participants	Mr. Puneet- M/s. Allengers

Prebid meeting against the Tender floated for pre-bid meeting for supply, installation and commissioning of ECG Machines (4 nos.) at Hindlabs at 4 locations was conducted on 12.02.2026 at 05.00 PM at corporate and head office of HLL Lifecare Limited at Thiruvananthapuram. Prospective bidders were given the option to participate either in person or through the online link provided in the tender document. During the meeting, the bidders raised various queries and were advised to submit the same in writing via email to hcstenders@lifecarehll.com for further consideration. It was also informed that all queries received would be examined by the HLL team and clarifications/ responses would be published subsequently as the official Pre-Bid Minutes. In continuation to the Prebid meeting, the prospective bidders have also mailed their respective queries. Also we have representations were received through GeM platform also. Details of response to the queries are as below.

REPRESENTATIONS IN GEM		
1	1) Request to amend Point 13: 12 Nos. Reusable button type electrode TO 12 PKTS DISPOSABLE BUTTON TYPE ELECTRODE. [Reason: As Reusable button-type electrodes are not readily available in the market. Replacing them with disposable electrodes ensures consistent ECG signal	Should be supplied with 2 patient cable sets, 8 clip on reusable electrodes, 12 chest reusable electrode with silicon rubber bulb, 12 packets recording paper, 1 bottle of jelly and 12 packets of disposable button type electrode with suitable cables.

	quality, hygiene, patient safety, and operational efficiency. Quality certification will confirm compliance with medical standards.]	
2	2) No Quality certification has been asked as it is a life saving product. Request to ask for Certification like Bureau of Indian Standards (BIS) or European CE from Notified Body or US FDA [Reason: ECG machine are life-saving medical devices. Considering their critical role in patient safety, it is imperative that only devices meeting recognized international/Indian quality and safety standards. Using certified devices protects the hospital from legal issues or liability in case of device failure. Certified devices undergo rigorous testing for electrical safety, electromagnetic compatibility, and biocompatibility. Reduces the risk of device malfunction or patient harm during clinical use. As many companies are selling Non standard ECG Machines in India which does not have any Quality Certification to prove that there machines are quality certified and are safe to use on Patients. Also Services are not provided by such companies at later stages.]	CDSCO is already required as per stipulated specifications. Added clause : The quoted model should have CE issued by notified body /US FDA/BIS certification and the same shall be submitted along with technical bid.

This minute of pre-bid meeting shall form part of bid document as Addendum/Amendment. All the bidders must attach a copy of this MoM, sealed and signed by authorized signatory, which indicates that all the points in this MoM is agreed by the bidder.