

**Ministry of External Affairs
Development Partnership Administration- I**

Tender No. DPA-I/230/95/2013/3-1

13 January, 2015

CORRIGENDUM - I

Subject: Tender for supply and installation of medical equipment to JFK Medical Centre, Monrovia, Liberia

The following amendments are hereby issued in connection with the above mentioned tender.

Amendment No. 1

Clause 5 on page 7 **for Group X-1 and X-2** is modified as below:

“Important Dates:

- *Sale of tender: up to 1600 hrs on **11/01/2016***
- *Deadline for submission of bids: up to 1130 hrs on **12/02/2016***
- *Opening Date of Technical Bid : at 1200 hrs on **12/02/2016***

Note: There is no change in the specified dates for **Group-Y** (CT scan machine).

Amendment No. 2

Clause 3 (b) on page 10 is modified as below:

*“Failure of the above by the Supplier may lead to forfeiture of the **Performance Bank Guarantee** and may face action as per Government rules.”*

Amendment No. 3

Clause 16 on page 14: (Performance Bank Guarantee): a new sub clause is added below Clause 16 c.

“d. The bid of the successful bidder will be treated as void if he does not furnish a Performance Bank Guarantee within the stipulated period of 10 working days.”

Amendment No. 4

Annexure II: technical specifications

X-1.14 para D on page 32 is modified as below:

“Detector Type may be CCD/PDA/Deuterium /Tungsten “.

X-1.33 para 2 on page 51 is modified as below:

*“System should be based on windows platform with 17" color monitor having minimum resolution of 1280 x 1024, **500 GB** HDD, DVD-RW, Mouse, UPS for analyser”.*

X-2.12 para C-6 on page 63 is modified as below:

*“Should have **bright colour** LCD / TFT display for viewing messages and ECG waveform of 4 seconds”*

X-2.14 para B on page 64 is modified as below:

"The unit should be operative on mains voltage from single phase 180-260 v AC with automatic main compensation or the unit with a servo controlled voltage stabilizer of suitable capacity from 180-260V AC".

X-2.14 para J on page 64 is modified as below:

"Stationary/Rotating Anode tube with focal spot 1.8 X 1.8 mm or better or 0.6/1.8mm or better".

X-2.15 para 3.21 on page 67 is now **deleted**.

X-2.16 para A-17 on page 68 is modified as below:

“Storage-should have >2,00,0000 image storage facility in the hard disk drive. Should have inbuilt hard disk for image storage”.

X-2.18 para D on page 71 is modified as below:

“it should be 25 cm minimum”.

X-2.19 para B.7 on page 72 is now **deleted**.

Amendment No. 5

Annexure VII (Manufacturer's Authorisation Form) on page 89 is modified as below. (Modified MAF is applicable to Group X-1 and Group X-2)

“

Annexure VII

Manufacturer's Authorization Form

Subject: Tender for Supply, Installation and Maintenance of Medical equipment at JFK Medical Centre, Monrovia in Liberia (Group-X / Group-Y).

To
(Name of the Purchaser)

Dear Sirs,

Ref. your TE document no _____,
dated _____

We, _____ who are proven and reputable manufacturers of _____ (name and description of the goods offered in the tender) having factories at _____ hereby authorize Messrs _____ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also hereby extend our full warranty of the General Conditions of the Contract, read with modification, if any, in the Special Conditions of the contract for the goods and services offered for supply by the above firm against this tender document.

Yours faithfully,

[Signature with date, name and designation]

For and on behalf of Messrs _____

[Name & address of the manufacturers]

Note:

1. This letter of authorization should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
2. Original letter may be sent.

“

Note:

1. MAF will only be acceptable, if Tender document no: DPA-1/230/95/2013/3 or Tender document no: DPA-1/230/95/2013/3-1 is specified in MAF.
2. This corrigendum shall be treated as a part of the original tender document.
3. All other terms and conditions of the original tender document shall remain unchanged.
