

**Amendment No. 1****Date: 22/01/2019****Sub: Amendment to the Tender Enquiry Document.****Ref: NIT No.: HLL/PCD/CDSCO-06/RT-III/18-19 Dated 26.12.2018**

The following changes are being incorporated in the referred tender enquiry.

**SECTION I  
NOTICE INVITING TENDERS (NIT)****1.) EXISTING:**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
c	Closing date & time for submission of <b>tender processing fee and EMD in physical form</b>	<b>25.01.2019, 1400 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender processing fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	<b>25.01.2019, 1300 hrs IST</b>
e	Time and date of opening of online bids	<b>25.01.2019, 1430 hrs IST</b>

**READ AS:**

<b>Sl. No.</b>	<b>Description</b>	<b>Schedule</b>
c	Closing date & time for submission of <b>tender processing fee and EMD in physical form</b>	<b>05.02.2019, 1400 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender processing fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	<b>05.02.2019, 1300 hrs IST</b>
e	Time and date of opening of online bids	<b>05.02.2019, 1430 hrs IST</b>

**SECTION – XII  
QUESTIONNAIRE****ADDED AS:****Fill up the Techno-Commercial Compliance Sheet Bid provided in spreadsheet (Excel file) and upload in the C-Folder.**

## SECTION-VII TECHNICAL SPECIFICATIONS

Para	TENDER SPECIFICATION	READ AS
Page 44	Item No 1. Analytical balance with 5 digits display	Item No 1. Analytical Balance (5 digits) with printer along with Anti Vibration table
<b>Item No. 02 Tensile Strength Tester</b>		
Para 22 Pg 45	Compatible PC and suitable printer also to be provided.	Suitable compatible laser printer to be supplied. Suitable online UPS with surge protector with 01 hr back up . The UPS battery should be covered during both warranty and CMC periods.
<b>Item No. 03 Amino Acid Analyzer</b>		
Pg 47	Compatible PC and suitable printer also to be provided.	Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor and laser printer to be supplied Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.
<b>Item No. 05 Ion Chromtographic System</b>		
Pg 48	PC & Printer with support Software Windows 7 or any other compatible version & colored printer with automatic both side printing facility. Compatible branded online 3 KVA UPS with 1 hour back up.	Computer with i7 processor, 4GB RAM 1 TB Hard disk, 21" monitor, windows software and colored printer with automatic both side printing facility to be supplied Suitable online UPS with surge protector with 01 hr back up should support both the instrument and computer. The UPS battery should be covered during both warranty and CMC periods.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**