## **Techno commercial Responsiveness & Non Responsiveness**

Procurement & Consultancy Division: HITES – HLL Lifecare Limited

For Tender Ref: HLL/PCD/PMSSY/AIIMS-II-11/13-14 dated 19.12.2013								
Sl. No	Quoted Item No.	Item Name	Name of the Bidder	Technical Evaluation Status	Reason for technical non- responsiveness	Commercial Responsiveness	Reason for commercial non- responsiveness	Overall Responsivenes s
			Elekta Medical Systems India Pvt. Ltd	Responsive	Responsive	Responsive	•	Responsive
1	12(a)	High energy LINAC	Varian Medical Systems International (India) Pvt. Ltd	Non- Responsive	Non Responsive: i) Offered diode based Ditector instead of ion chember based Detector ii) Not quoted Cylindrical and rotational Phantom with inclinometer iii)Not quoted Ion chamber array resulting in field size of 27 cm x 27cm iv) Not quoted an additional array device with 900 or above liquid field ionized chembers v)Not quoted for upgradation of	Not Evaluated		Non- Responsive

HLL/HITES/TC/Nov.'16 Page 1 of 3

## **Techno commercial Responsiveness & Non Responsiveness**

Procurement & Consultancy Division: HITES – HLL Lifecare Limited

					SRS/ SRT/immobilizati on device (4 sets) vi)Not quoted for integrated lifting carriage and water resorvoir		
2	12(b)	Low energy LINAC	Elekta Medical Systems India Pvt. Ltd	Non- Responsive	Non Responsive: i) Offered diode based Ditector instead of ion chember based Detector ii) Not quoted Cylindrical and rotational Phantom with inclinometer iii)Not quoted Ion chember array resulting in field size of 27 cm x 27cm iv) Not quoted an additional array device with 900 or above liquid field ionized chembers v)Not quoted for	Not Evaluated	Non- Responsive

## **Techno commercial Responsiveness & Non Responsiveness**

Procurement & Consultancy Division: HITES – HLL Lifecare Limited

					upgradation of SRS/ SRT/immobilizati on device ( 4 sets) vi)Not quoted for integrated lifting carriage and water resorvoir		
			Varian Medical Systems International (India) Pvt. Ltd	Responsive	Responsive	Responsive	Responsive
		HDR	Elekta Medical Systems India Pvt. Ltd	Responsive	Responsive	Responsive	Responsive
3	12(c)	Brachythe rapy	Varian Medical Systems International (India) Pvt. Ltd	Responsive	Responsive	Responsive	Responsive

HLL/HITES/TC/Nov.'16 Page 3 of 3