21-07-2017

## Amendment No. 3

### Sub: Amendment to the referred tender enquiry

# Ref.: Tender Enquiry HITES/PCD/PMSSY-III/07/CTVS/2017-18 dated 08-06-2017 read with Amendment No. 01 dated 03-07-2017 & Amendment No. 02 dated 10-07-2017

The following changes have been incorporated in the technical specifications of the referred tender enquiry.

# Section VII

## **Technical Specification**

	Sch 08. Ultrasonic Cutting and Coagulation System		
Para	Existing Specification	READ AS	
3	It should have both 5mm and 9-10mm instruments.	It should have compatibility for 5mm or 9-10mm instruments.	
6	It should have vibration range of 50-100 $\mu m.$	Should have vibration frequency in of 25 KHz or more	
8	The system should have an auto switch- off mechanism.	Deleted	
12	Necessary accessories required for cardiac procedures are to be provided.	Necessary accessories required for cardiac procedures are to be provided as detailed in the BOQ below. Also need to provide rates for various accessories compatible for the system.	
		Added Para: System should have Vessel sealing capacity with open and laparoscopic shears up to 5 mm or bigger vessel size.	
	BOQ	BOQ	
3	Dissecting Shears Curved Handpiece Laparoscopic with transducer a. Dissecting Shears 9-10mm Curved Handpiece Laparoscopic - qty 10 pcs.	Dissecting Shears Curved Handpiece Laparoscopic with transducer a. Dissecting Shears Curved Handpiece Laparoscopic - qty 10 pcs	
4	Vessel Welder Straight Handpiece Open with transducer a. Vessel Welder Straight Handpiece Open - qty 10 nos b. Vessel Welder Straight Transducer Open - 01 reusable or 25 disposables	Vessel Welder /cutting and sealing Straight Handpiece Open with transducer a. Vessel Welder/ cutting and sealing Straight Handpiece Open - qty 10 nos b. Vessel Welder/ cutting and sealing Straight Transducer Open - 01 reusable or 25 disposables	

Vessel Welder Straight Handpiece Laparoscopic with transducer a. Vessel Welder Straight Handpiece Iaparoscopic - qty 10 no's b. Vessel Welder Straight Transducer Iaparoscopic - 01 reusable or 25 disposablesVessel Welder / cutting and sealing Straight Handpiece Caproscopic - a. Vessel Welder/ cutting and sealing Straight Handpiece Open - qty 10 nos b. Vessel Welder Straight Transducer Iaparoscopic - 01 reusable or 25 disposablesVessel Welder / cutting and sealing Straight Handpiece Open - qty 10 nos b. Vessel Welder / cutting and sealing Straight Transducer Open - 01 reusable or 25 disposables
--

## All other contents of the tender enquiry including terms & conditions remain unaltered.

### Note:

i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids