

**Amendment No. 3****28.02.19****Sub: Amendment to the Bidding Document**

**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/38/18-19 dated 09.01.2019 read with its Amendment No. 1 & 2 dated 11.02.19 & 22.02.19 respectively.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document for item bearing 'Rfx no. 3000003725' only.

**SECTION – VII****TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:**

**Item No:- 2. (Rfx no. 3000003725)**

**Ultrasound Machine- High End**

<b>Sl. No.</b>	<b>Reference to the bidding document</b>	<b>Existing Specification</b>	<b>Amended as</b>
1	Point 4 , page 56	The system should include at least 21” LCD monitor with 1280 x 1024 resolution arm with tilt, swivel and height adjustment facility. A separate touchscreen should also be available.	The system should include at least 21” LCD monitor with 1280 x 1024 resolution arm with tilt, swivel and height adjustment facility. A separate <b>12” or more</b> touchscreen should also be available.

**Note:**

Technical Compliance Sheet in HLL e-bid portal 'for Rfx / Event number 3000003725' has been replaced by a new version i.e. 'Ver1.1'. The same need to be downloaded for submission.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**