

**HLL LIFECARE LIMITED**  
 (A Government of India Enterprise)  
 Projects Division  
 Corporate Head Office, Poojappura.P.O,  
 Thiruvananthapuram – 695012,  
 Kerala, India  
 Phn: 0471- 2354949, 2775588

**AMENDMENT NO. 2**

**IFB Ref No. HLL/CHO/PROJ/HCS/MCG -EQP/USG/2018-19 dated 10.10.2018**

**Amendment Date: 02-11-2018**

**AMENDMENTS BASED ON PRE-BID MEETING DATED 24.10.2018**

<b>Sub:</b>	<b>Supply, Installation, Testing and Commissioning of USG WITH ECHO PROBE &amp; its accessories for Diagnostic Lab Facility at Sector 39, Gurugram - reg.</b>		
<b>Ref No.</b>	<b>HLL/CHO/PROJ/HCS/MCG -EQP/USG/2018-19 dated 10.10.2018</b>		
<b>Sl No.</b>	<b>Tender Details (Page No, Clause etc.)</b>	<b>Original Clause</b>	<b>Amendment</b>
1	Schedule II - User Interface & Ergonomics - Pt. 1	The keyboard should have Height adjustment. The adjustment should also include Keyboard rotation.	The control panel should have Up and Down Height adjustment facility.
2	Schedule II - Scanning Parameters - Pt. 1	The system should have minimum 65,000 digital system processing channels.	The system should have minimum 1,50,000 or more digital system processing channels.
3	Schedule II - Image archive & Networking - Pt. 6	b. Linear, with biopsy attachment. Operating Frequency: 5 – 13 MHz.	b. Linear, with biopsy attachment. Operating Frequency: 4 – 12 MHz & Pediatric echo transducer with frequency from 3 - 8 MHz.

**All other terms and conditions of the tender remain unchanged.**

**SENIOR MANAGER (PROJECTS)**