

**Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)**

**NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018**

**Schedule No. 14 - Rigid Bronchoscope (Rfx no. 3000002790)**

Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
1	Page 78 Para 1,k	Adaptor from bronchoscope to any type of pediatric respiration equipment.	Karl Storz	Specs not clear, if adaptor for respiration is required for pediatric model then same adaptor will work	<b>Amended as:</b> Adaptor from bronchoscope to any type of adult respiration equipment
2	Page 79 Para 3.b	Should have motorized optical zoom up to 2x	Karl Storz	Kindly delete motorized optical zoom lens as it is company specific and amend it to integrated zoom lens.	<b>Amended as:</b> Should have integrated zoom lens
3	Page 79 Para 3.c	Should have motorized focus function which can be varied steplessly from coarse to fine image.	Karl Storz	Kindly delete motorized focus function as it is company specific and amend it to focus function.	<b>Amended as:</b> Should have focus function which can be varied steplessly from coarse to fine image.
4	Page 79 Para 3.e	The camera head must be autoclavable for achieving better Cleaning, Disinfection and Sterilisation.	Karl Storz	Kindly give option for autoclavable camera head/soakable camera head for disinfection and sterilization.	<b>Amended as:</b> The camera head must be sterilizable for achieving better Cleaning, Disinfection and Sterilisation
5	Page 79 Para 4.a	Separate/combined module from video processor	as per pre bid discussion		<b>Deleted</b>
6	Page 79 Para 4.c	Powerful 300 Watt Xenon Lamp with backup lamp of halogen/LED	Karl Storz	Kindly delete back up lamp of halogen/LED as it is company specific and amend it to "powerful 300 watt xenon lamp".	<b>Amended as:</b> Powerful 300 Watt Xenon Lamp/LED light source.
7	Page 79 Para 4.d	Lamp life should not be less than 500 hrs (approx), one spare 300W xenon lamp should be supplied in case of Xenon light source			<b>Amended as:</b> Lamp life should not be less than 500 hrs (approx). 2 nos. spare 300W xenon lamp should be supplied in case of Xenon light source

*Handwritten signatures and initials:*

*Sunder* (with signature)

*De* (with signature)

*Deo* (with signature)

*Chand* (with signature)