Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)					
NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018					
Schedule No. 09- Flexible Nasopharyngolaryngoscope (Rfx/Event number 3000002785)					
Sl. No	**Reference to the bidding document**	**Specification/Clause of the bidding documents**	**Name of firms**	**Queries/Clarification asked by firms**	**Recommendations/replies by the Committee**
1	B-2	angle of view 100-110 deg	Olymphus	Kindly amend it as 85 degree or more	**Amended as:** angle of view 85 degree or more
			Fujiflim	angle of view 100-deg or more. Justification: For better participation	
2	B-3	working length 20-25cm or batter	Olymphus	Kindly amend it as 30-40 cm or more.Remark: as 20-25 cm is too small length to visualize pharynx & larynx	**Amended as: Working length 25cm or more**
			Fujiflim	working length 25cm or batter,Justification: For better participation	
3	B-4	outer diameter 5-5.5mm	Olymphus	Kindly amend it 4.8-5.5 mm	**Amended as: Outer diameter: 5.5mm or less**
			Fujiflim	outer diameter 5-5.5mm or lesser	
4	B-6	deflection upwards 150 and downward 150	Olymphus	Kindly amend Deflection as Upward & Downward 130 degree or more	**Amended as: Deflection upwards 130deg or more and downward 100 deg or more**
			Fujiflim	deflection upwards 130 or more and downward 130 or more,Justification: For better participation	
			Karl Storz	Kindly amend this to "Deflection: Upward upto 180 deg, Downward upto 100 deg.".	