

Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)				
NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 24.04.2018				
Schedule No. 16 (Rfx/Event number 3000002792) Rigid Proctoscope				
Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms
1	Page 84 Para 15 & 15.a to 15.d	Light Source: Should have following technical specifications/features Separate/combined module from video processor Must be latest and high end model from bidder Powerful 300 Watt Xenon Lamp with backup lamp of Halogen/LED Lamp life should not be less than 500 hrs (approx), one spare 300W xenon lamp should be supplied in case of Xenon light source		
2	Page 85 BOQ Sl.no.15	Light Source as per technical specifications - 1 no.		
				Deleted







