Techno-Commercial Evaluation Result on Responsiveness for the Bids received against the tender enquiry no. HLL/PCD/CATS/NCT-RT/02/14-15

Tooling Commission Production Responsiveness for the Entertocated against the tortion stightly file. HELD Co. (1975)						
Name of Equipment	Name of Bidder	Commercial Responsiveness	Resaon for Commercial Non Responsiveness	Technical Responsiveness	Reason for Technical Non Responsiveness	Overall Responsiveness
ALS & BLS Ambulances	M/s Bafna Healthcare (P) Ltd. in consortium with M/s Tata Motors Ltd.	Responsive	-	Responsive	-	Responsive
	M/s Aeon Medical Pvt. Ltd. in consortium with M/s Ashok Leyland Ltd.	Responsive	1	Non-Responsive	Against the tender requirement of 'M' Category Chassis, bidder has quoted a model which is not in existance under 'M' Category as on date. However, firm has submitted the type approval certificate of ARAI in respect of N2 category vehicle which is "Goods Vehicle" and not in line with the tender enquiry condition	Non-Responsive
	M/s Instromedix (India) Pvt. Ltd. in consortium with M/s SML Isuzu Ltd	Responsive	-	Non-Responsive	1. CMVR Type approval certificate for the quoted model not submitted and M category chassis not evident. Firm has informed that the offered vehicle is presently under test run, which itself proves that as on date the offered vehicle does not have valid certification under CMVR. 2. Non compliant bubble type O2 flowmeter offered against the requirement of dial type which is not suitable for use in ambulance 3. Against the requirement of certificate of EN1789 of ambulance cot, bidder has submitted Partial Test report (static test only) of EN 1789 for ambulance cot 4. Valid European CE or US FDA certificate for Oxygen Delivery system not submitted	Non-Responsive
	M/s Tempo Automobiles Pvt. Ltd. in consortium with M/s BVG India Ltd.	Responsive	-	Responsive	-	Responsive