

Amendment No. 6

Date: 10/12/2015

Subject: Amendment to the Tender Enquiry Document

Ref: Tender Enquiry No.: HLL/HITES/PCD/RC-ME/01/2015 dated 06/10/2015 read with amendment no. 1, 2, 3, 4 & 5 dated 21.11.2015, 23.11.2015, 30.11.2015, 05.12.2015 & 07.12.2015 respectively.

The following changes have been incorporated in the referred Tender Enquiry Document (TED).

<p style="text-align: center;">SECTION VII Technical Specification</p>

Item No. 26
C-Arm SYSTEM (HF)

X-Ray Tube:

Existing Para:- Anode heat storage capacity should be **more than** 200KHU or more.

Read as:- Anode heat storage capacity should be 200KHU or more.

Item No. 102
Operation Table Orthopaedic

Existing Para:- 4.4 protection:- Should have over-charging cut-off **with visual symbol**.

Read as:- 4.4 protection:- Should have over-charging cut-off.

Existing Para:- 7.1 1. 7.1 1. Should be FDA/CE/BIS and ISO 13485 approved product.

Read as:- **7.1 1.** Should be FDA/CE from an European **Notified Body**/BIS and ISO 13485 approved product.

Item No. 108
Shadow Less Lamp Ceiling Type Major

Existing Para:- **2.1 Technical characteristics :-** Intensity Control in **9** steps for individual domes.

Read as:- **2.1 Technical characteristics:-** Intensity Control in steps for individual dome.

Existing Para:- Height Adjustment :600mm.

Read as:- Height Adjustment :**Approx** 600mm.

Existing Para:- Action Radius :1850mm.

Read as:- Action Radius :**Approx** 1850mm.

Existing Para:- 7.1 1. 7.1 1. Should be FDA/CE/BIS and ISO 13485 approved product.

Read as:- **7.1 1.** Should be FDA/CE from an European **Notified Body**/BIS and ISO 13485 approved product.

Item No. 113
Operation Table Electro-Hydraulic (Electrical With Manual Over Side)

Existing Para:- 4.4 protection:- Should have over-charging cut-off **with visual symbol.**

Read as:- 4.4 protection:- Should have over-charging cut-off.

Existing Para:- 7.1 1. 7.1 1. Should be FDA/CE/BIS and ISO 13485 approved product.

Read as:- **7.1 1.** Should be FDA/CE from an European **Notified Body**/BIS and ISO 13485 approved product.

Item No. 116
Shadow less lamp standing model

Existing Para:- 2.1 Technical characteristics :- Dome Head :515mm Dia.

Read as:- 2.1 Technical characteristics :- Dome Head :**Approx** 515mm Dia.

Existing Para:- 2.1 – 2. LED lights-2 nos.

Read as:- Deleted.

Existing Para:- 2.1 – 3. Lockable castor **stands with minor dome.**

Read as:- Should have lockable castor.

Existing Para:- 7.1 1. 7.1 1. Should be FDA/CE/BIS and ISO 13485 approved product.

Read as:- **7.1 1.** Should be FDA/CE from an European **Notified Body**/BIS and ISO 13485 approved product.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

Prospective bidders are advised to check the website regularly prior to the closing date and time of online submission of tenders.

-- End --