

Amendment No.2

Date: 03/05/2014

Subject: Amendment No.2 to the Tender Enquiry Document

**Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/12/13-14 dated 03/03/2014.
(ii) Amendment No.1 dated 15.04.2014**

The pre-bid meeting for the referred tender enquiry was held on 10/03/2014. Based on pre-bid discussions following amendments are being incorporated in the referred tender enquiry document.

Section – VII **Technical Specifications**

Schedule No. 1

ETO Sterilizer

1. Existing Specification:

Para : 9: Capacity: Should have capacity of **200- 250 litre**

Read as: Capacity: Should have capacity of **200 litre or more**

2. Existing Specification:

Para : The consumable required for 200 cycles to be provided.

Read as: The **consumables including packing material** required for 200 cycles to be provided.

3. Existing Specification:

Para 8d: Gas disposal arrangement/catalytic converter.

Read as: Appropriate pollution control device for safe disposal of E.O. like catalytic converters or equivalent technology gas disposal management to be as per local pollution control norms

ETO Room, Item no. 2 **Wire Storage Shelf Module for Steril Store**

Item to be read as **Compatible** Wire Storage Shelf module for Steril Store

All other terms and conditions of the tender enquiry remain unaltered.