

AMENDMENT NO.1 Dated 06.10.2025

REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

Amendment made to the details given in Tender Document.

REF NO: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

CPP TENDER ID : 2025_HLL_250885_1

The following amendment has been incorporated in the above mentioned **E-TENDER FOR SUPPLY OF LOW THROUGHPUT HbA1C (HPLC METHOD) ANALYZER FOR HINDLABS DIAGNOSTIC CENTERS ON REAGENT RENTAL BASIS.**

The following amendment / Inclusion have been incorporated to the bid document for the above tender:

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 2 - Complete ready to use kit should be provided with Buffers in transparent plastic tanks to view the level of buffers; columns, primers, calibrators & sample vials.

MAY BE READ AS

Point No.2 – Complete ready to use kit should be provided with Buffers in transparent plastic tanks to view the level of buffers; columns, primers, calibrators & sample vials / similar pack with facility to know the quantity of buffers; columns, primers, calibrators & sample vials.

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 6 - Should provide sample rack – 3No, Pipette rack – 1 No

MAY BE READ AS

Point No: 6 - Should provide Sample rack – 3 No, Pipette rack – 1 No (If required for testing).

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 13 - Assay time should not be more than 4 minutes for HbA1c testing

MAY BE READ AS

Point No: 13 - Assay time should not be more than 3 minutes for HbA1c testing

AMENDMENT NO.1 Dated 06.10.2025

REF TENDER: HLL/CHO/PSD/HCS/2025-26/TENDER 14 Dated 30th September 2025

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 19 - It should have a hard disk and a remote data access feature when connected to LAN or Intranet.

MAY BE READ AS

Point No: 19 - It should have internal storage capacity and a remote data access feature when connected to LAN or Intranet.

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 23 - Should have offline library of chromatograms for result interpretation.

MAY BE READ AS

Point No: 23 - Should have inbuilt database of chromatograms for result interpretation

FOR

ANNEXURE 4 – PRODUCT DETAILS AND SPECIFICATION REQUIRED

Point No: 26 – New point added

MAY BE READ AS

Point No: 26 - Should be CDSCO registered.

All relevant clauses of the Tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & Group Head (HCS)